



**Supplemental figure 1**

**Comparison of the biosensor output to sorbitol in buffer solution or blood plasma. The final concentration of sorbitol was 100  $\mu\text{mol/L}$  in all samples. When the output of the sorbitol biosensor for sorbitol standard solution was 100%, the sensor output in pre-treated blood plasma was 86.7%.**