

Supplemental information

ICP-OES

Element wavelength table

Element	wavelengths	Element	wavelengths	Element	wavelengths
Cu	324.752	Fe	238.204	Sr	421.552
Pb	220.353	Mn	257.610	Be	313.107
Zn	213.857	Al	396.153	Mg	279.077
Co	228.616	Ga	294.364	Rb	780.023
Cr	205.560	Ge	209.426	Ca	317.933
Cd	214.440	Sr	460.733	Ba	233.527
Ni	232.003	Sr	407.771	Nb	309.418

ICP-MS

Smart tune solution criteria

Mg 23.985	>8000
In 114.904	>40000
U 238.05	>30000
Bkgd 220	≤2
Bkgd 8.5	≤3
CeO 155.9/Ce 139.905	≤0.03
Ba ⁺⁺ 68.9525/Ba 137.905	≤0.03

Smart tune solution from Perkin Elemer contained 10 ppb Ba and 1 ppb Be, Ce, Co, Fe, In, Mg, Pb, Th, U.

ICP-MS Auto Lens settings

Analyte	Mass	# Points	Max	
			Dac Value	Intensity
Be	9.0122	45	4.25	1500.16
Co	58.9332	45	5.25	33420
In	114.904	45	6.25	51244.2
U	238.05	45	7.5	40974.2

ICP-MS Dual Detector calibration

analyte	mass	# Points	Coefficient	Gain	N (max)	Conversion
						Factor
Be	9.0122	17	0.99996	11715.8	1.07E+09	0.119782
Na	22.9898	8	0.999794	10518.5	1.19E+09	0.107541
Mg	23.985	18	0.999942	9745.94	1.28E+09	0.0996422

Al	26.9815	16	0.999933	9426.53	1.33E+09	0.0963765
Ca	43.9555	22	0.998705	12223.3	1.02E+09	0.124971
Cr	51.9405	14	0.999919	8871.48	1.41E+09	0.0907016
Mn	54.9381	12	0.99982	8569.42	1.46E+09	0.0876135
Fe	56.9354	13	0.999851	8740.44	1.43E+09	0.0893619
Co	58.9332	11	0.999747	8123.28	1.54E+09	0.0830521
Ni	59.9332	26	0.999947	8063.86	1.55E+09	0.0824446
Cu-1	62.9298	21	0.999861	7751.99	1.62E+09	0.0792561
Cu	64.9278	27	0.999942	7795.78	1.61E+09	0.0797037
Zn-1	65.926	26	0.999951	7967.14	1.57E+09	0.0814557
Zn	67.9249	20	0.999959	8012.08	1.56E+09	0.0819153
Ga	68.9257	16	0.99996	8031.53	1.56E+09	0.082114
Ge	73.9219	29	0.999951	7963.11	1.57E+09	0.0814145
Rb	84.9117	12	0.99982	7903.87	1.58E+09	0.0808089
Sr	87.9056	9	0.999827	7893.91	1.59E+09	0.0807071
Nb	92.906	10	0.999768	8172.64	1.53E+09	0.0835568
Cd	110.904	32	0.999961	7330.59	1.71E+09	0.0749476
Ba	137.905	15	0.999774	7112.79	1.76E+09	0.0727209
Pb	207.977	27	0.999865	6365.53	1.97E+09	0.065081

ICP-MS Interference Correction equation

*analyte original = intensity original – abundance interference original/abundance interfereent correction *interferent correction*

*Mass, Corrections: Sr 83.9, -4.96 *Kr-83;*

*Sr-83.9 = intensity 83.9 – Kr-84 which is Sr-83.9 = intensity 83.9 – (57%/11.5%).96(Kr-83)
intensity 83.9 – 4.96(Kr-83)*

*Mass, Corrections: Sr 85.9, -1.505*Kr-83;*

*Sr-85.9= intensity 85.9– Kr-86 which is Sr-85.9= intensity 85.9 – (17.3%/11.5%) Kr-83=
intensity 85-1.505*Kr-83*

*Mass, Corrections: Sr 86.9, -0.385*Rb-85;*

*Sr-86.9= intensity 86.9– Rb-87 which is Sr-86.9= intensity 86.9-(27.8%/75.16%)*Rb-85=
intensity 86.9-0.385*Rb-85*

*Mass, Corrections: In 114.9, -0.014*Sn118*

*In 114.9= intensity 114.9– Sn115 which is In 114.9= intensity 114.9-(0.34%/24.22%)*Sn118=
intensity 114.9-0.014*Sn118*

ICP-MS Analysis of Sr-84

Analysis of Sr-84 in purified Sr-82, QC samples and the calibration curve (inset). Sample data was not on the calibration curve, so the data and calibration curve was validated using a 1, 10, 50, 1000, 5000, 10000 ppb strontium QC standards which were within 100+/-10% of the true value and %RSD were <5%.

Gamma Spectroscopy

Energy and efficiency was calibrated multinuclide standard. Quality control checks of the detector were determined with Eu-152 standard. Both standards were NIST traceable and purchase from Eckert & Ziegler Analytics Inc (GA, USA)

Eu-152 standard:

Detected energy of gamma 1408.05 keV (true value 1408.08 keV) acceptance range 1407.58-1408.58 keV

FWHM 1.79 % acceptance range +/- 5%

Detected energy of gamma 344.3 keV (true value 344.3 keV) acceptance range 343.5-344.8 keV

FWHM 1.16 % acceptance range +/- 5%