

Supplemental Information

Figure S1 Effect of electrospray voltage on glutathione signal intensity/response. While the spray voltage of 1.85 kV showed the highest signal intensity, it was not robust for all chips used. Therefore, 2.5 kV was used for our experiments.

