

Electronic Supplemental Information for:

Control of Nitrosamines During Non-Potable and *de Facto* Wastewater
Reuse with Medium Pressure Ultraviolet Light and Preformed
Monochloramine

Daniel L. McCurry¹, Stuart W. Krasner², and William A. Mitch^{1,}*

¹Department of Civil and Environmental Engineering,
Stanford University, Stanford, CA 94305

²Metropolitan Water District of Southern California
La Verne, CA 91750

*Corresponding author phone: (650) 725-9298; fax: (650) 723-7058;
e-mail: wamitch@stanford.edu

Table S1: Nitrosamine GC/MS analytical details

Compound	Retention Time (min)	Parent Ion	Quantification Ions
<i>Internal Standards</i>			
d6-NDMA	14.12	81	49+50+63+64
d8NMOR	20.56	125	93+95
<i>Analytes</i>			
NDMA	14.16	75	44+47+58
NMEA	15.74	89	61
NDEA	17.01	103	75
NDPA	19.85	131	89
NPIP	21.3	115	69
NPYR	20.93	101	55
NMOR	20.59	117	86+87
NDBA	22.75	159	57