

Supplementary Materials

In-vitro and ex-vivo measurements of biophysical properties of blood using microfluidic platforms and animal models










Yang Jun Kang¹ and Sang-Joon Lee^{2*}

1. Department of Mechanical Engineering, Chosun University, Gwangju, Republic of Korea
2. Department of Mechanical Engineering, Pohang University of Science and Technology (POSTECH), Pohang, Republic of Korea

*Address: Department of Mechanical Engineering, Pohang University of Science and Technology (POSTECH), 77 Cheongam-Ro, Nam-gu, Pohang, Gyeongbuk 790-784, Republic of South Korea. Fax: +82-54-279-3199

*E-mail: sjlee@postech.ac.kr

S1. Summary of commercialized products for measurement of haemorheological properties

	Commercialized Product			
Viscosity or viscoelasticity	Health Onvector (USA)  Hemathix Scv-300	Kyoto electronics (Japan)  EMS-1000	Nano Biz (Korea)  Viscore	Vilastic Scientific (USA)  Vilastic-3
Deformability or Aggregation	RR Mechatronics (Netherlands)  LORRCA	Myrenne GmbH (Germany)  Rheodyn SSD	Rheomeditech (Korea)  RheoScan-D	
ESR	LabONE (India)  ESR 2010	RR Mechatronics (Netherlands)  Starsed ST		