

Table 1S

Tyrosol determination of simulated intestinal fluid with known concentrations of added tyrosol by the HPLC method.

$C_{(\text{TYR})}$ added (μM)	$C_{(\text{TYR})}$ founded (μM)	Accuracy (%)	Intraday coefficient variations (%)	Interday coefficient variations (%)
5.00	5.20 ± 0.50	104.6	10.1	11.7
50.00	49.35 ± 0.73	98.8	5.9	7.3
100.00	93.57 ± 0.94	94.6	9.5	9.9