

Standards	K-factor				
		$^{207}/^{206}\text{Pb}$	$^{208}/^{207}\text{Pb}$	$^{208}/^{204}\text{Pb}$	$^{207}/^{204}\text{Pb}$
Internal standard (Tl)	K-factor	0.9954	0.9954	0.9818	0.9863
	Standard deviation	0.0003	0.0003	0.0011	0.0009
External standard (SRM-981, NIST)	K-factor	0.9966	0.9973	0.9951	0.9978
	Standard deviation	0.0006	0.0003	0.0014	0.0013

Table of calculated K-factors for mass bias correction including their associated errors.