Electronic Supplementary Material (ESI) for Lab on a Chip. This journal is © The Royal Society of Chemistry 2018

On-chip oocyte denudation from cumulus-oocyte complexes for assisted reproductive therapy

Lindong Weng^{1,2}, Gloria Y. Lee^{2,3}, Jie Liu^{2,3}, Ravi Kapur^{1,2}, Thomas L. Toth⁴ and Mehmet Toner^{1,2,3}

¹BioMEMS Resource Center, The Center for Engineering in Medicine, Massachusetts General Hospital and Harvard Medical School, Boston, MA 02129, USA

²Department of Surgery, Massachusetts General Hospital and Harvard Medical School, Boston, MA 02114, USA

³Shriners Hospital for Children, Boston, MA 02114, USA

⁴Department of Obstetrics and Gynecology, Massachusetts General Hospital and Harvard Medical School, Boston, MA 02114, USA

Supporting Information




Figure S1. Count of COCs falling within different diameter ranges (A) and the effect of COC size on denudation efficiency (B). A total of 103 GV stage COCs from 11 mice were measured to generate the size distribution. In (B), the group of small COCs have an average diameter of 125 μ m whereas those large COCs have an average diameter of 175 μ m.




Figure S2. Breakdown of denudation outcome into three categories. Complete: oocytes are completely denuded; Partial: oocytes are partially denuded with only a portion of zona pellucida exposed; None: oocytes are fully enclosed within the corona radiate. (A) The optimal condition features W_3 =90 µm, N=100 and θ =141°. (B) Effect of constriction width (W_3) on the denudation score. (C) Effect of repeating number (N). (D) Effect of teeth tilting angle (θ). The percentages are the mean of multiple replicates for each condition.

Table S1. Results of in vitro fertilization (IVF) on oocytes denuded on a chip or manually

Denudation method	Experiment No.	Number of oocytes inseminated	Number of two- cells (%ª)	Number of blastocysts (%b)
On-chip	#1	42	24 (57.1)	24 (100)
	#2	59	48 (81.4)	46 (95.8)
	#3	61	50 (81.9)	47 (94)
	#4	40	19 (47.5)	18 (94.7)
	#5	65	43 (66.2)	41 (95.3)
	#6	59	46 (77.9)	45 (97.8)
Manual	#1	37	23 (62.2)	22 (95.7)
	#2	54	37 (68.5)	35 (94.6)

#3	52	27 (51.9)	23 (85.2)
#4	51	36 (70.6)	36 (100)
#5	71	30 (42.3)	25 (83.3)
#6	54	45 (83.3)	45 (100)

^a Percent of survived oocytes developed to two-cell embryos ^b Percent of two-cell embryos developed to blastocysts

Table S2. Results of intracytoplasmic sperm injection (ICSI) on oocytes denuded on a chip or manually

Denudation method	Experiment No.	Number of oocytes survived injection	Number of two- cells (%ª)	Number of blastocysts (%b)
On-chip	#1	14	14 (100)	11 (78.6)
	#2	16	15 (93.8)	10 (66.7)
	#3	18	18 (100)	13 (72.2)
	#4	16	13 (81.3)	10 (76.9)
	#5	10	10 (100)	9 (90)
	#6	12	12 (100)	10 (83.3)
	#7	12	12 (100)	9 (75)
	#8	14	13 (92.9)	8 (61.5)
Manual	#1	7	6 (85.7)	5 (83.3)
	#2	13	13 (100)	10 (76.9)
	#3	11	11 (100)	8 (72.7)
	#4	10	8 (80)	7 (87.5)
	#5	6	4 (66.7)	2 (50)
	#6	11	9 (81.8)	7 (77.8)
	#7	14	13 (92.9)	12 (92.3)
	#8	11	11 (100)	8 (72.7)

^a Percent of survived oocytes developed to two-cell embryos ^b Percent of two-cell embryos developed to blastocysts