

## checkCIF/PLATON report

You have not supplied any structure factors. As a result the full set of tests cannot be run.

THIS REPORT IS FOR GUIDANCE ONLY. IF USED AS PART OF A REVIEW PROCEDURE FOR PUBLICATION, IT SHOULD NOT REPLACE THE EXPERTISE OF AN EXPERIENCED CRYSTALLOGRAPHIC REFEREE.

No syntax errors found.      CIF dictionary      Interpreting this report

### Datablock: DIBrAdaRietveld294k

---

Bond precision:    C-C = 0.0028 Å                      Wavelength=1.54056

Cell:                      a=14.9792(9)              b=9.7677(6)              c=7.2734(5)  
                            alpha=90              beta=90              gamma=90

Temperature:              294 K

	Calculated	Reported
Volume	1064.19(12)	1064.19(12)
Space group	P n m a	P n m a
Hall group	-P 2ac 2n	-P 2ac 2n
Moiety formula	C10 H14 Br2	C10 H14 Br2
Sum formula	C10 H14 Br2	C10 H14 Br2
Mr	294.01	294.01
Dx,g cm-3	1.835	1.835
Z	4	4
Mu (mm-1)	9.212	0.000
F000	576.0	0.0
F000'	571.32	
h,k,lmax		
Nref		
Tmin,Tmax		
Tmin'		

Correction method= Not given

Data completeness=                      Theta(max)=

R(reflections)=                      wR2(reflections)=

S =                      Npar=

---

The following ALERTS were generated. Each ALERT has the format  
**test-name\_ALERT\_alert-type\_alert-level.**  
Click on the hyperlinks for more details of the test.

---

### Alert level A

ATOM007\_ALERT\_1\_A \_atom\_site\_aniso\_label is missing  
Unique label identifying the atom site.  
PLAT029\_ALERT\_3\_A \_diffrn\_measured\_fraction\_theta\_full value Low . 0.000 Why?  
PLAT183\_ALERT\_1\_A Missing \_cell\_measurement\_reflns\_used Value .... Please Do !  
PLAT880\_ALERT\_1\_A NO datum for \_diffrn\_reflns\_number ..... Please Do !  
PLAT881\_ALERT\_1\_A No Datum for \_diffrn\_reflns\_av\_R\_equivalents ... Please Do !

---

### Alert level B

PLAT201\_ALERT\_2\_B Isotropic non-H Atoms in Main Residue(s) ..... 9 Report  
Br1 Br2 C1 C2 C3 C4 etc.

---

### Alert level C

PLAT162\_ALERT\_4\_C Missing or Zero s.u. (esd) on y-coordinate for . C4 Check  
PLAT162\_ALERT\_4\_C Missing or Zero s.u. (esd) on y-coordinate for . C5 Check  
PLAT162\_ALERT\_4\_C Missing or Zero s.u. (esd) on y-coordinate for . C6 Check

---

### Alert level G

PLAT005\_ALERT\_5\_G No Embedded Refinement Details Found in the CIF Please Do !  
PLAT164\_ALERT\_4\_G Nr. of Refined C-H H-Atoms in Heavy-Atom Struct. 8 Note  
PLAT808\_ALERT\_5\_G No Parseable SHELXL Style Weighting Scheme Found Please Check  
PLAT882\_ALERT\_1\_G No Datum for \_diffrn\_reflns\_av\_unetI/netI ..... Please Do !  
PLAT883\_ALERT\_1\_G No Info/Value for \_atom\_sites\_solution\_primary . Please Do !  
PLAT980\_ALERT\_1\_G No Anomalous Scattering Factors Found in CIF ... Please Check

---

- 5 **ALERT level A** = Most likely a serious problem - resolve or explain  
1 **ALERT level B** = A potentially serious problem, consider carefully  
3 **ALERT level C** = Check. Ensure it is not caused by an omission or oversight  
6 **ALERT level G** = General information/check it is not something unexpected

- 7 ALERT type 1 CIF construction/syntax error, inconsistent or missing data  
1 ALERT type 2 Indicator that the structure model may be wrong or deficient  
1 ALERT type 3 Indicator that the structure quality may be low  
4 ALERT type 4 Improvement, methodology, query or suggestion  
2 ALERT type 5 Informative message, check
-

It is advisable to attempt to resolve as many as possible of the alerts in all categories. Often the minor alerts point to easily fixed oversights, errors and omissions in your CIF or refinement strategy, so attention to these fine details can be worthwhile. In order to resolve some of the more serious problems it may be necessary to carry out additional measurements or structure refinements. However, the purpose of your study may justify the reported deviations and the more serious of these should normally be commented upon in the discussion or experimental section of a paper or in the "special\_details" fields of the CIF. checkCIF was carefully designed to identify outliers and unusual parameters, but every test has its limitations and alerts that are not important in a particular case may appear. Conversely, the absence of alerts does not guarantee there are no aspects of the results needing attention. It is up to the individual to critically assess their own results and, if necessary, seek expert advice.

### **Publication of your CIF in IUCr journals**

A basic structural check has been run on your CIF. These basic checks will be run on all CIFs submitted for publication in IUCr journals (*Acta Crystallographica*, *Journal of Applied Crystallography*, *Journal of Synchrotron Radiation*); however, if you intend to submit to *Acta Crystallographica Section C* or *E* or *IUCrData*, you should make sure that full publication checks are run on the final version of your CIF prior to submission.

### **Publication of your CIF in other journals**

Please refer to the *Notes for Authors* of the relevant journal for any special instructions relating to CIF submission.

