





**Figure (1S):** Plots of RMSECV against the number of factors for calibration set of the binary mixtures using the developed PLSR models for the raw spectral data.



**Figure (2S):** The analytes reference concentrations against the predicted concentrations

**Table 1S:** The recovery percent obtained after applying the PLSR model to the test set in the presence of the expected excipients.

Mixture no <sup>#</sup>	Analytes (R%)		Mixture no <sup>#</sup>	Analytes (R%)	
	HYS	KTP		BENZ	DEX
<b>1</b>	97.98	98.87	<b>4</b>	102.12	97.82
<b>2</b>	97.88	101.58	<b>6</b>	97.54	99.51
<b>9</b>	98.13	99.93	<b>10</b>	98.80	101.87
<b>16</b>	98.43	87.11	<b>12</b>	99.67	99.86
<b>17</b>	99.76	99.41	<b>15</b>	97.09	99.03
<b>19</b>	98.98	98.59	<b>17</b>	98.88	98.77
<b>22</b>	97.98	97.79	<b>18</b>	101.87	97.66
<b>23</b>	98.55	102.10	<b>21</b>	101.65	98.34
<b>24</b>	98.56	97.99	<b>24</b>	98.82	100.82
<b>25</b>	99.76	99.69	<b>25</b>	97.71	101.92

<sup>#</sup> The concentration levels provided in (Table 1).

**Table 2S:** System suitability parameters for the determination of HYS and KTP by the HPLC method.

<b>Parameters</b>	<b>Obtained value</b>		<b>Reference value</b>
	<b>HYS</b>	<b>KTP</b>	
<i>Resolution (R)</i>	6.78		$R > 0.8$
<i>Capacity factor (K)</i>	2.85	5.08	1-10 acceptable
<i>Selectivity factor (<math>\alpha</math>)</i>	1.78		$> 1$
<i>Number of theoretical plates (N)</i>	2002.23	4916.46	Increases with increasing the efficiency of separation
<i>Tailing factor (T)</i>	1.79	1.93	$T \leq 2$ T= 1 for a typical symmetric peak
<i>Height equivalent to one theoretical plate (H)</i>	0.12	0.05	The smaller the value, the higher the column efficacy

**Table 3S:** Validation parameters of the results obtained by applying the HPLC method for HYS and KTP determination.

<b>Parameter</b>	<b>Value</b>	
	<b>HYS</b>	<b>KTP</b>
<b>•Linearity:</b>		
Slope	3760.3	18983
Intercept	-11537	196793
Correlation coefficient (r)	0.9999	0.9998
Range ( $\mu\text{g/ml}$ )	10-200	50-1000
<b>•Accuracy</b> <i>(Mean <math>\pm</math> S.D):</i>	$99.63 \pm 0.92$	$99.5 \pm 1.4$
<b>•Precision (R.S.D%):</b>		
Repeatability	1.13	1.94
Intermediate Precision	1.29	1.98
<b>•Specificity:</b> <i>(mean <math>\pm</math> SD)</i>	$100.33 \pm 1.67$	$101.11 \pm 0.73$

**Table 4S:** System suitability parameters for the determination of DEX and BENZ by the HPLC method.

<b>Parameters</b>	<b>Obtained value</b>		<b>Reference value</b>
	<b>DEX</b>	<b>BENZ</b>	
<i>Resolution (R)</i>	11.88		$R > 0.8$
<i>Capacity factor (K)</i>	1.89	4.25	1-10 acceptable
<i>Selectivity factor (<math>\alpha</math>)</i>	2.25		$> 1$
<i>Number of theoretical plates (N)</i>	4332.89	8613.83	Increases with increasing the efficiency of separation
<i>Tailing factor (T)</i>	1.38	1.20	$T \leq 2$ $T = 1$ for a typical symmetric peak
<i>Height equivalent to one theoretical plate (H)</i>	0.06	0.03	The smaller the value, the higher the column efficacy

**Table 5S : Validation parameters of the results obtained by applying the HPLC method for DEX and BENZ determination.**

<b>Parameter</b>	<b>Value</b>	
	<b>DEX</b>	<b>BENZ</b>
<b>•Linearity:</b>		
Slope	17638	28580
Intercept	-13938	30915
Correlation coefficient (r)	0.9998	0.9998
Range (µg/ml)	5-100	5-100
<b>•Accuracy</b> (Mean $\pm$ S.D):		
	100.73 $\pm$ 2.03	100.61 $\pm$ 1.61
<b>•Precision (R.S.D%):</b>		
Repeatability	2.43	1.70
Intermediate Precision	2.05	1.69
<b>•Specificity:</b> (mean $\pm$ SD)		
	99.29 $\pm$ 1.20	99.82 $\pm$ 1.66