

## Supporting Information

# Detecting hydrogen peroxide reliably in water by ion chromatography: a method evaluation update and comparison in the presence of interfering components

**Yuanxi Huang<sup>a</sup>, Lei Wang<sup>a</sup>, Baiyang Chen<sup>a\*</sup>, Qi Zhang<sup>a</sup>, Rongshu Zhu<sup>a\*</sup>**


<sup>a</sup> Shenzhen Key Laboratory of Organic Pollution Prevention and Control, Harbin Institute of Technology, Shenzhen 518055, China

\*Corresponding author. Phone: 86-134-8072-7605; Emails: chen.baiyang@hit.edu.cn, rszhu@hit.edu.cn


Total pages: 6, including cover page

Number of figures: 3


Number of tables: 2



**Figure S1. A depiction of  $\text{H}_2\text{O}_2$  and typical anions' chromatograms in water by IC with (a) conductivity meter (CM) and (b) UV detector.**



**Figure S2.** The spectra of waters spiked with  $\text{H}_2\text{O}_2$  and KI as measured by the colorimetric method.



**Figure S3. A comparison of the IC and KI methods under the influences of nitrate for  $\text{H}_2\text{O}_2$  measurement at (a) 2.0 mg/L and (b) 10.0 mg/L  $\text{H}_2\text{O}_2$ .**

**Table S1. H<sub>2</sub>O<sub>2</sub> MDLs calculation for (a) the KI method and (b) the IC method.**

<b>(a) Test concentration = 30 µg/L</b>	
Sample #	C (µg/L)
1	33.60
2	29.40
3	34.65
4	36.75
5	34.65
6	27.30
7	28.35
8	30.45
9	43.04
10	44.09
n	10
Average	34.23
Variance	33.61
StDev	5.80
t (n-1, 0.01)	2.82
t value for 99% at n (from Table)	
MDL(t *StDev)	
16.35	
C-MDL=	13.65
MDL*10-C=	133.53
S/N=	5.90
MDL (µg/L)=	16.35
LOQ (µg/L)=	57.97
X <sup>2</sup> (n-1, 0.975)	2.70
X <sup>2</sup> (n-1, 0.025)	19.02
	X <sup>2</sup> /df
LCL	0.30
UCL	2.11
	(X <sup>2</sup> /df) (MDL)
	4.91
	34.6
<b>(b) Test concentration= 5 µg/L</b>	
Sample #	C (µg/L)
1	2.79
2	2.32
3	4.40
4	3.26
5	2.70
6	3.36
7	4.49
8	4.21
9	2.88
10	3.45
n	10
Average	3.39
Variance	0.57
StDev	0.76
t (n-1, 0.01)	2.82
t value for 99% at n (from Table)	
MDL(t *StDev)	
2.14	
C-MDL=	2.86
MDL*10-C=	16.38
S/N=	4.47
MDL (µg/L)=	2.14
LOQ (µg/L)=	7.58
X <sup>2</sup> (n-1, 0.975)	2.70
X <sup>2</sup> (n-1, 0.025)	19.02
	X <sup>2</sup> /df
LCL	0.3
UCL	2.1
	(X <sup>2</sup> /df) (MDL)
	0.6
	4.5

**Table S2. Common characteristics of the tap water used for the photolysis tests.**

Parameters	Unit	Value
NPOC	mg-C/L	1.44
TN	mg-N/L	1.43
Free chlorine	mg-Cl <sub>2</sub> /L	0.16
UV <sub>254</sub>	1/cm	0.03
NO <sub>3</sub> <sup>-</sup>	mg/L	6.02
NO <sub>2</sub> <sup>-</sup>	mg/L	0.00
pH	/	6.80

NPOC: non-purgeable organic carbon; TN: total nitrogen.