Electronic Supplementary Material (ESI) for Lab on a Chip. This journal is © The Royal Society of Chemistry 2021

Supplementary Information

A New Paper-Based Biosensor for Therapeutic Drug Monitoring

Darlin Lantigua^{a, b}, Jamie Trimper^{b, c}, Baris Unal^d, and Gulden Camci-Unal^{b,e*}

^a Biomedical Engineering and Biotechnology Program, University of Massachusetts Lowell, One University Avenue, Lowell, MA, 01854, USA

^b Department of Chemical Engineering, University of Massachusetts Lowell, One University Avenue, Lowell, MA, 01854, USA

^c Department of Biomedical Engineering, University of Massachusetts Lowell, One University Avenue, Lowell, MA, 01854, USA

^d Triton Systems, Inc. 200 Turnpike Road, Chelmsford, MA 01824, USA

^e Department of Surgery, University of Massachusetts Medical School, 55 Lake Avenue North, Worcester, MA, 01605, USA

*Correspondence: Gulden_CamciUnal@uml.edu



Figure S1. Colorimetric detection and quantification of tacrolimus in human blood. Detection and quantification of tacrolimus in 20 μ L of unmodified whole human blood at 10 and 3 ng/mL concentrations. Red color formation increased with decreasing tacrolimus concentration as expected for a competitive assay. The detection time was <10 minutes. The images were taken using an Android cellphone camera. Quantification of red color formation was performed using the NIH ImageJ. Statistical analysis was performed by unpaired t-test in GraphPad Prism (La Jolla, CA, USA). Error bars: \pm SD, *p < 0.05.