

ARTICLE

Computer Vision Enabled Funnel Adapted Sensing Tube (FAST) for Power-Free and Pipette-Free Nucleic Acid Detection

Mengdi Bao,^a Shuhuan Zhang,^a Chad ten Pas,^a Stephen J. Dollery,^b Ruth V. Bushnell,^b FNU Yuqing,^a Rui Liu,^a Guoyu Lu,^c Gregory J. Tobin,^b and Ke Du^a

Table 1. Sequences of DNA and crRNA used in this study.

Forward primer	5'-GCCATTGTCATGATGGAAAAGCACACTTCCT-3'
Reverse primer	5'-CTCATTGAGGCGGTCAATTCTTTGAATGTT-3'
crRNA	5'- UAAUUUCUACUAAGUGUAGAUUGUCUGGUACUGU GAUGUUGUA-3'
SARS-CoV-2	TS: 5'... TTTG TGTCTGGTAACGTGATGTTGTA ...-3'
crRNA	uaauuuacuacuaaguguagauAUCGGUAAGAAUAGGUUU GC

Calculation process to convert detection sensitivity in fM to that in copies/μL

$$10 \text{ fM} \sim \frac{10 \text{ fmole} \times 6 \times 10^{23} \text{ copies/mole}}{1 \text{ L}} \quad (1)$$

$$\sim \frac{10^{-15} \text{ mole} \times 6 \times 10^{23} \text{ copies/mole}}{10^6 \mu\text{L}} \quad (2)$$

$$\sim 6 \times 10^{23} \text{ copies}/\mu\text{L} \quad (3)$$

^a Department of Mechanical Engineering, Kate Gleason Collage of Engineering, Rochester Institute of Technology, Rochester, NY 14623, United States

^b Biological Mimetics, Inc. 124 Byte Drive, Frederick, MD 21702, United States

^c Department of Electrical and Computer Engineering, University of Georgia, Athens, GA 30602, United States