Supplementary data

Preparation of a new composite based on multilayer fullerene with mesoporous carbon nitride and its application in extraction of tacrolimus and everolimus from plasma prior to LC-MS/MS analysis

Rana Honarnezhad ^a, Mohammad Reza Afshar Mogaddam^{a, b,*}, Elnaz Marzi Khosrowshahi^{a,*}, Mir Ali Farajzadeh ^{b,c}, Mahboob Nemati^{a,}

^a Food and Drug Safety Research Center, Tabriz University of Medical Science, Tabriz, Iran

^b Pharmaceutical Analysis Research Center, Tabriz University of Medical Science, Tabriz, Iran

^c Department of Analytical Chemistry, Faculty of Chemistry, University of Tabriz, Tabriz, Iran

^d Engineering Faculty, Near East University, 99138 Nicosia, North Cyprus, Mersin 10, Turkey

*Corresponding authors: Dr. M.R. Afshar Mogaddam and Dr. E. Marzi Khosrowshahi Email addresses: mr.afsharmogaddam@yahoo.com and elnazmarzi@yahoo.com Tel.: +98 4131772354;

DµSPE-LC-MS/MS					
Reagents	PPs				
ACN	4				
Adsorbent	4				
Deionized water	0				
	$\sum 8$				
Instruments	PPs				
HPLC-MS/MS	2				
Waste	3				
Heater	2				
	$\sum 7$				
Total penalty points	15				
AES score	85				

Table S1. Evaluation of greenness of the proposed method using Analytical Eco-Scale (AES).

Table S2	2. Evalua	ation of	the pro	posed r	nethod a	applica	bility ı	using]	BAC	ЪI.
							2	0		

Factors	Values
Type of analysis	10
Number of analytes that are simultaneously determined	5.0
Analytical technique- instrumentation	5.0
Number of samples that can be simultaneously treated	5.0
The sample preparation scale	7.5
Number of samples that can be analysed per hour	5.0
Type of reagents and materials	5.0
Requirement for preconcentration	7.5
The automation degree	5.0
The amount of sample	10
BAGI score	65.0

CCSHLLE -MDES-DLLME-HPLC-DAD