

The necessity of matrix-matched standards usage in quantitative imaging of specimens with calcareous skeleton by LA-ICP-MS

Ewelina Pollak-Kowa^{a,b}, Anna Telk^{a*}, Marcin Wieczorek^a

^aJagiellonian University, Faculty of Chemistry, Gronostajowa St 2, 30-387 Cracow, Poland

^bJagiellonian University, Doctoral School of Exact and Natural Sciences, Prof. St. Łojasiewicza
St 11, 30-348, Cracow, Poland

*corresponding author e-mail: anna.telk@uj.edu.pl

Table S1 Operating conditions

parameter	ICP-MS spectrometer			Laser system			
	wet analysis		LA analysis		parameter	UP-266	ESL 213
	ELAN DRC-e	ELAN DRC-e	NexION 5000				
RF power [W]	1300	1300	1600	Wavelength [nm]	266	213	
Nebuliser gas [L min^{-1}]	1.08	1.08	0.65	Ablation mode	line	line (imaging)	
Auxiliary gas [L min^{-1}]	1.02	1.02	1.20	Spot size [μm]	20	20 x 20 (XYR Shutter)	
Plasma gas [L min^{-1}]	17.00	17.00	16.00	Scan speed [$\mu\text{m s}^{-1}$]	80		
Sweeps	30	1	1	Offset between lines [μm]	20	(no gap)	
Reading	2	adjusted to the size of the analysed sample		Repetition rate [Hz]	10		
Replicate	3	1	1	Power [%]	100	25	
Dwell time [ms]	50	40		Energy density [J cm^{-2}]	7.22	8.86	
Measured masses	^{24}Mg , ^{88}Sr , ^{138}Ba , ^{43}Ca			He carrier gas flow rate [L min^{-1}]	0.8		

Table S2 Digestion programme

Stage	Time [min]	Max. pressure [bar]	Min. pressure [bar]	Max. temperature [°C]	Min. temperature [°C]	Power [%]
I	5:00	20	18	260	250	60
II	5:00	28	26	280	270	85
III	10:00	40	38	290	285	100

Fig. S1 The comparison of elements distribution without and with internal standardisation




Fig. S2 Examples of 3D distribution maps: Ca, Mg, and Mg standardised against Ca




Table S3 Minimal, maximal, mean intensity and coefficient of variation (standard deviation/ mean value x 100%) of the normalised against

Ca signals in the obtained pellets images and of the averaged standardised signals from the five ablated lines located randomly at each pellet


Ablation type	Pellet	24Mg				88Sr				138Ba			
		Min. int.	Max. int.	Mean int.	CV (%)	Min. int.	Max. int.	Mean int.	CV (%)	Min. int.	Max. int.	Mean int.	CV (%)
rectangular area of 15 lines	1	0.154	0.553	0.268	18	0.131	3.847	0.488	65	0.036	7.675	0.810	138
	2	0.140	0.523	0.267	17	0.119	1.992	0.375	51	0.065	4.519	0.326	144
	3	0.125	0.553	0.256	25	0.134	2.178	0.488	62	0.051	6.426	0.752	148
5 lines located randomly	1	0.151	0.450	0.261	8	0.675	6.839	0.502	14	0.300	5.072	0.770	28
	2	0.169	0.452	0.268	7	0.472	8.948	0.481	54	0.178	5.803	0.758	12
	3	0.176	0.494	0.272	4	0.896	3.896	0.560	15	0.269	5.732	0.882	41

Table S4 Minimal, maximal, mean intensity and coefficient of variation (standard deviation/ mean value x 100%) of the non-normalised signals in the obtained pellets images and of the averaged signals from the five ablated lines located randomly at each pellet


Ablation type	Pellet	²⁴ Mg				⁸⁸ Sr				¹³⁸ Ba			
		Min. int. [cps]	Max. int. [cps]	Mean int. [cps]	CV [%]	Min. int. [cps]	Max. int. [cps]	Mean int. [cps]	CV [%]	Min. int. [cps]	Max. int. [cps]	Mean int. [cps]	CV [%]
rectangular area of 15 lines	1	325,207	1,210,994	632,866	18	295,474	7,743,583	1,162,928	67	81,210	18,453,375	1,938,846	139
	2	307,089	2,804,460	567,611	24	280,820	3,215,789	786,164	49	149,645	9,218,855	684,701	147
	3	308,553	1,201,083	600,207	27	310,591	4,416,103	1,161,799	65	114,440	14,825,949	1,831,928	153
5 lines located randomly	1	315,382	1,138,577	586,165	12	301,053	5,254,136	1,129,936	18	129,490	12,495,567	1,738,197	31
	2	371,406	996,389	607,622	4	260,375	7,299,223	1,078,714	50	163,632	20,764,737	1,673,484	16
	3	278,660	1,088,650	605,332	8	291,584	3,516,461	1,243,942	15	151,657	13,157,906	1,963,881	41

Fig. S3 Calibration curves used for LOD and LOQ establishment

24Mg



88Sr



138Ba




Table S5 The concentration of elements in the in-house reference material established against various calibration standards without internal standardisation

Element	Laser wavelength (nm)	calibration standard NIST 610		calibration standard NIST 612		liquid-powder calibration standards	
		c (mg kg ⁻¹)	RE (%)	c (mg kg ⁻¹)	RE (%)	c (mg kg ⁻¹)	RE (%)
²⁴ Mg	266	141.9 ± 4.9	30	28.8 ± 0.9	-74	<LOD	-
	213	97.2 ± 0.5	-11	11.7 ± 0.2	-89		
⁸⁸ Sr	266	19,840 ± 790	721	25,900 ± 1,000	973	1,787 ± 69	-26
	213	11,380 ± 260	371	8,650 ± 200	258	1,741 ± 39	-28
¹³⁸ Ba	266	193.8 ± 8.1	883	213 ± 9	985	19.6 ± 0.9	-1
	213	114.9 ± 2.9	483	61.3 ± 1.6	211	26.8 ± 0.7	36

Fig. S4 Internal standardisation problem with a settlement plate

