

**Development of plasmid reference material to improve the accuracy of
quantitative detection of *Bovine Parvovirus***

Caihang Zhang^{a, b, †}, Mengke Lv^{a, b, †}, Yanmin Lin^a, Jiayi Yang^a, Yu Zhang^a, Xia
Wang^a, Jingya Yang^{b, c, *}, Lianhua Dong^{a, *}

^a Center for Advanced Measurement of Science, National Institute of Metrology, 100029, Beijing,
P.R. China

^b College of Food Science and Technology, Shanghai Ocean University, 201306, Shanghai, P.R.
China

^c Marine Biomedical Science and Technology Innovation Platform of Lin-gang Special Area,
Shanghai 201306, China

*Corresponding author: 1) Lianhua Dong, Tel: 86-10-64524959, FAX: 86-10-64524959, E-mail:
lianhuadong@126.com.

2) Jingya Yang, Tel: 86-21-61900382, FAX: 86-21-61900365, E-mail: jyayang@shou.edu.cn.

[†] These authors made equal contributions to this work and are considered co-first authors.

Supplementary Table

Table S1. Nanodrop detection results of BPV plasmid RM

Gene	A260/A280	A260/A230
BPV-VP1	1.92	2.01

Table S2. Primer and probe sequences of BPV-VP1 dPCR

Assay	Type	Sequence (5'-3')	length (bp)
Assay A	Forward	CAAGCACATCCAATCAAC	198
	Reverse	CCACAATGTTCTCGCTAA	
	Probe	FAM-CGTCCATCCGCCAGTTGAGTA-BHQ1	
Assay B	Forward	CCAGTACCAGGAAACGGAGAC	118
	Reverse	GCATGTATTCCGGTCTCCAA	
	Probe	HEX-CCTCAACATCTACGTCACCGGACAA-BHQ1	

Table S3. The partition volumes of different dPCR platforms

Platform	Partition volumes	Partition number	Individual reaction volume(μL)
A ¹	0.8 nL	Approx.23000	22
B ²	0.85 nL	Approx.20000	20
C ³	0.82 nL	Approx.26000	40

Table S4. Detection results of dPCR positivity in low concentration samples

Concentration (copies/reaction)	Mean (copies/reaction)	Positive/Total
4	3.76	18/20
6	6.05	20/20
8	8.26	20/20

Table S5. DPCR detection results for low concentration samples (6 copies/reaction)

Repetitions	Concentration (copies/reaction)	Repetitions	concentration (copies/reaction)
1	1.76	11	7.48
2	3.74	12	11
3	9.24	13	3.74
4	7.26	14	1.76

5	5.72	15	5.5
6	11.44	16	3.74
7	5.5	17	3.74
8	3.74	18	3.52
9	9.46	19	3.96
10	9.24	20	9.46
Mean		6.05	
standard deviation (s)		3.04	
LOD		9	
LOQ		30	

Table S6. Homogeneity assessment of BPV plasmid RM

Unit	Concentration (copies/ μ l)	
	Rep1	Rep2
1	2.24×10^6	2.09×10^6
2	2.15×10^6	2.09×10^6
3	2.06×10^6	2.22×10^6
4	2.11×10^6	2.18×10^6
5	2.20×10^6	2.32×10^6
6	2.25×10^6	2.44×10^6
7	2.16×10^6	2.26×10^6
8	2.16×10^6	2.16×10^6
9	2.20×10^6	2.07×10^6
10	2.27×10^6	2.40×10^6
11	2.22×10^6	2.18×10^6
Mean	2.20×10^6	
M_{between}	1.27×10^{10}	
M_{within}	6.87×10^9	
F	1.85	
$F_{0.05(10,11)}$	2.85	
S_{bb}	5.39×10^4	
S_r	8.29×10^4	
u'_{bb}	3.83×10^4	
$u_{\text{hom rel}}$	1.74%	

Table S7. The long-term stability assessment of BPV plasmid RM

Events		Concentration (copies/ μ L)
Time (month)	0	2.21×10^6
	1	2.27×10^6
	2	2.25×10^6
	3	2.12×10^6

	6	2.39×10^6
	9	2.19×10^6
	12	2.39×10^6
Mean		2.26×10^6
$ \beta_1 $		1.09×10^4
$s(\beta_1)$		9.04×10^3
$t_{0.95,3} \times s(\beta_1)$		2.32×10^4
conclusion		$ \beta_1 < t_{0.95,3} \times s(\beta_1)$
u_{Its}		1.09×10^5
$u_{Its \text{ rel}}$		4.80%

Table S8. The reference value of the BPV plasmid RM.

Unit	Concentration (copies/ μ L)		
	Rep1	Rep2	Rep3
1	2.22×10^6	2.24×10^6	2.09×10^6
2	2.04×10^6	2.15×10^6	2.09×10^6
3	2.38×10^6	2.44×10^6	2.47×10^6
4	2.04×10^6	2.06×10^6	2.22×10^6
5	2.11×10^6	2.16×10^6	2.18×10^6
6	2.05×10^6	2.08×10^6	2.04×10^6
7	2.20×10^6	2.32×10^6	2.32×10^6
8	2.34×10^6	2.25×10^6	2.44×10^6
9	2.07×10^6	2.16×10^6	2.26×10^6
10	2.16×10^6	2.16×10^6	2.10×10^6
11	2.20×10^6	2.20×10^6	2.07×10^6
12	2.27×10^6	2.40×10^6	2.39×10^6
mean		2.21×10^6	
u_A		3.67×10^4	
$u_{A \text{ rel}}$		1.67%	

Table S9. Components of the uncertainty of BPV plasmid RM

Components	value
$u_{A \text{ rel}}$	1.67%
$u_{B \text{ rel}}$	1.71%
$u_{\text{char rel}}$	2.39%
$u_{Its \text{ rel}}$	4.80%
$u_{\text{hom rel}}$	1.74%
$u_{\text{CRM rel}}$	5.64%
$U_{\text{rel}} = u_{\text{CRM rel}} \cdot k$	11.27%
Reference values (copies/ μ L)	2.21×10^6
Expanded uncertainty (copies/ μ L)	2.49×10^5

References

- ¹ Sniper. DQ24 All-in-one Automation Digital PCR System [Technical document]. Retrieved from <https://www.sniper-tech.com/software/Brochure.pdf>.
- ² BIO-RAD. QX200™ Droplet Reader and QX Manager Software Standard Edition [Technical document]. Retrieved from <https://www.bio-rad.com/sites/default/files/2025-02/10000107223.pdf>.
- ³ QIAGEN. QIAcuity® User Manual (Version 3.1) [Technical document]. Retrieved from <https://www.qiagen.com/us/resources/resourcedetail?id=948d868f-81ed-4f51-9692-149df3557f73&lang=en>.