

Supplementary Material

Expanding Luminol Activity: UV-Enhanced Peroxidase-Like Activity for Dual-Mode Wide-pH Biosensing

Mohamed Hassan Mahana ^a, Syed Rahin Ahmed ^b, Poushali Das ^a, Seshasai Srinivasan ^{*a,b} and Amin Reza Rajabzadeh ^{*a,b}

^a School of Biomedical Engineering, McMaster University, 1280 Main Street West Hamilton, Ontario, Canada

^b School of Engineering Practice and Technology, McMaster University, 1280 Main Street West Hamilton, Ontario, Canada

*E-mail address: SS:ssriniv@mcmaster.ca, ARR: rajaba@mcmaster.ca




Fig S1: Optimization of luminol concentration for H₂O₂ detection




Fig S2: Optimization of TMB concentration for H₂O₂ detection




Fig S3: Type of buffer for H₂O₂ detection




Fig S4: Time optimization for H₂O₂ detection

Table S1: Regression data and statistical parameters for detection of H₂O₂ using the developed colorimetric method.

| Parameter | value |
|--|-------------|
| Linearity range (mM) | (0.1-14 mM) |
| Coefficient of determination (r^2) | 0.9969 |
| LOD (mM) | 0.0149 |
| slope | 0.2261 |
| Intercept | 0.1086 |
| SE of slope (S_b) | 0.0056 |
| SE of intercept (S_a) | 0.0032 |
| SE of residuals ($S_{y/x}$) | 0.0052 |




Fig S5: Optimization of luminol concentration for dopamine detection




Fig S6: Optimization of TMB concentration for dopamine detection




Fig S7: Optimization of H₂O₂ concentration for dopamine detection

Table S2: Regression data and statistical parameters for detection of dopamine using the developed colorimetric method.

| Parameter | value |
|--|----------|
| Linearity range (μM) | (5–1000) |
| Coefficient of determination (r^2) | 0.9992 |
| LOD (μM) | 0.5197 |
| slope | 0.0108 |
| Intercept | 1.026 |
| SE of slope (S_b) | 0.0001 |
| SE of intercept (S_a) | 0.0002 |
| SE of residuals ($S_{y/x}$) | 0.0126 |

Table S3: Regression data and statistical parameters for determination of dopamine in simulated blood matrix using the developed colorimetric method.

| Parameter | value |
|--|--------------|
| Linearity range (μM) | (5–1000) |
| Coefficient of determination (r^2) | 0.9977 |
| LOD (μM) | 1.5089 |
| slope | 0.0109 |
| Intercept | 0.9685 |
| SE of slope (S_b) | 0.0003 |
| SE of intercept (S_a) | 0.0158 |
| SE of residuals ($S_{y/x}$) | 0.021 |