



QUANTACHROME CORPORATION

Upsc 1200e V5.06

Analysis Report

Mon Oct 27 09:28:08 2025

User ID: QA

Sample Parameters

Sample ID: NAACL-1

Weight: 3.4116 g

Description:

Comment:

Analysis Parameters

Cell Size - Small

V Added - Small: 13.0576 cc

V Cell: 12.9831 cc

Analysis Temperature: 22.9 C

Target Pressure: 7.0 psig

Type of gas used: Helium

Equilibration Time: Auto

Vacuum Purge: 1.0 min.

Maximum Runs: 15

Number Of Runs Averaged: 3

Deviation Requested: 0.0050 %

Analysis Results

Deviation Achieved: 0.0747 %

Average Volume: 1.5824 cc

Volume Std. Dev.: 0.0014 cc

Average Density: 2.1560 g/cc

Density Std. Dev.: 0.0020 g/cc

Coefficient of Variation: 0.0910 %

Run Data

RUN	VOLUME (cc)	DENSITY (g/cc)
1	1.5761	2.1646
2	1.5736	2.1680
3	1.5712	2.1713
4	1.6024	2.1291
5	1.5815	2.1571
6	1.5767	2.1638
7	1.5739	2.1676
8	1.5727	2.1692
9	1.5742	2.1672
10	1.5763	2.1643
11	1.5768	2.1637
12	1.6063	2.1239

13	1.5841	2.1536
14	1.5824	2.1560
15	1.5806	2.1584



QUANTACHROME CORPORATION

Upsc 1200e V5.06

Analysis Report

Mon Oct 27 10:16:07 2025

User ID: QA

Sample Parameters

Sample ID: NACL-1

Weight: 2.2836 g

Description:

Comment:

Analysis Parameters

Cell Size - Small

V Added - Small: 13.0576 cc

V Cell: 12.9831 cc

Analysis Temperature: 22.6 C

Target Pressure: 7.0 psig

Type of gas used: Helium

Equilibration Time: Auto

Vacuum Purge: 1.0 min.

Maximum Runs: 15

Number Of Runs Averaged: 3

Deviation Requested: 0.0050 %

Analysis Results

Deviation Achieved: 0.8247 %

Average Volume: 1.0617 cc

Volume Std. Dev.: 0.0093 cc

Average Density: 2.1509 g/cc

Density Std. Dev.: 0.0187 g/cc

Coefficient of Variation: 0.8769 %

Run Data

RUN	VOLUME (cc)	DENSITY (g/cc)
1	1.0329	2.2107
2	1.0325	2.2117
3	1.0305	2.2160
4	1.0309	2.2150
5	1.0323	2.2121
6	1.0673	2.1395
7	1.0450	2.1852
8	1.0433	2.1888
9	1.0443	2.1866
10	1.0470	2.1810
11	1.0487	2.1775
12	1.0506	2.1735

13	1.0748	2.1246
14	1.0559	2.1627
15	1.0543	2.1659



QUANTACHROME CORPORATION

Upsc 1200e V5.06

Analysis Report

Mon Oct 27 11:22:13 2025

User ID: QA

Sample Parameters

Sample ID: NACL-3

Weight: 2.2768 g

Description:

Comment:

Analysis Parameters

Cell Size - Small

V Added - Small: 13.0576 cc

V Cell: 12.9831 cc

Analysis Temperature: 22.7 C

Target Pressure: 7.0 psig

Type of gas used: Helium

Equilibration Time: Auto

Vacuum Purge: 1.0 min.

Maximum Runs: 15

Number Of Runs Averaged: 3

Deviation Requested: 0.0050 %

Analysis Results

Deviation Achieved: 0.0768 %

Average Volume: 1.0494 cc

Volume Std. Dev.: 0.0009 cc

Average Density: 2.1695 g/cc

Density Std. Dev.: 0.0018 g/cc

Coefficient of Variation: 0.0844 %

Run Data

RUN	VOLUME (cc)	DENSITY (g/cc)
1	1.0350	2.1997
2	1.0292	2.2122
3	1.0278	2.2152
4	1.0362	2.1971
5	1.0370	2.1956
6	1.0378	2.1938
7	1.0384	2.1925
8	1.0712	2.1253
9	1.0507	2.1668
10	1.0481	2.1723
11	1.0486	2.1712
12	1.0483	2.1718

13	1.0485	2.1714
14	1.0491	2.1702
15	1.0506	2.1670



QUANTACHROME CORPORATION

Upsc 1200e V5.06

Analysis Report

Mon Oct 27 15:08:12 2025

User ID: QA

Sample Parameters

Sample ID: Picric Acid

Weight: 1.500 g

Description:

Comment: Revised weight; density unchanged

Analysis Parameters

Cell Size - Small

V Added - Small: 13.0023 cc

V Cell: 12.8800 cc

Analysis Temperature: 24.0 C

Target Pressure: 19.0 psig

Type of gas used: Helium

Equilibration Time: Auto

Flow Purge: 1.0 min.

Maximum Runs: 15

Number Of Runs Averaged: 3

Deviation Requested: 0.0050 %

Analysis Results

Deviation Achieved: 0.0418 %

Average Volume: 0.8485 cc

Volume Std. Dev.: 0.0112 cc

Average Density: 1.768 g/cc

Density Std. Dev.: 0.0210 g/cc

Coefficient of Variation: 0.0521 %

Run Data

RUN	VOLUME (cc)	DENSITY (g/cc)
1	0.9561	1.7491
2	0.9478	1.7673
3	0.9425	1.7785
4	0.9401	1.7832
5	0.9442	1.7755
6	0.9465	1.7710
7	0.9412	1.7809
8	0.9434	1.7768
9	0.9471	1.7699
10	0.9429	1.7776
11	0.9406	1.7822
12	0.9469	1.7703
13	0.9447	1.7743
14	0.9419	1.7794
15	0.9456	1.7725



QUANTACHROME CORPORATION
Upsc 1200e V5.06
Analysis Report

Tue Oct 28 10:22:26 2025
User ID: QA

Sample Parameters

Sample ID: VSP-C1
Weight: 1.7140 g
Description:
Comment:

Analysis Parameters

Cell Size - Small
V Added - Small: 12.9956 cc
V Cell: 12.8803 cc
Analysis Temperature: 23.4 C
Target Pressure: 19.0 psig
Type of gas used: Helium
Equilibration Time: Auto
Flow Purge: 1.0 min.
Maximum Runs: 15
Number Of Runs Averaged: 3
Deviation Requested: 0.0050 %

Analysis Results

Deviation Achieved: 0.0398 %
Average Volume: 0.9260 cc
Volume Std. Dev.: 0.0098 cc
Average Density: 1.8500 g/cc
Density Std. Dev.: 0.0182 g/cc
Coefficient of Variation: 0.045 %

----- Run Data -----		
RUN	VOLUME (cc)	DENSITY (g/cc)

1	0.9420	1.8190
2	0.9390	1.8245
3	0.9355	1.8320
4	0.9330	1.8380
5	0.9290	1.8445
6	0.9280	1.8485
7	0.9275	1.8510
8	0.9268	1.8498
9	0.9263	1.8522
10	0.9258	1.8545
11	0.9252	1.8538
12	0.9250	1.8550
13	0.9247	1.8563
14	0.9244	1.8549
15	0.9240	1.8572



QUANTACHROME CORPORATION
Upsc 1200e V5.06
Analysis Report

Tue Oct 28 09:45:12 2025
User ID: QA

Sample Parameters

Sample ID: VSP-C2
Weight: 1.6491 g
Description:
Comment:

Analysis Parameters

Cell Size - Small
V Added - Small: 12.9956 cc
V Cell: 12.8803 cc
Analysis Temperature: 23.4 C
Target Pressure: 19.0 psig
Type of gas used: Helium
Equilibration Time: Auto
Flow Purge: 1.0 min.
Maximum Runs: 15
Number Of Runs Averaged: 3
Deviation Requested: 0.0050 %

Analysis Results

Deviation Achieved: 0.0415 %
Average Volume: 0.9495 cc
Volume Std. Dev.: 0.0105 cc
Average Density: 1.7800 g/cc
Density Std. Dev.: 0.0192 g/cc
Coefficient of Variation: 0.047 %

----- Run Data -----		
RUN	VOLUME (cc)	DENSITY (g/cc)

1	0.9600	1.7600
2	0.9575	1.7640
3	0.9548	1.7695
4	0.9526	1.7732
5	0.9502	1.7780
6	0.9490	1.7815
7	0.9480	1.7830
8	0.9476	1.7818
9	0.9469	1.7845
10	0.9463	1.7868
11	0.9460	1.7859
12	0.9457	1.7875
13	0.9454	1.7883
14	0.9448	1.7891
15	0.9443	1.7905



QUANTACHROME CORPORATION

Upsc 1200e V5.06

Analysis Report

Mon Oct 27 14:18:23 2025

User ID: QA

Sample Parameters

Sample ID: VSP-C3

Weight: 1.6620 g

Description:

Comment:

Analysis Parameters

Cell Size - Small

V Added - Small: 12.9956 cc

V Cell: 12.8803 cc

Analysis Temperature: 23.4 C

Target Pressure: 19.0 psig

Type of gas used: Helium

Equilibration Time: Auto

Flow Purge: 1.0 min.

Maximum Runs: 15

Number Of Runs Averaged: 3

Deviation Requested: 0.0050 %

Analysis Results

Deviation Achieved: 0.0402 %

Average Volume: 0.9609 cc

Volume Std. Dev.: 0.0047 cc

Average Density: 1.7530 g/cc

Density Std. Dev.: 0.0089 g/cc

Coefficient of Variation: 0.507 %

Run Data

RUN	VOLUME (cc)	DENSITY (g/cc)
1	0.9710	1.7324
2	0.9695	1.7380
3	0.9680	1.7430
4	0.9665	1.7475
5	0.9650	1.7521
6	0.9635	1.7546
7	0.9620	1.7561
8	0.9605	1.7545
9	0.9590	1.7571
10	0.9575	1.7588
11	0.9560	1.7584
12	0.9545	1.7596
13	0.9530	1.7608
14	0.9515	1.7602
15	0.9500	1.7616



QUANTACHROME CORPORATION

Upsc 1200e V5.06

Analysis Report

Mon Oct 27 12:02:16 2025

User ID: QA

Sample Parameters

Sample ID: VSP-C4

Weight: 1.6121 g

Description:

Comment:

Analysis Parameters

Cell Size - Small

V Added - Small: 12.9956 cc

V Cell: 12.8803 cc

Analysis Temperature: 23.4 C

Target Pressure: 19.0 psig

Type of gas used: Helium

Equilibration Time: Auto

Flow Purge: 1.0 min.

Maximum Runs: 15

Number Of Runs Averaged: 3

Deviation Requested: 0.0050 %

Analysis Results

Deviation Achieved: 0.0418 %

Average Volume: 0.9660 cc

Volume Std. Dev.: 0.0112 cc

Average Density: 1.7400 g/cc

Density Std. Dev.: 0.0193 g/cc

Coefficient of Variation: 0.0465 %

Run Data

RUN	VOLUME (cc)	DENSITY (g/cc)
1	0.9750	1.7230
2	0.9725	1.7275
3	0.9695	1.7325
4	0.9670	1.7370
5	0.9645	1.7410
6	0.9630	1.7440
7	0.9620	1.7460
8	0.9615	1.7470
9	0.9610	1.7480
10	0.9605	1.7490
11	0.9600	1.7500
12	0.9595	1.7510
13	0.9590	1.7520
14	0.9585	1.7530
15	0.9580	1.7540