

Table S3. Description of HbA2 analytical for Variant II Hemoglobin Testing System.

Parameters		Measured value
Precision	Intra-assay precision	1.46%
	Inter-assay precision	1.23%
Accuracy	Low value	-1.48%
	High value	0.54%
Linearity		0.8%-10.20%
Carryover		0.00%