

Electronic Supplementary Material

Table S1 Results of artificially contaminated samples detection

Number	Addition of <i>H. pylori</i> (CFU/mL)	<i>H. pylori</i> found (CFU/mL)	Recovery rate (%)	RSD (n=3, %)
1	1	0.984	98.4	3.7
2	100	104.00	104.0	4.0
3	10000	10491.97	104.9	4.6