

Cdate	Method	QC_Measure	Result	RL	Units	Samp1_Conc	Samp2_Conc	Recovery	LowerConc	UpperConc
15/04/2002	EPA 415.1 (T) Ox	RPD - Check Standard	0.7258		%	9.61	9.68		0	0.15
15/04/2002	EPA 415.1 (T) Ox	Method Blank	0.1		0.1 mg/L as C	0	0			
15/04/2002	EPA 415.1 (T) Ox	Method Blank	0.1		0.1 mg/L as C	0	0			
15/04/2002	EPA 415.1 (T) Ox	Check Standard	9.61		0.1 mg/L as C	0	0	0.961	0.85	1.15
15/04/2002	EPA 415.1 (T) Ox	Check Standard	9.68		0.1 mg/L as C	0	0	0.968	0.85	1.15
15/04/2002	EPA 415.1 (T) Ox	Check Standard	9.61		0.1 mg/L as C	0	0	0.961	0.85	1.15
15/04/2002	EPA 415.1 (T) Ox	Check Standard	9.68		0.1 mg/L as C	0	0	0.968	0.85	1.15
15/04/2002	EPA 415.1 (T) Ox	Low Level Standard	0.56		0.1 mg/L as C	0	0	1.12	0.5	1.5
15/04/2002	EPA 415.1 (T) Ox	Low Level Standard	1.09		0.1 mg/L as C	0	0	1.09	0.5	1.5
15/04/2002	EPA 415.1 (T) Ox	Low Level Standard	1.09		0.1 mg/L as C	0	0	1.09	0.5	1.5
15/04/2002	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.6742		%	25.3	25.13		0	0.15
15/04/2002	EPA 415.1 (T) Ox	RPD - Check Standard	0.7258		%	9.61	9.68		0	0.15
15/04/2002	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.6742		%	25.3	25.13		0	0.15
15/04/2002	EPA 415.1 (T) Ox	Low Level Standard	0.56		0.1 mg/L as C	0	0	1.12	0.5	1.5
22/04/2002	EPA 415.1 (T) Ox	Check Standard	9.7		0.1 mg/L as C	0	0	0.97	0.85	1.15
22/04/2002	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.2504		%	23.99	23.93		0	0.15
22/04/2002	EPA 415.1 (T) Ox	RPD - Check Standard	1.558		%	9.55	9.7		0	0.15
22/04/2002	EPA 415.1 (T) Ox	Low Level Standard	0.57		0.1 mg/L as C	0	0	1.14	0.5	1.5
22/04/2002	EPA 415.1 (T) Ox	Check Standard	9.55		0.1 mg/L as C	0	0	0.955	0.85	1.15
22/04/2002	EPA 415.1 (T) Ox	Method Blank	0.11		0.1 mg/L as C	0	0			
22/04/2002	EPA 415.1 (T) Ox	Low Level Standard	1.04		0.1 mg/L as C	0	0	1.04	0.5	1.5
29/04/2002	EPA 415.1 (T) Ox	Check Standard	9.63		0.1 mg/L as C	0	0	0.963	0.85	1.15
29/04/2002	EPA 415.1 (T) Ox	Check Standard	9.6		0.1 mg/L as C	0	0	0.96	0.85	1.15
29/04/2002	EPA 415.1 (T) Ox	RPD - Check Standard	0.312		%	9.6	9.63		0	0.15
29/04/2002	EPA 415.1 (T) Ox	RPD - Check Standard	5.941		%	0.98	1.04		0	0.15
29/04/2002	EPA 415.1 (T) Ox	Check Standard	1.04		0.1 mg/L as C	0	0	1.04	0.85	1.15
29/04/2002	EPA 415.1 (T) Ox	Check Standard	0.98		0.1 mg/L as C	0	0	0.98	0.85	1.15
29/04/2002	EPA 415.1 (T) Ox	Low Level Standard	0.56		0.1 mg/L as C	0	0	1.12	0.5	1.5
29/04/2002	EPA 415.1 (T) Ox	Method Blank	0.05		0.1 mg/L as C	0	0			
29/04/2002	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.6159		%	24.43	24.28		0	0.15
06/05/2002	EPA 415.1 (T) Ox	Low Level Standard	0.55		0.1 mg/L as C	0	0	1.1	0.5	1.5
06/05/2002	EPA 415.1 (T) Ox	Low Level Standard	0.55		0.1 mg/L as C	0	0	1.1	0.5	1.5
06/05/2002	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.416		%	21.68	21.59		0	0.15
06/05/2002	EPA 415.1 (T) Ox	RPD - Check Standard	2.642		%	8.59	8.82		0	0.15
06/05/2002	EPA 415.1 (T) Ox	RPD - Check Standard	2.642		%	8.59	8.82		0	0.15
06/05/2002	EPA 415.1 (T) Ox	Low Level Standard	0.55		0.1 mg/L as C	0	0	1.1	0.5	1.5

06/05/2002	EPA 415.1 (T) Ox	Check Standard	0.98	0.1 mg/L as C	0	0	0.98	0.85	1.15
06/05/2002	EPA 415.1 (T) Ox	Check Standard	8.59	0.1 mg/L as C	0	0	0.859	0.85	1.15
06/05/2002	EPA 415.1 (T) Ox	Check Standard	8.82	0.1 mg/L as C	0	0	0.882	0.85	1.15
06/05/2002	EPA 415.1 (T) Ox	Check Standard	0.98	0.1 mg/L as C	0	0	0.98	0.85	1.15
06/05/2002	EPA 415.1 (T) Ox	Check Standard	8.59	0.1 mg/L as C	0	0	0.859	0.85	1.15
06/05/2002	EPA 415.1 (T) Ox	Check Standard	8.82	0.1 mg/L as C	0	0	0.882	0.85	1.15
06/05/2002	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.416	%	21.68	21.59		0	0.15
06/05/2002	EPA 415.1 (T) Ox	Method Blank	0.09	0.1 mg/L as C	0	0			
06/05/2002	EPA 415.1 (T) Ox	Method Blank	0.09	0.1 mg/L as C	0	0			
06/05/2002	EPA 415.1 (T) Ox	Low Level Standard	0.55	0.1 mg/L as C	0	0	1.1	0.5	1.5
13/05/2002	EPA 415.1 (T) Ox	RPD - Check Standard	3.855	%	8.65	8.99		0	0.15
13/05/2002	EPA 415.1 (T) Ox	Check Standard	8.65	0.1 mg/L as C	0	0	0.865	0.85	1.15
13/05/2002	EPA 415.1 (T) Ox	Check Standard	8.99	0.1 mg/L as C	0	0	0.899	0.85	1.15
13/05/2002	EPA 415.1 (T) Ox	Check Standard	0.94	0.1 mg/L as C	0	0	0.94	0.85	1.15
13/05/2002	EPA 415.1 (T) Ox	Low Level Standard	0.53	0.1 mg/L as C	0	0	1.06	0.5	1.5
13/05/2002	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.1781	%	22.48	22.44		0	0.15
13/05/2002	EPA 415.1 (T) Ox	RPD - Check Standard	8.163	%	0.94	1.02		0	0.15
13/05/2002	EPA 415.1 (T) Ox	Check Standard	1.02	0.1 mg/L as C	0	0	1.02	0.85	1.15
13/05/2002	EPA 415.1 (T) Ox	Method Blank	0.09	0.1 mg/L as C	0	0			
20/05/2002	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.1781	%	22.48	22.44		0	0.15
20/05/2002	EPA 415.1 (T) Ox	Method Blank	0.09	0.1 mg/L as C	0	0			
20/05/2002	EPA 415.1 (T) Ox	RPD - Check Standard	3.855	%	8.65	8.99		0	0.15
20/05/2002	EPA 415.1 (T) Ox	Low Level Standard	0.53	0.1 mg/L as C	0	0	1.06	0.5	1.5
20/05/2002	EPA 415.1 (T) Ox	Check Standard	1.02	0.1 mg/L as C	0	0	1.02	0.85	1.15
20/05/2002	EPA 415.1 (T) Ox	Check Standard	0.94	0.1 mg/L as C	0	0	0.94	0.85	1.15
20/05/2002	EPA 415.1 (T) Ox	RPD - Check Standard	8.163	%	0.94	1.02		0	0.15
20/05/2002	EPA 415.1 (T) Ox	Check Standard	8.99	0.1 mg/L as C	0	0	0.899	0.85	1.15
20/05/2002	EPA 415.1 (T) Ox	Check Standard	8.65	0.1 mg/L as C	0	0	0.865	0.85	1.15
28/05/2002	EPA 415.1 (T) Ox	Check Standard	1	0.1 mg/L as C	0	0	1	0.85	1.15
28/05/2002	EPA 415.1 (T) Ox	Check Standard	8.89	0.1 mg/L as C	0	0	0.889	0.85	1.15
28/05/2002	EPA 415.1 (T) Ox	Check Standard	1	0.1 mg/L as C	0	0	1	0.85	1.15
28/05/2002	EPA 415.1 (T) Ox	Check Standard	0.91	0.1 mg/L as C	0	0	0.91	0.85	1.15
28/05/2002	EPA 415.1 (T) Ox	Check Standard	8.71	0.1 mg/L as C	0	0	0.871	0.85	1.15
28/05/2002	EPA 415.1 (T) Ox	Check Standard	8.89	0.1 mg/L as C	0	0	0.889	0.85	1.15
28/05/2002	EPA 415.1 (T) Ox	Check Standard	8.71	0.1 mg/L as C	0	0	0.871	0.85	1.15
28/05/2002	EPA 415.1 (T) Ox	Check Standard	0.91	0.1 mg/L as C	0	0	0.91	0.85	1.15
28/05/2002	EPA 415.1 (T) Ox	Low Level Standard	0.54	0.1 mg/L as C	0	0	1.08	0.5	1.5

28/05/2002	EPA 415.1 (T) Ox	RPD - Check Standard	2.045	%	8.89	8.71	0	0.15	
28/05/2002	EPA 415.1 (T) Ox	Method Blank	0.08	0.1 mg/L as C	0	0			
28/05/2002	EPA 415.1 (T) Ox	Method Blank	0.08	0.1 mg/L as C	0	0			
28/05/2002	EPA 415.1 (T) Ox	Low Level Standard	0.54	0.1 mg/L as C	0	0	1.08	0.5	1.5
28/05/2002	EPA 415.1 (T) Ox	RPD - Check Standard	9.424	%	1	0.91	0	0.15	
28/05/2002	EPA 415.1 (T) Ox	RPD - Check Standard	2.045	%	8.89	8.71	0	0.15	
28/05/2002	EPA 415.1 (T) Ox	RPD - LCS Duplicate	6.636	%	22.29	23.82	0	0.15	
28/05/2002	EPA 415.1 (T) Ox	RPD - LCS Duplicate	6.636	%	22.29	23.82	0	0.15	
28/05/2002	EPA 415.1 (T) Ox	RPD - Check Standard	9.424	%	1	0.91	0	0.15	
03/06/2002	EPA 415.1 (T) Ox	RPD - LCS Duplicate	6.636	%	22.29	23.82	0	0.15	
03/06/2002	EPA 415.1 (T) Ox	RPD - Check Standard	9.424	%	1	0.91	0	0.15	
03/06/2002	EPA 415.1 (T) Ox	Low Level Standard	0.54	0.1 mg/L as C	0	0	1.08	0.5	1.5
03/06/2002	EPA 415.1 (T) Ox	Method Blank	0.08	0.1 mg/L as C	0	0			
03/06/2002	EPA 415.1 (T) Ox	RPD - Check Standard	2.045	%	8.89	8.71	0	0.15	
03/06/2002	EPA 415.1 (T) Ox	Check Standard	1	0.1 mg/L as C	0	0	1	0.85	1.15
03/06/2002	EPA 415.1 (T) Ox	Check Standard	0.91	0.1 mg/L as C	0	0	0.91	0.85	1.15
03/06/2002	EPA 415.1 (T) Ox	Check Standard	8.89	0.1 mg/L as C	0	0	0.889	0.85	1.15
03/06/2002	EPA 415.1 (T) Ox	Check Standard	8.71	0.1 mg/L as C	0	0	0.871	0.85	1.15
10/06/2002	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.7135	%	23.91	23.74	0	0.15	
10/06/2002	EPA 415.1 (T) Ox	Check Standard	10.3	0.1 mg/L as C	0	0	1.03	0.85	1.15
10/06/2002	EPA 415.1 (T) Ox	Low Level Standard	0.62	0.1 mg/L as C	0	0	1.24	0.5	1.5
10/06/2002	EPA 415.1 (T) Ox	Check Standard	1.06	0.1 mg/L as C	0	0	1.06	0.85	1.15
10/06/2002	EPA 415.1 (T) Ox	RPD - Check Standard	6.413	%	9.66	10.3	0	0.15	
10/06/2002	EPA 415.1 (T) Ox	Check Standard	9.66	0.1 mg/L as C	0	0	0.966	0.85	1.15
10/06/2002	EPA 415.1 (T) Ox	Method Blank	0.09	0.1 mg/L as C	0	0			
17/06/2002	EPA 415.1 (T) Ox	Low Level Standard	0.58	0.1 mg/L as C	0	0	1.16	0.5	1.5
17/06/2002	EPA 415.1 (T) Ox	Low Level Standard	0.58	0.1 mg/L as C	0	0	1.16	0.5	1.5
17/06/2002	EPA 415.1 (T) Ox	Method Blank	0	0.1 mg/L as C	0	0			
17/06/2002	EPA 415.1 (T) Ox	Method Blank	0	0.1 mg/L as C	0	0			
17/06/2002	EPA 415.1 (T) Ox	Check Standard	10.2	0.1 mg/L as C	0	0	1.02	0.85	1.15
17/06/2002	EPA 415.1 (T) Ox	RPD - Check Standard	6.165	%	9.59	10.2	0	0.15	
17/06/2002	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0	%	24.29	24.29	0	0.15	
17/06/2002	EPA 415.1 (T) Ox	RPD - Check Standard	6.165	%	9.59	10.2	0	0.15	
17/06/2002	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0	%	24.29	24.29	0	0.15	
17/06/2002	EPA 415.1 (T) Ox	Check Standard	1.09	0.1 mg/L as C	0	0	1.09	0.85	1.15
17/06/2002	EPA 415.1 (T) Ox	Check Standard	9.59	0.1 mg/L as C	0	0	0.959	0.85	1.15
17/06/2002	EPA 415.1 (T) Ox	Check Standard	1.09	0.1 mg/L as C	0	0	1.09	0.85	1.15

17/06/2002	EPA 415.1 (T) Ox	Check Standard	10.2	0.1 mg/L as C	0	0	1.02	0.85	1.15
17/06/2002	EPA 415.1 (T) Ox	Check Standard	9.59	0.1 mg/L as C	0	0	0.959	0.85	1.15
01/07/2002	EPA 415.1 (T) Ox	Low Level Standard	0.53	0.1 mg/L as C	0	0	1.06	0.5	1.5
01/07/2002	EPA 415.1 (T) Ox	RPD - Check Standard	0	%	9.79	9.79		0	0.15
01/07/2002	EPA 415.1 (T) Ox	Check Standard	9.79	0.1 mg/L as C	0	0	0.979	0.85	1.15
01/07/2002	EPA 415.1 (T) Ox	Method Blank	0.07	0.1 mg/L as C	0	0			
01/07/2002	EPA 415.1 (T) Ox	Check Standard	9.79	0.1 mg/L as C	0	0	0.979	0.85	1.15
01/07/2002	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.332	%	24.14	24.06		0	0.15
01/07/2002	EPA 415.1 (T) Ox	Check Standard	1.06	0.1 mg/L as C	0	0	1.06	0.85	1.15
08/07/2002	EPA 415.1 (T) Ox	Check Standard	0.88	0.1 mg/L as C	0	0	0.88	0.85	1.15
08/07/2002	EPA 415.1 (T) Ox	Check Standard	1.1	0.1 mg/L as C	0	0	1.1	0.85	1.15
08/07/2002	EPA 415.1 (T) Ox	Low Level Standard	0.43	0.1 mg/L as C	0	0	0.86	0.5	1.5
08/07/2002	EPA 415.1 (T) Ox	Check Standard	9.4	0.1 mg/L as C	0	0	0.94	0.85	1.15
08/07/2002	EPA 415.1 (T) Ox	Check Standard	9.79	0.1 mg/L as C	0	0	0.979	0.85	1.15
08/07/2002	EPA 415.1 (T) Ox	RPD - Check Standard	4.065	%	9.4	9.79		0	0.15
08/07/2002	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.04201	%	23.81	23.8		0	0.15
08/07/2002	EPA 415.1 (T) Ox	Method Blank	0.05	0.1 mg/L as C	0	0			
15/07/2002	EPA 415.1 (T) Ox	Check Standard	9.46	0.1 mg/L as C	0	0	0.946	0.85	1.15
15/07/2002	EPA 415.1 (T) Ox	Method Blank	0.06	0.1 mg/L as C	0	0			
15/07/2002	EPA 415.1 (T) Ox	Check Standard	9.41	0.1 mg/L as C	0	0	0.941	0.85	1.15
15/07/2002	EPA 415.1 (T) Ox	Check Standard	1.01	0.1 mg/L as C	0	0	1.01	0.85	1.15
15/07/2002	EPA 415.1 (T) Ox	Check Standard	1	0.1 mg/L as C	0	0	1	0.85	1.15
15/07/2002	EPA 415.1 (T) Ox	Check Standard	9.46	0.1 mg/L as C	0	0	0.946	0.85	1.15
15/07/2002	EPA 415.1 (T) Ox	Check Standard	9.41	0.1 mg/L as C	0	0	0.941	0.85	1.15
15/07/2002	EPA 415.1 (T) Ox	Check Standard	1.01	0.1 mg/L as C	0	0	1.01	0.85	1.15
15/07/2002	EPA 415.1 (T) Ox	Check Standard	1	0.1 mg/L as C	0	0	1	0.85	1.15
15/07/2002	EPA 415.1 (T) Ox	Low Level Standard	0.56	0.1 mg/L as C	0	0	1.12	0.5	1.5
15/07/2002	EPA 415.1 (T) Ox	Low Level Standard	0.56	0.1 mg/L as C	0	0	1.12	0.5	1.5
15/07/2002	EPA 415.1 (T) Ox	RPD - LCS Duplicate	1.158	%	23.45	23.18		0	0.15
15/07/2002	EPA 415.1 (T) Ox	RPD - Check Standard	0.5299	%	9.41	9.46		0	0.15
15/07/2002	EPA 415.1 (T) Ox	RPD - Check Standard	0.995	%	1	1.01		0	0.15
15/07/2002	EPA 415.1 (T) Ox	RPD - LCS Duplicate	1.158	%	23.45	23.18		0	0.15
15/07/2002	EPA 415.1 (T) Ox	RPD - Check Standard	0.995	%	1	1.01		0	0.15
15/07/2002	EPA 415.1 (T) Ox	RPD - Check Standard	0.5299	%	9.41	9.46		0	0.15
15/07/2002	EPA 415.1 (T) Ox	Method Blank	0.06	0.1 mg/L as C	0	0			
22/07/2002	EPA 415.1 (T) Ox	Check Standard	9.68	0.1 mg/L as C			0.968	0.85	1.15
22/07/2002	EPA 415.1 (T) Ox	Check Standard	1.01	0.1 mg/L as C			1.01	0.85	1.15

22/07/2002	EPA 415.1 (T) Ox	Low Level Standard	0.55	0.1 mg/L as C	0	0	1.1	0.5	1.5
22/07/2002	EPA 415.1 (T) Ox	Low Level Standard	0.57	0.1 mg/L as C			1.14	0.5	1.5
22/07/2002	EPA 415.1 (T) Ox	Low Level Standard	0.57	0.1 mg/L as C			1.14	0.5	1.5
22/07/2002	EPA 415.1 (T) Ox	Check Standard	9.68	0.1 mg/L as C			0.968	0.85	1.15
22/07/2002	EPA 415.1 (T) Ox	RPD - Internal Duplicate	0	0.1 %	1.4	1.4		0	0.3
22/07/2002	EPA 415.1 (T) Ox	RPD - Check Standard	1.13	%	9.68	9.79		0	0.15
22/07/2002	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.2066	%	24.23	24.18		0	0.15
22/07/2002	EPA 415.1 (T) Ox	RPD - Check Standard	1.13	%	9.68	9.79		0	0.15
22/07/2002	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.2066	%	24.23	24.18		0	0.15
22/07/2002	EPA 415.1 (T) Ox	RPD - Check Standard	1.13	%	9.68	9.79		0	0.15
22/07/2002	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.2066	%	24.23	24.18		0	0.15
22/07/2002	EPA 415.1 (T) Ox	RPD - LCS Duplicate	1.522	%	23.48	23.84		0	0.15
22/07/2002	EPA 415.1 (T) Ox	RPD - Check Standard	1.798	%	9.37	9.54		0	0.15
22/07/2002	EPA 415.1 (T) Ox	Low Level Standard	0.57	0.1 mg/L as C			1.14	0.5	1.5
22/07/2002	EPA 415.1 (T) Ox	RPD - Check Standard	0.995	%	1.01	1		0	0.15
22/07/2002	EPA 415.1 (T) Ox	Method Blank	0	0.1 mg/L as C	0	0			
22/07/2002	EPA 415.1 (T) Ox	Check Standard	9.79	0.1 mg/L as C			0.979	0.85	1.15
22/07/2002	EPA 415.1 (T) Ox	Check Standard	9.54	0.1 mg/L as C	0	0	0.954	0.85	1.15
22/07/2002	EPA 415.1 (T) Ox	Check Standard	1.01	0.1 mg/L as C			1.01	0.85	1.15
22/07/2002	EPA 415.1 (T) Ox	Check Standard	9.37	0.1 mg/L as C	0	0	0.937	0.85	1.15
22/07/2002	EPA 415.1 (T) Ox	Method Blank	0.08	0.1 mg/L as C					
22/07/2002	EPA 415.1 (T) Ox	Method Blank	0.08	0.1 mg/L as C					
22/07/2002	EPA 415.1 (T) Ox	Check Standard	1	0.1 mg/L as C	0	0	1	0.85	1.15
22/07/2002	EPA 415.1 (T) Ox	Check Standard	9.79	0.1 mg/L as C			0.979	0.85	1.15
22/07/2002	EPA 415.1 (T) Ox	Check Standard	9.68	0.1 mg/L as C			0.968	0.85	1.15
22/07/2002	EPA 415.1 (T) Ox	Check Standard	1.01	0.1 mg/L as C			1.01	0.85	1.15
22/07/2002	EPA 415.1 (T) Ox	Check Standard	9.79	0.1 mg/L as C			0.979	0.85	1.15
22/07/2002	EPA 415.1 (T) Ox	Method Blank	0.08	0.1 mg/L as C					
22/07/2002	EPA 415.1 (T) Ox	Check Standard	1.01	0.1 mg/L as C	0	0	1.01	0.85	1.15
29/07/2002	EPA 415.1 (T) Ox	Low Level Standard	0.51	0.1 mg/L as C	0	0	1.02	0.5	1.5
29/07/2002	EPA 415.1 (T) Ox	Method Blank	0.02	0.1 mg/L as C	0	0			
29/07/2002	EPA 415.1 (T) Ox	RPD - Check Standard	1.98	%	1.02	1		0	0.15
29/07/2002	EPA 415.1 (T) Ox	Low Level Standard	0.51	0.1 mg/L as C	0	0	1.02	0.5	1.5
29/07/2002	EPA 415.1 (T) Ox	RPD - Check Standard	1.98	%	1.02	1		0	0.15
29/07/2002	EPA 415.1 (T) Ox	RPD - Check Standard	1.72	%	9.22	9.38		0	0.15
29/07/2002	EPA 415.1 (T) Ox	RPD - LCS Duplicate	1.073	%	24.35	24.09		0	0.15
29/07/2002	EPA 415.1 (T) Ox	RPD - LCS Duplicate	1.073	%	24.35	24.09		0	0.15

29/07/2002	EPA 415.1 (T) Ox	Method Blank	0.02	0.1 mg/L as C	0	0			
29/07/2002	EPA 415.1 (T) Ox	Check Standard	9.22	0.1 mg/L as C	0	0	0.922	0.85	1.15
29/07/2002	EPA 415.1 (T) Ox	RPD - Check Standard	1.72	%	9.22	9.38		0	0.15
29/07/2002	EPA 415.1 (T) Ox	Check Standard	9.38	0.1 mg/L as C	0	0	0.938	0.85	1.15
29/07/2002	EPA 415.1 (T) Ox	Check Standard	1	0.1 mg/L as C	0	0	1	0.85	1.15
29/07/2002	EPA 415.1 (T) Ox	Check Standard	1.02	0.1 mg/L as C	0	0	1.02	0.85	1.15
29/07/2002	EPA 415.1 (T) Ox	Check Standard	9.38	0.1 mg/L as C	0	0	0.938	0.85	1.15
29/07/2002	EPA 415.1 (T) Ox	Check Standard	9.22	0.1 mg/L as C	0	0	0.922	0.85	1.15
29/07/2002	EPA 415.1 (T) Ox	Check Standard	1	0.1 mg/L as C	0	0	1	0.85	1.15
29/07/2002	EPA 415.1 (T) Ox	Check Standard	1.02	0.1 mg/L as C	0	0	1.02	0.85	1.15
06/08/2002	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.5364	%	24.3	24.17		0	0.15
06/08/2002	EPA 415.1 (T) Ox	RPD - Check Standard	2.725	%	9.41	9.67		0	0.15
06/08/2002	EPA 415.1 (T) Ox	Low Level Standard	1.03	0.1 mg/L as C	0	0	1.03	0.5	1.5
06/08/2002	EPA 415.1 (T) Ox	Method Blank	0.07	0.1 mg/L as C	0	0			
06/08/2002	EPA 415.1 (T) Ox	Check Standard	9.41	0.1 mg/L as C	0	0	0.941	0.85	1.15
06/08/2002	EPA 415.1 (T) Ox	Check Standard	9.67	0.1 mg/L as C	0	0	0.967	0.85	1.15
06/08/2002	EPA 415.1 (T) Ox	Low Level Standard	0.65	0.1 mg/L as C	0	0	1.3	0.5	1.5
12/08/2002	EPA 415.1 (T) Ox	Method Blank	0.07	0.1 mg/L as C	0	0			
12/08/2002	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.5364	%	24.3	24.17		0	0.15
12/08/2002	EPA 415.1 (T) Ox	Low Level Standard	1.03	0.1 mg/L as C	0	0	1.03	0.5	1.5
12/08/2002	EPA 415.1 (T) Ox	Low Level Standard	1.03	0.1 mg/L as C	0	0	1.03	0.5	1.5
12/08/2002	EPA 415.1 (T) Ox	Low Level Standard	0.65	0.1 mg/L as C	0	0	1.3	0.5	1.5
12/08/2002	EPA 415.1 (T) Ox	Low Level Standard	1.03	0.1 mg/L as C	0	0	1.03	0.5	1.5
12/08/2002	EPA 415.1 (T) Ox	Method Blank	0.07	0.1 mg/L as C	0	0			
12/08/2002	EPA 415.1 (T) Ox	Low Level Standard	0.65	0.1 mg/L as C	0	0	1.3	0.5	1.5
12/08/2002	EPA 415.1 (T) Ox	Check Standard	9.41	0.1 mg/L as C	0	0	0.941	0.85	1.15
12/08/2002	EPA 415.1 (T) Ox	Check Standard	9.67	0.1 mg/L as C	0	0	0.967	0.85	1.15
12/08/2002	EPA 415.1 (T) Ox	Check Standard	9.67	0.1 mg/L as C	0	0	0.967	0.85	1.15
12/08/2002	EPA 415.1 (T) Ox	Check Standard	9.41	0.1 mg/L as C	0	0	0.941	0.85	1.15
12/08/2002	EPA 415.1 (T) Ox	Check Standard	9.67	0.1 mg/L as C	0	0	0.967	0.85	1.15
12/08/2002	EPA 415.1 (T) Ox	Method Blank	0.07	0.1 mg/L as C	0	0			
12/08/2002	EPA 415.1 (T) Ox	RPD - Check Standard	2.725	%	9.41	9.67		0	0.15
12/08/2002	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.5364	%	24.3	24.17		0	0.15
12/08/2002	EPA 415.1 (T) Ox	Check Standard	9.41	0.1 mg/L as C	0	0	0.941	0.85	1.15
12/08/2002	EPA 415.1 (T) Ox	RPD - Check Standard	2.725	%	9.41	9.67		0	0.15
12/08/2002	EPA 415.1 (T) Ox	Low Level Standard	0.65	0.1 mg/L as C	0	0	1.3	0.5	1.5
12/08/2002	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.5364	%	24.3	24.17		0	0.15

12/08/2002	EPA 415.1 (T) Ox	RPD - Check Standard	2.725	%	9.41	9.67	0	0.15	
19/08/2002	EPA 415.1 (T) Ox	Check Standard	9.43	0.1 mg/L as C			0.943	0.85	1.15
19/08/2002	EPA 415.1 (T) Ox	Check Standard	1.01	0.1 mg/L as C			1.01	0.85	1.15
19/08/2002	EPA 415.1 (T) Ox	Check Standard	9.48	0.1 mg/L as C			0.948	0.85	1.15
19/08/2002	EPA 415.1 (T) Ox	RPD - Check Standard	3.883	%	1.05	1.01	0	0.15	
19/08/2002	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.6771	%	23.71	23.55	0	0.15	
19/08/2002	EPA 415.1 (T) Ox	RPD - Check Standard	0.5288	%	9.43	9.48	0	0.15	
19/08/2002	EPA 415.1 (T) Ox	Method Blank	0.06	0.1 mg/L as C					
19/08/2002	EPA 415.1 (T) Ox	Low Level Standard	0.73	0.1 mg/L as C			1.46	0.5	1.5
19/08/2002	EPA 415.1 (T) Ox	Check Standard	1.05	0.1 mg/L as C			1.05	0.85	1.15
26/08/2002	EPA 415.1 (T) Ox	RPD - LCS Duplicate	1.483	%	23.78	23.43	0	0.15	
26/08/2002	EPA 415.1 (T) Ox	RPD - Check Standard	1.602	%	9.29	9.44	0	0.15	
26/08/2002	EPA 415.1 (T) Ox	Method Blank	0.07	0.1 mg/L as C					
26/08/2002	EPA 415.1 (T) Ox	RPD - Check Standard	0.9479	%	1.05	1.06	0	0.15	
26/08/2002	EPA 415.1 (T) Ox	Check Standard	1.05	0.1 mg/L as C			1.05	0.85	1.15
26/08/2002	EPA 415.1 (T) Ox	Check Standard	1.06	0.1 mg/L as C			1.06	0.85	1.15
26/08/2002	EPA 415.1 (T) Ox	Check Standard	9.29	0.1 mg/L as C			0.929	0.85	1.15
26/08/2002	EPA 415.1 (T) Ox	Check Standard	9.44	0.1 mg/L as C			0.944	0.85	1.15
26/08/2002	EPA 415.1 (T) Ox	Low Level Standard	0.52	0.1 mg/L as C			1.04	0.5	1.5
03/09/2002	EPA 415.1 (T) Ox	Check Standard	9.27	0.1 mg/L as C			0.927	0.85	1.15
03/09/2002	EPA 415.1 (T) Ox	Check Standard	9.41	0.1 mg/L as C			0.941	0.85	1.15
03/09/2002	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0	%	23.63	23.63	0	0.15	
03/09/2002	EPA 415.1 (T) Ox	Method Blank	0.02	0.1 mg/L as C					
03/09/2002	EPA 415.1 (T) Ox	Low Level Standard	0.47	0.1 mg/L as C			0.94	0.5	1.5
03/09/2002	EPA 415.1 (T) Ox	Check Standard	1.05	0.1 mg/L as C			1.05	0.85	1.15
03/09/2002	EPA 415.1 (T) Ox	RPD - Check Standard	1.499	%	9.27	9.41	0	0.15	
09/09/2002	EPA 415.1 (T) Ox	Method Blank	0.06	0.1 mg/L as C					
09/09/2002	EPA 415.1 (T) Ox	Method Blank	0.06	0.1 mg/L as C					
09/09/2002	EPA 415.1 (T) Ox	Low Level Standard	0.5	0.1 mg/L as C			1	0.5	1.5
09/09/2002	EPA 415.1 (T) Ox	RPD - Check Standard	1.796	%	9.55	9.38	0	0.15	
09/09/2002	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.5747	%	24.43	24.29	0	0.15	
09/09/2002	EPA 415.1 (T) Ox	Check Standard	9.55	0.1 mg/L as C			0.955	0.85	1.15
09/09/2002	EPA 415.1 (T) Ox	Check Standard	9.38	0.1 mg/L as C			0.938	0.85	1.15
09/09/2002	EPA 415.1 (T) Ox	RPD - Check Standard	1.796	%	9.55	9.38	0	0.15	
09/09/2002	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.5747	%	24.43	24.29	0	0.15	
09/09/2002	EPA 415.1 (T) Ox	Low Level Standard	0.5	0.1 mg/L as C			1	0.5	1.5
09/09/2002	EPA 415.1 (T) Ox	Check Standard	9.55	0.1 mg/L as C			0.955	0.85	1.15

09/09/2002	EPA 415.1 (T) Ox	Check Standard	1	0.1 mg/L as C			1	0.85	1.15
09/09/2002	EPA 415.1 (T) Ox	Check Standard	9.38	0.1 mg/L as C			0.938	0.85	1.15
09/09/2002	EPA 415.1 (T) Ox	Check Standard	1	0.1 mg/L as C			1	0.85	1.15
16/09/2002	EPA 415.1 (T) Ox	Check Standard	9.14	0.1 mg/L as C			0.914	0.85	1.15
16/09/2002	EPA 415.1 (T) Ox	RPD - Check Standard	3.441	%	9.14	9.46		0	0.15
16/09/2002	EPA 415.1 (T) Ox	Check Standard	9.14	0.1 mg/L as C			0.914	0.85	1.15
16/09/2002	EPA 415.1 (T) Ox	Low Level Standard	0.51	0.1 mg/L as C			1.02	0.5	1.5
16/09/2002	EPA 415.1 (T) Ox	Low Level Standard	0.51	0.1 mg/L as C			1.02	0.5	1.5
16/09/2002	EPA 415.1 (T) Ox	Check Standard	0.94	0.1 mg/L as C			0.94	0.85	1.15
16/09/2002	EPA 415.1 (T) Ox	Check Standard	0.94	0.1 mg/L as C			0.94	0.85	1.15
16/09/2002	EPA 415.1 (T) Ox	Check Standard	9.46	0.1 mg/L as C			0.946	0.85	1.15
16/09/2002	EPA 415.1 (T) Ox	Check Standard	9.46	0.1 mg/L as C			0.946	0.85	1.15
16/09/2002	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.2572	%	23.3	23.36		0	0.15
16/09/2002	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.2572	%	23.3	23.36		0	0.15
16/09/2002	EPA 415.1 (T) Ox	Method Blank	0.05	0.1 mg/L as C					
16/09/2002	EPA 415.1 (T) Ox	Method Blank	0.05	0.1 mg/L as C					
16/09/2002	EPA 415.1 (T) Ox	RPD - Check Standard	3.441	%	9.14	9.46		0	0.15
23/09/2002	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.756	%	23.9	23.72		0	0.15
23/09/2002	EPA 415.1 (T) Ox	Check Standard	9.22	0.1 mg/L as C	0	0	0.922	0.85	1.15
23/09/2002	EPA 415.1 (T) Ox	RPD - Check Standard	0.213	%	9.38	9.4		0	0.15
23/09/2002	EPA 415.1 (T) Ox	RPD - Check Standard	2.985	%	1.02	0.99		0	0.15
23/09/2002	EPA 415.1 (T) Ox	RPD - Check Standard	0.6486	%	9.28	9.22		0	0.15
23/09/2002	EPA 415.1 (T) Ox	Check Standard	9.28	0.1 mg/L as C	0	0	0.928	0.85	1.15
23/09/2002	EPA 415.1 (T) Ox	Check Standard	9.22	0.1 mg/L as C	0	0	0.922	0.85	1.15
23/09/2002	EPA 415.1 (T) Ox	Check Standard	9.28	0.1 mg/L as C	0	0	0.928	0.85	1.15
23/09/2002	EPA 415.1 (T) Ox	Check Standard	9.4	0.1 mg/L as C	0	0	0.94	0.85	1.15
23/09/2002	EPA 415.1 (T) Ox	Check Standard	9.38	0.1 mg/L as C	0	0	0.938	0.85	1.15
23/09/2002	EPA 415.1 (T) Ox	Check Standard	0.99	0.1 mg/L as C	0	0	0.99	0.85	1.15
23/09/2002	EPA 415.1 (T) Ox	Check Standard	1.02	0.1 mg/L as C	0	0	1.02	0.85	1.15
23/09/2002	EPA 415.1 (T) Ox	Check Standard	9.22	0.1 mg/L as C	0	0	0.922	0.85	1.15
23/09/2002	EPA 415.1 (T) Ox	Check Standard	9.28	0.1 mg/L as C	0	0	0.928	0.85	1.15
23/09/2002	EPA 415.1 (T) Ox	RPD - LCS Duplicate	1.452	%	24.28	23.93		0	0.15
23/09/2002	EPA 415.1 (T) Ox	RPD - LCS Duplicate	1.452	%	24.28	23.93		0	0.15
23/09/2002	EPA 415.1 (T) Ox	RPD - LCS Duplicate	1.452	%	24.28	23.93		0	0.15
23/09/2002	EPA 415.1 (T) Ox	RPD - Check Standard	0.6486	%	9.28	9.22		0	0.15
23/09/2002	EPA 415.1 (T) Ox	Low Level Standard	0.56	0.1 mg/L as C	0	0	1.12	0.5	1.5
23/09/2002	EPA 415.1 (T) Ox	Low Level Standard	0.499	0.1 mg/L as C	0	0	0.998	0.5	1.5

23/09/2002	EPA 415.1 (T) Ox	Low Level Standard	0.499	0.1 mg/L as C	0	0	0.998	0.5	1.5
23/09/2002	EPA 415.1 (T) Ox	Low Level Standard	0.499	0.1 mg/L as C	0	0	0.998	0.5	1.5
23/09/2002	EPA 415.1 (T) Ox	Method Blank	0.03	0.1 mg/L as C	0	0			
23/09/2002	EPA 415.1 (T) Ox	Method Blank	0.03	0.1 mg/L as C	0	0			
23/09/2002	EPA 415.1 (T) Ox	Method Blank	0.03	0.1 mg/L as C	0	0			
23/09/2002	EPA 415.1 (T) Ox	Method Blank	0.07	0.1 mg/L as C	0	0			
23/09/2002	EPA 415.1 (T) Ox	RPD - Check Standard	0.6486	%	9.28	9.22		0	0.15
30/09/2002	EPA 415.1 (T) Ox	Check Standard	0.99	0.1 mg/L as C	0	0	0.99	0.85	1.15
30/09/2002	EPA 415.1 (T) Ox	Check Standard	9.41	0.1 mg/L as C	0	0	0.941	0.85	1.15
30/09/2002	EPA 415.1 (T) Ox	Check Standard	0.99	0.1 mg/L as C	0	0	0.99	0.85	1.15
30/09/2002	EPA 415.1 (T) Ox	Check Standard	9.16	0.1 mg/L as C	0	0	0.916	0.85	1.15
30/09/2002	EPA 415.1 (T) Ox	Method Blank	0.07	0.1 mg/L as C	0	0			
30/09/2002	EPA 415.1 (T) Ox	Check Standard	9.41	0.1 mg/L as C	0	0	0.941	0.85	1.15
30/09/2002	EPA 415.1 (T) Ox	Method Blank	0.07	0.1 mg/L as C	0	0			
30/09/2002	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.4692	%	23.5	23.39		0	0.15
30/09/2002	EPA 415.1 (T) Ox	Check Standard	9.16	0.1 mg/L as C	0	0	0.916	0.85	1.15
30/09/2002	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.4692	%	23.5	23.39		0	0.15
30/09/2002	EPA 415.1 (T) Ox	RPD - Check Standard	2.693	%	9.41	9.16		0	0.15
30/09/2002	EPA 415.1 (T) Ox	Low Level Standard	0.53	0.1 mg/L as C	0	0	1.06	0.5	1.5
30/09/2002	EPA 415.1 (T) Ox	Low Level Standard	0.53	0.1 mg/L as C	0	0	1.06	0.5	1.5
30/09/2002	EPA 415.1 (T) Ox	RPD - Check Standard	2.693	%	9.41	9.16		0	0.15
07/10/2002	EPA 415.1 (T) Ox	Method Blank	0.03	0.1 mg/L as C	0	0			
07/10/2002	EPA 415.1 (T) Ox	Check Standard	0.91	0.1 mg/L as C	0	0	0.91	0.85	1.15
07/10/2002	EPA 415.1 (T) Ox	Low Level Standard	0.51	0.1 mg/L as C	0	0	1.02	0.5	1.5
07/10/2002	EPA 415.1 (T) Ox	RPD - Check Standard	0.1063	%	9.41	9.4		0	0.15
07/10/2002	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.8361	%	23.82	24.02		0	0.15
07/10/2002	EPA 415.1 (T) Ox	Check Standard	9.4	0.1 mg/L as C	0	0	0.94	0.85	1.15
07/10/2002	EPA 415.1 (T) Ox	Check Standard	9.41	0.1 mg/L as C	0	0	0.941	0.85	1.15
15/10/2002	EPA 415.1 (T) Ox	Method Blank	0.09	0.1 mg/L as C	0	0			
15/10/2002	EPA 415.1 (T) Ox	Low Level Standard	0.64	0.1 mg/L as C	0	0	1.28	0.5	1.5
15/10/2002	EPA 415.1 (T) Ox	Method Blank	0.09	0.1 mg/L as C	0	0			
15/10/2002	EPA 415.1 (T) Ox	Check Standard	9.55	0.1 mg/L as C	0	0	0.955	0.85	1.15
15/10/2002	EPA 415.1 (T) Ox	Low Level Standard	0.64	0.1 mg/L as C	0	0	1.28	0.5	1.5
15/10/2002	EPA 415.1 (T) Ox	RPD - Check Standard	5.93	%	9.55	9		0	0.15
15/10/2002	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.4083	%	24.44	24.54		0	0.15
15/10/2002	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.4083	%	24.44	24.54		0	0.15
15/10/2002	EPA 415.1 (T) Ox	Check Standard	1.02	0.1 mg/L as C	0	0	1.02	0.85	1.15

15/10/2002	EPA 415.1 (T) Ox	RPD - Check Standard	5.93	%	9.55	9	0	0.15
15/10/2002	EPA 415.1 (T) Ox	Check Standard	9	0.1 mg/L as C	0	0	0.9	0.85
15/10/2002	EPA 415.1 (T) Ox	Check Standard	1.02	0.1 mg/L as C	0	0	1.02	0.85
15/10/2002	EPA 415.1 (T) Ox	Check Standard	9.55	0.1 mg/L as C	0	0	0.955	0.85
15/10/2002	EPA 415.1 (T) Ox	Check Standard	9	0.1 mg/L as C	0	0	0.9	0.85
28/10/2002	EPA 415.1 (T) Ox	RPD - LCS Duplicate	1.363	%	23.63	23.31	0	0.15
28/10/2002	EPA 415.1 (T) Ox	Check Standard	9.19	0.1 mg/L as C	0	0	0.919	0.85
28/10/2002	EPA 415.1 (T) Ox	RPD - Check Standard	2.259	%	9.4	9.19	0	0.15
28/10/2002	EPA 415.1 (T) Ox	RPD - LCS Duplicate	1.363	%	23.63	23.31	0	0.15
28/10/2002	EPA 415.1 (T) Ox	RPD - Check Standard	2.259	%	9.4	9.19	0	0.15
28/10/2002	EPA 415.1 (T) Ox	RPD - Check Standard	1.026	%	0.97	0.98	0	0.15
28/10/2002	EPA 415.1 (T) Ox	RPD - Check Standard	1.026	%	0.97	0.98	0	0.15
28/10/2002	EPA 415.1 (T) Ox	Check Standard	9.4	0.1 mg/L as C	0	0	0.94	0.85
28/10/2002	EPA 415.1 (T) Ox	Method Blank	0.06	0.1 mg/L as C	0	0		
28/10/2002	EPA 415.1 (T) Ox	Method Blank	0.06	0.1 mg/L as C	0	0		
28/10/2002	EPA 415.1 (T) Ox	Check Standard	0.97	0.1 mg/L as C	0	0	0.97	0.85
28/10/2002	EPA 415.1 (T) Ox	Check Standard	0.98	0.1 mg/L as C	0	0	0.98	0.85
28/10/2002	EPA 415.1 (T) Ox	Check Standard	0.98	0.1 mg/L as C	0	0	0.98	0.85
28/10/2002	EPA 415.1 (T) Ox	Check Standard	0.97	0.1 mg/L as C	0	0	0.97	0.85
28/10/2002	EPA 415.1 (T) Ox	Check Standard	9.4	0.1 mg/L as C	0	0	0.94	0.85
28/10/2002	EPA 415.1 (T) Ox	Check Standard	9.19	0.1 mg/L as C	0	0	0.919	0.85
28/10/2002	EPA 415.1 (T) Ox	Low Level Standard	0.5	0.1 mg/L as C	0	0	1	0.5
28/10/2002	EPA 415.1 (T) Ox	Low Level Standard	0.5	0.1 mg/L as C	0	0	1	0.5
04/11/2002	EPA 415.1 (T) Ox	Check Standard	9.76	0.1 mg/L as C	0	0	0.976	0.85
04/11/2002	EPA 415.1 (T) Ox	Method Blank	0.07	0.1 mg/L as C	0	0		
04/11/2002	EPA 415.1 (T) Ox	RPD - LCS Duplicate	1.224	%	24.36	24.66	0	0.15
04/11/2002	EPA 415.1 (T) Ox	RPD - Check Standard	1.653	%	9.6	9.76	0	0.15
04/11/2002	EPA 415.1 (T) Ox	Check Standard	9.6	0.1 mg/L as C	0	0	0.96	0.85
04/11/2002	EPA 415.1 (T) Ox	Low Level Standard	0.54	0.1 mg/L as C	0	0	1.08	0.5
04/11/2002	EPA 415.1 (T) Ox	Check Standard	1.04	0.1 mg/L as C	0	0	1.04	0.85
12/11/2002	EPA 415.1 (T) Ox	Check Standard	0.98	0.1 mg/L as C	0	0	0.98	0.85
12/11/2002	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.5373	%	24.26	24.13	0	0.15
12/11/2002	EPA 415.1 (T) Ox	RPD - Check Standard	0.8565	%	9.38	9.3	0	0.15
12/11/2002	EPA 415.1 (T) Ox	Low Level Standard	0.52	0.1 mg/L as C	0	0	1.04	0.5
12/11/2002	EPA 415.1 (T) Ox	Method Blank	0.09	0.1 mg/L as C	0	0		
12/11/2002	EPA 415.1 (T) Ox	Check Standard	9.3	0.1 mg/L as C	0	0	0.93	0.85
12/11/2002	EPA 415.1 (T) Ox	Check Standard	9.38	0.1 mg/L as C	0	0	0.938	0.85

18/11/2002	EPA 415.1 (T) Ox	Check Standard	9.6	0.1 mg/L as C	0	0	0.96	0.85	1.15
18/11/2002	EPA 415.1 (T) Ox	Method Blank	0.07	0.1 mg/L as C	0	0			
18/11/2002	EPA 415.1 (T) Ox	Check Standard	8.82	0.1 mg/L as C	0	0	0.882	0.85	1.15
18/11/2002	EPA 415.1 (T) Ox	RPD - LCS Duplicate	3.54	%	22.2	23		0	0.15
18/11/2002	EPA 415.1 (T) Ox	Low Level Standard	0.51	0.1 mg/L as C	0	0	1.02	0.5	1.5
18/11/2002	EPA 415.1 (T) Ox	RPD - Check Standard	8.469	%	8.82	9.6		0	0.15
02/12/2002	EPA 415.1 (T) Ox	RPD - Check Standard	0.4386	%	9.14	9.1		0	0.15
02/12/2002	EPA 415.1 (T) Ox	Low Level Standard	0.54	0.1 mg/L as C	0	0	1.08	0.5	1.5
02/12/2002	EPA 415.1 (T) Ox	Method Blank	0.09	0.1 mg/L as C	0	0			
02/12/2002	EPA 415.1 (T) Ox	Check Standard	9.14	0.1 mg/L as C	0	0	0.914	0.85	1.15
02/12/2002	EPA 415.1 (T) Ox	Check Standard	0.96	0.1 mg/L as C	0	0	0.96	0.85	1.15
02/12/2002	EPA 415.1 (T) Ox	Check Standard	9.1	0.1 mg/L as C	0	0	0.91	0.85	1.15
02/12/2002	EPA 415.1 (T) Ox	RPD - LCS Duplicate	1.201	%	23.46	23.18		0	0.15
09/12/2002	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.8701	%	24.03	24.24		0	0.15
09/12/2002	EPA 415.1 (T) Ox	RPD - Check Standard	1.879	%	9.67	9.49		0	0.15
09/12/2002	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.8701	%	24.03	24.24		0	0.15
09/12/2002	EPA 415.1 (T) Ox	RPD - Check Standard	1.879	%	9.67	9.49		0	0.15
09/12/2002	EPA 415.1 (T) Ox	Low Level Standard	0.65	0.1 mg/L as C	0	0	1.3	0.5	1.5
09/12/2002	EPA 415.1 (T) Ox	Check Standard	9.67	0.1 mg/L as C	0	0	0.967	0.85	1.15
09/12/2002	EPA 415.1 (T) Ox	Low Level Standard	0.65	0.1 mg/L as C	0	0	1.3	0.5	1.5
09/12/2002	EPA 415.1 (T) Ox	Method Blank	0.46	0.1 mg/L as C	0	0			
09/12/2002	EPA 415.1 (T) Ox	Method Blank	0.46	0.1 mg/L as C	0	0			
09/12/2002	EPA 415.1 (T) Ox	Check Standard	1.09	0.1 mg/L as C	0	0	1.09	0.85	1.15
09/12/2002	EPA 415.1 (T) Ox	Check Standard	1.09	0.1 mg/L as C	0	0	1.09	0.85	1.15
09/12/2002	EPA 415.1 (T) Ox	Check Standard	9.67	0.1 mg/L as C	0	0	0.967	0.85	1.15
09/12/2002	EPA 415.1 (T) Ox	Check Standard	9.49	0.1 mg/L as C	0	0	0.949	0.85	1.15
09/12/2002	EPA 415.1 (T) Ox	Check Standard	9.49	0.1 mg/L as C	0	0	0.949	0.85	1.15
16/12/2002	EPA 415.1 (T) Ox	Low Level Standard	0.67	0.1 mg/L as C	0	0	1.34	0.5	1.5
16/12/2002	EPA 415.1 (T) Ox	Check Standard	9.68	0.1 mg/L as C	0	0	0.968	0.85	1.15
16/12/2002	EPA 415.1 (T) Ox	Check Standard	9.32	0.1 mg/L as C	0	0	0.932	0.85	1.15
16/12/2002	EPA 415.1 (T) Ox	Method Blank	0.16	0.1 mg/L as C	0	0			
23/12/2002	EPA 415.1 (T) Ox	Check Standard	9.56	0.1 mg/L as C	0	0	0.956	0.85	1.15
23/12/2002	EPA 415.1 (T) Ox	Low Level Standard	0.61	0.1 mg/L as C	0	0	1.22	0.5	1.5
23/12/2002	EPA 415.1 (T) Ox	Low Level Standard	1	0.1 mg/L as C	0	0	1	0.5	1.5
23/12/2002	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.1214	%	24.73	24.7		0	0.15
23/12/2002	EPA 415.1 (T) Ox	Check Standard	9.15	0.1 mg/L as C	0	0	0.915	0.85	1.15
23/12/2002	EPA 415.1 (T) Ox	Method Blank	0.07	0.1 mg/L as C	0	0			

23/12/2002	EPA 415.1 (T) Ox	RPD - Check Standard	4.383	%	9.56	9.15	0	0.15	
30/12/2002	EPA 415.1 (T) Ox	Low Level Standard	1	0.1 mg/L as C	0	0	1	0.5	1.5
30/12/2002	EPA 415.1 (T) Ox	Check Standard	9.32	0.1 mg/L as C	0	0	0.932	0.85	1.15
30/12/2002	EPA 415.1 (T) Ox	Check Standard	9.39	0.1 mg/L as C	0	0	0.939	0.85	1.15
30/12/2002	EPA 415.1 (T) Ox	Check Standard	9.32	0.1 mg/L as C	0	0	0.932	0.85	1.15
30/12/2002	EPA 415.1 (T) Ox	Check Standard	9.39	0.1 mg/L as C	0	0	0.939	0.85	1.15
30/12/2002	EPA 415.1 (T) Ox	Low Level Standard	1	0.1 mg/L as C	0	0	1	0.5	1.5
30/12/2002	EPA 415.1 (T) Ox	Method Blank	0.02	0.1 mg/L as C	0	0			
30/12/2002	EPA 415.1 (T) Ox	Method Blank	0.02	0.1 mg/L as C	0	0			
30/12/2002	EPA 415.1 (T) Ox	RPD - LCS Duplicate	1.131	%	23.74	24.01	0	0.15	
30/12/2002	EPA 415.1 (T) Ox	RPD - Check Standard	0.7483	%	9.32	9.39	0	0.15	
30/12/2002	EPA 415.1 (T) Ox	RPD - LCS Duplicate	1.131	%	23.74	24.01	0	0.15	
30/12/2002	EPA 415.1 (T) Ox	Low Level Standard	0.52	0.1 mg/L as C	0	0	1.04	0.5	1.5
30/12/2002	EPA 415.1 (T) Ox	RPD - Check Standard	0.7483	%	9.32	9.39	0	0.15	
30/12/2002	EPA 415.1 (T) Ox	Low Level Standard	0.993	0.1 mg/L as C	0	0	0.993	0.5	1.5
30/12/2002	EPA 415.1 (T) Ox	Low Level Standard	0.993	0.1 mg/L as C	0	0	0.993	0.5	1.5
30/12/2002	EPA 415.1 (T) Ox	Low Level Standard	0.52	0.1 mg/L as C	0	0	1.04	0.5	1.5
06/01/2003	EPA 415.1 (T) Ox	Check Standard	9.98	0.1 mg/L as C	0	0	0.998	0.85	1.15
06/01/2003	EPA 415.1 (T) Ox	Check Standard	0.98	0.1 mg/L as C	0	0	0.98	0.85	1.15
06/01/2003	EPA 415.1 (T) Ox	Check Standard	9.72	0.1 mg/L as C	0	0	0.972	0.85	1.15
06/01/2003	EPA 415.1 (T) Ox	Low Level Standard	0.57	0.1 mg/L as C	0	0	1.14	0.5	1.5
06/01/2003	EPA 415.1 (T) Ox	Method Blank	0.05	0.1 mg/L as C	0	0			
06/01/2003	EPA 415.1 (T) Ox	RPD - Check Standard	2.64	%	9.72	9.98	0	0.15	
06/01/2003	EPA 415.1 (T) Ox	RPD - LCS Duplicate	1.23	%	23.43	23.72	0	0.15	
14/01/2003	EPA 415.1 (T) Ox	Check Standard	9.04	0.1 mg/L as C	0	0	0.904	0.85	1.15
14/01/2003	EPA 415.1 (T) Ox	Method Blank	0.02	0.1 mg/L as C	0	0			
14/01/2003	EPA 415.1 (T) Ox	Check Standard	9.12	0.1 mg/L as C	0	0	0.912	0.85	1.15
14/01/2003	EPA 415.1 (T) Ox	Check Standard	0.88	0.1 mg/L as C	0	0	0.88	0.85	1.15
14/01/2003	EPA 415.1 (T) Ox	RPD - Check Standard	0.8811	%	9.04	9.12	0	0.15	
14/01/2003	EPA 415.1 (T) Ox	Low Level Standard	0.44	0.1 mg/L as C	0	0	0.88	0.5	1.5
14/01/2003	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.3225	%	21.67	21.74	0	0.15	
21/01/2003	EPA 415.1 (T) Ox	RPD - Check Standard	3.06	%	9.01	9.29	0	0.15	
21/01/2003	EPA 415.1 (T) Ox	Method Blank	0.06	0.1 mg/L as C	0	0			
21/01/2003	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.5141	%	23.4	23.28	0	0.15	
21/01/2003	EPA 415.1 (T) Ox	Check Standard	9.29	0.1 mg/L as C	0	0	0.929	0.85	1.15
21/01/2003	EPA 415.1 (T) Ox	Check Standard	9.01	0.1 mg/L as C	0	0	0.901	0.85	1.15
21/01/2003	EPA 415.1 (T) Ox	Check Standard	0.98	0.1 mg/L as C	0	0	0.98	0.85	1.15

21/01/2003	EPA 415.1 (T) Ox	Low Level Standard	0.52	0.1 mg/L as C	0	0	1.04	0.5	1.5
27/01/2003	EPA 415.1 (T) Ox	Check Standard	8.53	0.1 mg/L as C	0	0	0.853	0.85	1.15
27/01/2003	EPA 415.1 (T) Ox	RPD - Check Standard	5.291	%	0.92	0.97		0	0.15
27/01/2003	EPA 415.1 (T) Ox	Check Standard	9.01	0.1 mg/L as C	0	0	0.901	0.85	1.15
27/01/2003	EPA 415.1 (T) Ox	RPD - LCS Duplicate	1.68	%	22.21	21.84		0	0.15
27/01/2003	EPA 415.1 (T) Ox	Low Level Standard	0.49	0.1 mg/L as C	0	0	0.98	0.5	1.5
27/01/2003	EPA 415.1 (T) Ox	Check Standard	0.92	0.1 mg/L as C	0	0	0.92	0.85	1.15
27/01/2003	EPA 415.1 (T) Ox	RPD - Check Standard	5.473	%	8.53	9.01		0	0.15
27/01/2003	EPA 415.1 (T) Ox	Method Blank	0.06	0.1 mg/L as C	0	0			
27/01/2003	EPA 415.1 (T) Ox	Check Standard	0.97	0.1 mg/L as C	0	0	0.97	0.85	1.15
03/02/2003	EPA 415.1 (T) Ox	Check Standard	8.6	0.1 mg/L as C	0	0	0.86	0.85	1.15
03/02/2003	EPA 415.1 (T) Ox	Check Standard	0.9	0.1 mg/L as C	0	0	0.9	0.85	1.15
03/02/2003	EPA 415.1 (T) Ox	Check Standard	9.05	0.1 mg/L as C	0	0	0.905	0.85	1.15
03/02/2003	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.8819	%	21.45	21.64		0	0.15
03/02/2003	EPA 415.1 (T) Ox	RPD - Check Standard	5.099	%	8.6	9.05		0	0.15
03/02/2003	EPA 415.1 (T) Ox	Low Level Standard	0.5	0.1 mg/L as C	0	0	1	0.5	1.5
03/02/2003	EPA 415.1 (T) Ox	Method Blank	0.07	0.1 mg/L as C	0	0			
10/02/2003	EPA 415.1 (T) Ox	Check Standard	9.45	0.1 mg/L as C	0	0	0.945	0.85	1.15
10/02/2003	EPA 415.1 (T) Ox	Check Standard	1	0.1 mg/L as C	0	0	1	0.85	1.15
10/02/2003	EPA 415.1 (T) Ox	Method Blank	0.08	0.1 mg/L as C	0	0			
10/02/2003	EPA 415.1 (T) Ox	Check Standard	1	0.1 mg/L as C	0	0	1	0.85	1.15
10/02/2003	EPA 415.1 (T) Ox	RPD - Check Standard	0	%	9.45	9.45		0	0.15
10/02/2003	EPA 415.1 (T) Ox	RPD - Check Standard	0	%	1	1		0	0.15
10/02/2003	EPA 415.1 (T) Ox	RPD - LCS Duplicate	1.282	%	24.03	24.34		0	0.15
10/02/2003	EPA 415.1 (T) Ox	Check Standard	9.45	0.1 mg/L as C	0	0	0.945	0.85	1.15
10/02/2003	EPA 415.1 (T) Ox	Low Level Standard	0.56	0.1 mg/L as C	0	0	1.12	0.5	1.5
18/02/2003	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.2014	%	24.85	24.8		0	0.15
18/02/2003	EPA 415.1 (T) Ox	Check Standard	1.07	0.1 mg/L as C	0	0	1.07	0.85	1.15
18/02/2003	EPA 415.1 (T) Ox	Check Standard	9.57	0.1 mg/L as C	0	0	0.957	0.85	1.15
18/02/2003	EPA 415.1 (T) Ox	Check Standard	9.34	0.1 mg/L as C	0	0	0.934	0.85	1.15
18/02/2003	EPA 415.1 (T) Ox	RPD - Check Standard	2.433	%	9.57	9.34		0	0.15
18/02/2003	EPA 415.1 (T) Ox	Method Blank	0.09	0.1 mg/L as C	0	0			
18/02/2003	EPA 415.1 (T) Ox	Low Level Standard	0.56	0.1 mg/L as C	0	0	1.12	0.5	1.5
24/02/2003	EPA 415.1 (T) Ox	RPD - Check Standard	0.9756	%	9.18	9.27		0	0.15
24/02/2003	EPA 415.1 (T) Ox	RPD - Check Standard	7.254	%	0.93	1		0	0.15
24/02/2003	EPA 415.1 (T) Ox	RPD - LCS Duplicate	2.995	%	24.01	24.74		0	0.15
24/02/2003	EPA 415.1 (T) Ox	Check Standard	0.93	0.1 mg/L as C	0	0	0.93	0.85	1.15

24/02/2003	EPA 415.1 (T) Ox	Low Level Standard	0.48	0.1 mg/L as C	0	0	0.96	0.5	1.5
24/02/2003	EPA 415.1 (T) Ox	Method Blank	0.04	0.1 mg/L as C	0	0			
24/02/2003	EPA 415.1 (T) Ox	Check Standard	9.27	0.1 mg/L as C	0	0	0.927	0.85	1.15
24/02/2003	EPA 415.1 (T) Ox	Check Standard	9.18	0.1 mg/L as C	0	0	0.918	0.85	1.15
24/02/2003	EPA 415.1 (T) Ox	Check Standard	1	0.1 mg/L as C	0	0	1	0.85	1.15
03/03/2003	EPA 415.1 (T) Ox	Method Blank	0.08	0.1 mg/L as C	0	0			
03/03/2003	EPA 415.1 (T) Ox	LCS - Recovery	9.22	0.1 mg/L as C	0	0	0.922	0.85	1.15
03/03/2003	EPA 415.1 (T) Ox	LCS - Recovery	9.16	0.1 mg/L as C	0	0	0.916	0.85	1.15
03/03/2003	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.6529	%	9.16	9.22		0	0.15
03/03/2003	EPA 415.1 (T) Ox	Check Standard	9.2	0.1 mg/L as C	0	0	0.92	0.85	1.15
03/03/2003	EPA 415.1 (T) Ox	Low Level Standard	0.52	0.1 mg/L as C	0	0	1.04	0.5	1.5
03/03/2003	EPA 415.1 (T) Ox	Check Standard	9.12	0.1 mg/L as C	0	0	0.912	0.85	1.15
03/03/2003	EPA 415.1 (T) Ox	Check Standard	0.96	0.1 mg/L as C	0	0	0.96	0.85	1.15
03/03/2003	EPA 415.1 (T) Ox	RPD - Check Standard	0.8734	%	9.2	9.12		0	0.15
10/03/2003	EPA 415.1 (T) Ox	LCS - Recovery	9.19	0.1 mg/L as C	0	0	0.919	0.85	1.15
10/03/2003	EPA 415.1 (T) Ox	RPD - LCS Duplicate	1.405	%	9.19	9.32		0	0.15
10/03/2003	EPA 415.1 (T) Ox	LCS - Recovery	9.32	0.1 mg/L as C	0	0	0.932	0.85	1.15
10/03/2003	EPA 415.1 (T) Ox	RPD - Check Standard	1.208	%	9.16	9.05		0	0.15
10/03/2003	EPA 415.1 (T) Ox	Check Standard	9.05	0.1 mg/L as C	0	0	0.905	0.85	1.15
10/03/2003	EPA 415.1 (T) Ox	Low Level Standard	0.47	0.1 mg/L as C	0	0	0.94	0.5	1.5
10/03/2003	EPA 415.1 (T) Ox	Check Standard	9.16	0.1 mg/L as C	0	0	0.916	0.85	1.15
10/03/2003	EPA 415.1 (T) Ox	Method Blank	0.05	0.1 mg/L as C	0	0			
10/03/2003	EPA 415.1 (T) Ox	Check Standard	0.92	0.1 mg/L as C	0	0	0.92	0.85	1.15
17/03/2003	EPA 415.1 (T) Ox	Check Standard	9.01	0.1 mg/L as C	0	0	0.901	0.85	1.15
17/03/2003	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.8811	%	9.04	9.12		0	0.15
17/03/2003	EPA 415.1 (T) Ox	Check Standard	0.89	0.1 mg/L as C	0	0	0.89	0.85	1.15
17/03/2003	EPA 415.1 (T) Ox	Method Blank	0	0.1 mg/L as C	0	0			
17/03/2003	EPA 415.1 (T) Ox	LCS - Recovery	9.04	0.1 mg/L as C	0	0	0.904	0.85	1.15
17/03/2003	EPA 415.1 (T) Ox	RPD - Check Standard	0	%	9.01	9.01		0	0.15
17/03/2003	EPA 415.1 (T) Ox	LCS - Recovery	9.12	0.1 mg/L as C	0	0	0.912	0.85	1.15
17/03/2003	EPA 415.1 (T) Ox	Low Level Standard	0.41	0.1 mg/L as C	0	0	0.82	0.5	1.5
17/03/2003	EPA 415.1 (T) Ox	Check Standard	9.01	0.1 mg/L as C	0	0	0.901	0.85	1.15
24/03/2003	EPA 415.1 (T) Ox	LCS - Recovery	9.72	0.1 mg/L as C	0	0	0.972	0.85	1.15
24/03/2003	EPA 415.1 (T) Ox	Check Standard	9.67	0.1 mg/L as C	0	0	0.967	0.85	1.15
24/03/2003	EPA 415.1 (T) Ox	Check Standard	9.47	0.1 mg/L as C	0	0	0.947	0.85	1.15
24/03/2003	EPA 415.1 (T) Ox	Check Standard	0.94	0.1 mg/L as C	0	0	0.94	0.85	1.15
24/03/2003	EPA 415.1 (T) Ox	Method Blank	0.01	0.1 mg/L as C	0	0			

24/03/2003	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0	%	9.72	9.72	0	0.15	
24/03/2003	EPA 415.1 (T) Ox	LCS - Recovery	9.72	0.1 mg/L as C	0	0	0.972	0.85	1.15
24/03/2003	EPA 415.1 (T) Ox	Low Level Standard	0.47	0.1 mg/L as C	0	0	0.94	0.5	1.5
24/03/2003	EPA 415.1 (T) Ox	RPD - Check Standard	2.09	%	9.47	9.67	0	0.15	
07/04/2003	EPA 415.1 (T) Ox	LCS - Recovery	9.04	0.1 mg/L as C	0	0	0.904	0.85	1.15
07/04/2003	EPA 415.1 (T) Ox	LCS - Recovery	9.14	0.1 mg/L as C	0	0	0.914	0.85	1.15
07/04/2003	EPA 415.1 (T) Ox	Check Standard	0.87	0.1 mg/L as C	0	0	0.87	0.85	1.15
07/04/2003	EPA 415.1 (T) Ox	Check Standard	9.14	0.1 mg/L as C	0	0	0.914	0.85	1.15
07/04/2003	EPA 415.1 (T) Ox	Check Standard	9.21	0.1 mg/L as C	0	0	0.921	0.85	1.15
07/04/2003	EPA 415.1 (T) Ox	RPD - Check Standard	0.7629	%	9.14	9.21	0	0.15	
07/04/2003	EPA 415.1 (T) Ox	RPD - LCS Duplicate	1.1	%	9.14	9.04	0	0.15	
07/04/2003	EPA 415.1 (T) Ox	Method Blank	0	0.1 mg/L as C	0	0			
07/04/2003	EPA 415.1 (T) Ox	Low Level Standard	0.43	0.1 mg/L as C	0	0	0.86	0.5	1.5
14/04/2003	EPA 415.1 (T) Ox	Check Standard	9.71	0.1 mg/L as C	0	0	0.971	0.85	1.15
14/04/2003	EPA 415.1 (T) Ox	Method Blank	0.08	0.1 mg/L as C	0	0			
14/04/2003	EPA 415.1 (T) Ox	LCS - Recovery	9.72	0.1 mg/L as C	0	0	0.972	0.85	1.15
14/04/2003	EPA 415.1 (T) Ox	LCS - Recovery	9.6	0.1 mg/L as C	0	0	0.96	0.85	1.15
14/04/2003	EPA 415.1 (T) Ox	Check Standard	1.06	0.1 mg/L as C	0	0	1.06	0.85	1.15
14/04/2003	EPA 415.1 (T) Ox	Check Standard	9.61	0.1 mg/L as C	0	0	0.961	0.85	1.15
14/04/2003	EPA 415.1 (T) Ox	RPD - Check Standard	1.035	%	9.71	9.61	0	0.15	
14/04/2003	EPA 415.1 (T) Ox	RPD - LCS Duplicate	1.242	%	9.6	9.72	0	0.15	
14/04/2003	EPA 415.1 (T) Ox	Low Level Standard	0.57	0.1 mg/L as C	0	0	1.14	0.5	1.5
21/04/2003	EPA 415.1 (T) Ox	Check Standard	10.42	0.1 mg/L as C	0	0	1.042	0.85	1.15
21/04/2003	EPA 415.1 (T) Ox	Check Standard	0.97	0.1 mg/L as C	0	0	0.97	0.85	1.15
21/04/2003	EPA 415.1 (T) Ox	Check Standard	9.51	0.1 mg/L as C	0	0	0.951	0.85	1.15
21/04/2003	EPA 415.1 (T) Ox	RPD - Check Standard	9.132	%	9.51	10.42	0	0.15	
21/04/2003	EPA 415.1 (T) Ox	LCS - Recovery	9.44	0.1 mg/L as C	0	0	0.944	0.85	1.15
21/04/2003	EPA 415.1 (T) Ox	Low Level Standard	0.51	0.1 mg/L as C	0	0	1.02	0.5	1.5
21/04/2003	EPA 415.1 (T) Ox	Method Blank	0.05	0.1 mg/L as C	0	0			
21/04/2003	EPA 415.1 (T) Ox	RPD - LCS Duplicate	1.054	%	9.54	9.44	0	0.15	
21/04/2003	EPA 415.1 (T) Ox	LCS - Recovery	9.54	0.1 mg/L as C	0	0	0.954	0.85	1.15
28/04/2003	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.5736	%	10.49	10.43	0	0.15	
28/04/2003	EPA 415.1 (T) Ox	Check Standard	10.21	0.1 mg/L as C	0	0	1.021	0.85	1.15
28/04/2003	EPA 415.1 (T) Ox	RPD - Check Standard	1.77	%	1.14	1.12	0	0.15	
28/04/2003	EPA 415.1 (T) Ox	RPD - Check Standard	2.479	%	10.21	9.96	0	0.15	
28/04/2003	EPA 415.1 (T) Ox	Check Standard	9.96	0.1 mg/L as C	0	0	0.996	0.85	1.15
28/04/2003	EPA 415.1 (T) Ox	Check Standard	1.14	0.1 mg/L as C	0	0	1.14	0.85	1.15

28/04/2003	EPA 415.1 (T) Ox	Method Blank	0.09	0.1 mg/L as C	0	0			
28/04/2003	EPA 415.1 (T) Ox	Check Standard	1.12	0.1 mg/L as C	0	0	1.12	0.85	1.15
28/04/2003	EPA 415.1 (T) Ox	Low Level Standard	0.64	0.1 mg/L as C	0	0	1.28	0.5	1.5
28/04/2003	EPA 415.1 (T) Ox	LCS - Recovery	10.43	0.1 mg/L as C	0	0	1.043	0.85	1.15
28/04/2003	EPA 415.1 (T) Ox	LCS - Recovery	10.49	0.1 mg/L as C	0	0	1.049	0.85	1.15
05/05/2003	EPA 415.1 (T) Ox	Check Standard	1.1	0.1 mg/L as C	0	0	1.1	0.85	1.15
05/05/2003	EPA 415.1 (T) Ox	Low Level Standard	0.63	0.1 mg/L as C	0	0	1.26	0.5	1.5
05/05/2003	EPA 415.1 (T) Ox	LCS - Recovery	9.99	0.1 mg/L as C	0	0	0.999	0.85	1.15
05/05/2003	EPA 415.1 (T) Ox	LCS - Recovery	9.88	0.1 mg/L as C	0	0	0.988	0.85	1.15
05/05/2003	EPA 415.1 (T) Ox	Check Standard	9.82	0.1 mg/L as C	0	0	0.982	0.85	1.15
05/05/2003	EPA 415.1 (T) Ox	Method Blank	0.1	0.1 mg/L as C	0	0			
12/05/2003	EPA 415.1 (T) Ox	Check Standard	0.87	0.1 mg/L as C	0	0	0.87	0.85	1.15
12/05/2003	EPA 415.1 (T) Ox	Low Level Standard	0.41	0.1 mg/L as C	0	0	0.82	0.5	1.5
12/05/2003	EPA 415.1 (T) Ox	LCS - Recovery	9.78	0.1 mg/L as C	0	0	0.978	0.85	1.15
12/05/2003	EPA 415.1 (T) Ox	LCS - Recovery	9.74	0.1 mg/L as C	0	0	0.974	0.85	1.15
12/05/2003	EPA 415.1 (T) Ox	Check Standard	9.78	0.1 mg/L as C	0	0	0.978	0.85	1.15
12/05/2003	EPA 415.1 (T) Ox	Check Standard	10.43	0.1 mg/L as C	0	0	1.043	0.85	1.15
12/05/2003	EPA 415.1 (T) Ox	Method Blank	0	0.1 mg/L as C	0	0			
19/05/2003	EPA 415.1 (T) Ox	Check Standard	10.1	0.1 mg/L as C	0	0	1.01	0.85	1.15
19/05/2003	EPA 415.1 (T) Ox	Check Standard	9.8	0.1 mg/L as C	0	0	0.98	0.85	1.15
19/05/2003	EPA 415.1 (T) Ox	LCS - Recovery	9.73	0.1 mg/L as C	0	0	0.973	0.85	1.15
19/05/2003	EPA 415.1 (T) Ox	Check Standard	1.04	0.1 mg/L as C	0	0	1.04	0.85	1.15
19/05/2003	EPA 415.1 (T) Ox	Method Blank	0.07	0.1 mg/L as C	0	0			
19/05/2003	EPA 415.1 (T) Ox	Low Level Standard	0.53	0.1 mg/L as C	0	0	1.06	0.5	1.5
19/05/2003	EPA 415.1 (T) Ox	Check Standard	1.03	0.1 mg/L as C	0	0	1.03	0.85	1.15
19/05/2003	EPA 415.1 (T) Ox	RPD - Check Standard	0.9142	%	9.89	9.8		0	0.15
19/05/2003	EPA 415.1 (T) Ox	Check Standard	9.89	0.1 mg/L as C	0	0	0.989	0.85	1.15
19/05/2003	EPA 415.1 (T) Ox	Check Standard	9.8	0.1 mg/L as C	0	0	0.98	0.85	1.15
19/05/2003	EPA 415.1 (T) Ox	LCS - Recovery	9.87	0.1 mg/L as C	0	0	0.987	0.85	1.15
19/05/2003	EPA 415.1 (T) Ox	LCS - Recovery	9.94	0.1 mg/L as C	0	0	0.994	0.85	1.15
19/05/2003	EPA 415.1 (T) Ox	Low Level Standard	0.52	0.1 mg/L as C	0	0	1.04	0.5	1.5
19/05/2003	EPA 415.1 (T) Ox	Check Standard	1.02	0.1 mg/L as C	0	0	1.02	0.85	1.15
19/05/2003	EPA 415.1 (T) Ox	Method Blank	0.04	0.1 mg/L as C	0	0			
19/05/2003	EPA 415.1 (T) Ox	RPD - Check Standard	1.942	%	1.02	1.04		0	0.15
19/05/2003	EPA 415.1 (T) Ox	RPD - Check Standard	3.015	%	9.8	10.1		0	0.15
19/05/2003	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.6186	%	9.67	9.73		0	0.15
19/05/2003	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.7067	%	9.87	9.94		0	0.15

19/05/2003	EPA 415.1 (T) Ox	LCS - Recovery	9.67	0.1 mg/L as C	0	0	0.967	0.85	1.15
27/05/2003	EPA 415.1 (T) Ox	RPD - Check Standard	1.03	%	9.66	9.76		0	0.15
27/05/2003	EPA 415.1 (T) Ox	RPD - Check Standard	1.942	%	1.04	1.02		0	0.15
27/05/2003	EPA 415.1 (T) Ox	RPD - Check Standard	1.03	%	9.66	9.76		0	0.15
27/05/2003	EPA 415.1 (T) Ox	RPD - LCS Duplicate	1.362	%	9.61	9.48		0	0.15
27/05/2003	EPA 415.1 (T) Ox	Check Standard	9.76	0.1 mg/L as C	0	0	0.976	0.85	1.15
27/05/2003	EPA 415.1 (T) Ox	Method Blank	0.08	0.1 mg/L as C	0	0			
27/05/2003	EPA 415.1 (T) Ox	Check Standard	9.66	0.1 mg/L as C	0	0	0.966	0.85	1.15
27/05/2003	EPA 415.1 (T) Ox	Check Standard	1.02	0.1 mg/L as C	0	0	1.02	0.85	1.15
27/05/2003	EPA 415.1 (T) Ox	Check Standard	1.04	0.1 mg/L as C	0	0	1.04	0.85	1.15
27/05/2003	EPA 415.1 (T) Ox	Check Standard	9.76	0.1 mg/L as C	0	0	0.976	0.85	1.15
27/05/2003	EPA 415.1 (T) Ox	Check Standard	9.66	0.1 mg/L as C	0	0	0.966	0.85	1.15
27/05/2003	EPA 415.1 (T) Ox	Method Blank	0.08	0.1 mg/L as C	0	0			
27/05/2003	EPA 415.1 (T) Ox	Check Standard	1.04	0.1 mg/L as C	0	0	1.04	0.85	1.15
27/05/2003	EPA 415.1 (T) Ox	RPD - LCS Duplicate	1.362	%	9.61	9.48		0	0.15
27/05/2003	EPA 415.1 (T) Ox	Low Level Standard	0.54	0.1 mg/L as C	0	0	1.08	0.5	1.5
27/05/2003	EPA 415.1 (T) Ox	LCS - Recovery	9.48	0.1 mg/L as C	0	0	0.948	0.85	1.15
27/05/2003	EPA 415.1 (T) Ox	LCS - Recovery	9.61	0.1 mg/L as C	0	0	0.961	0.85	1.15
27/05/2003	EPA 415.1 (T) Ox	Low Level Standard	0.54	0.1 mg/L as C	0	0	1.08	0.5	1.5
27/05/2003	EPA 415.1 (T) Ox	LCS - Recovery	9.48	0.1 mg/L as C	0	0	0.948	0.85	1.15
27/05/2003	EPA 415.1 (T) Ox	LCS - Recovery	9.61	0.1 mg/L as C	0	0	0.961	0.85	1.15
27/05/2003	EPA 415.1 (T) Ox	Check Standard	1.02	0.1 mg/L as C	0	0	1.02	0.85	1.15
27/05/2003	EPA 415.1 (T) Ox	RPD - Check Standard	1.942	%	1.04	1.02		0	0.15
02/06/2003	EPA 415.1 (T) Ox	Method Blank	0.09	0.1 mg/L as C	0	0			
02/06/2003	EPA 415.1 (T) Ox	RPD - Check Standard	3.744	%	9.96	10.34		0	0.15
02/06/2003	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.2041	%	9.81	9.79		0	0.15
02/06/2003	EPA 415.1 (T) Ox	LCS - Recovery	9.81	0.1 mg/L as C	0	0	0.981	0.85	1.15
02/06/2003	EPA 415.1 (T) Ox	LCS - Recovery	9.79	0.1 mg/L as C	0	0	0.979	0.85	1.15
02/06/2003	EPA 415.1 (T) Ox	Low Level Standard	0.6	0.1 mg/L as C	0	0	1.2	0.5	1.5
02/06/2003	EPA 415.1 (T) Ox	Check Standard	1.11	0.1 mg/L as C	0	0	1.11	0.85	1.15
02/06/2003	EPA 415.1 (T) Ox	Check Standard	9.96	0.1 mg/L as C	0	0	0.996	0.85	1.15
02/06/2003	EPA 415.1 (T) Ox	Check Standard	10.34	0.1 mg/L as C	0	0	1.034	0.85	1.15
10/06/2003	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.1998	%	10	10.02		0	0.15
10/06/2003	EPA 415.1 (T) Ox	RPD - Check Standard	3.835	%	10.1	9.72		0	0.15
10/06/2003	EPA 415.1 (T) Ox	Check Standard	1.06	0.1 mg/L as C	0	0	1.06	0.85	1.15
10/06/2003	EPA 415.1 (T) Ox	RPD - Check Standard	3.835	%	10.1	9.72		0	0.15
10/06/2003	EPA 415.1 (T) Ox	Check Standard	9.72	0.1 mg/L as C	0	0	0.972	0.85	1.15

10/06/2003 EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.1998	%	10	10.02		0	0.15
10/06/2003 EPA 415.1 (T) Ox	Check Standard	1.06	0.1 mg/L as C	0	0	1.06	0.85	1.15
10/06/2003 EPA 415.1 (T) Ox	Check Standard	10.1	0.1 mg/L as C	0	0	1.01	0.85	1.15
10/06/2003 EPA 415.1 (T) Ox	Check Standard	9.72	0.1 mg/L as C	0	0	0.972	0.85	1.15
10/06/2003 EPA 415.1 (T) Ox	Check Standard	10.1	0.1 mg/L as C	0	0	1.01	0.85	1.15
10/06/2003 EPA 415.1 (T) Ox	LCS - Recovery	10.02	0.1 mg/L as C	0	0	1.002	0.85	1.15
10/06/2003 EPA 415.1 (T) Ox	Method Blank	0.03	0.1 mg/L as C	0	0			
10/06/2003 EPA 415.1 (T) Ox	Method Blank	0.03	0.1 mg/L as C	0	0			
10/06/2003 EPA 415.1 (T) Ox	Low Level Standard	0.52	0.1 mg/L as C	0	0	1.04	0.5	1.5
10/06/2003 EPA 415.1 (T) Ox	LCS - Recovery	10.02	0.1 mg/L as C	0	0	1.002	0.85	1.15
10/06/2003 EPA 415.1 (T) Ox	LCS - Recovery	10	0.1 mg/L as C	0	0	1	0.85	1.15
10/06/2003 EPA 415.1 (T) Ox	Low Level Standard	0.52	0.1 mg/L as C	0	0	1.04	0.5	1.5
10/06/2003 EPA 415.1 (T) Ox	LCS - Recovery	10	0.1 mg/L as C	0	0	1	0.85	1.15
16/06/2003 EPA 415.1 (T) Ox	Check Standard	9.7	0.1 mg/L as C	0	0	0.97	0.85	1.15
16/06/2003 EPA 415.1 (T) Ox	Check Standard	9.77	0.1 mg/L as C	0	0	0.977	0.85	1.15
16/06/2003 EPA 415.1 (T) Ox	Check Standard	1.05	0.1 mg/L as C	0	0	1.05	0.85	1.15
16/06/2003 EPA 415.1 (T) Ox	LCS - Recovery	9.7	0.1 mg/L as C	0	0	0.97	0.85	1.15
16/06/2003 EPA 415.1 (T) Ox	LCS - Recovery	9.52	0.1 mg/L as C	0	0	0.952	0.85	1.15
16/06/2003 EPA 415.1 (T) Ox	Low Level Standard	0.56	0.1 mg/L as C	0	0	1.12	0.5	1.5
16/06/2003 EPA 415.1 (T) Ox	Method Blank	0.08	0.1 mg/L as C	0	0			
16/06/2003 EPA 415.1 (T) Ox	RPD - Check Standard	0.7191	%	9.77	9.7		0	0.15
16/06/2003 EPA 415.1 (T) Ox	RPD - LCS Duplicate	1.873	%	9.7	9.52		0	0.15
16/06/2003 EPA 415.1 (T) Ox	RPD - LCS Duplicate	1.873	%	9.7	9.52		0	0.15
16/06/2003 EPA 415.1 (T) Ox	Check Standard	9.77	0.1 mg/L as C	0	0	0.977	0.85	1.15
16/06/2003 EPA 415.1 (T) Ox	LCS - Recovery	9.7	0.1 mg/L as C	0	0	0.97	0.85	1.15
16/06/2003 EPA 415.1 (T) Ox	LCS - Recovery	9.52	0.1 mg/L as C	0	0	0.952	0.85	1.15
16/06/2003 EPA 415.1 (T) Ox	Method Blank	0.08	0.1 mg/L as C	0	0			
16/06/2003 EPA 415.1 (T) Ox	Low Level Standard	0.56	0.1 mg/L as C	0	0	1.12	0.5	1.5
16/06/2003 EPA 415.1 (T) Ox	Check Standard	9.7	0.1 mg/L as C	0	0	0.97	0.85	1.15
16/06/2003 EPA 415.1 (T) Ox	RPD - Check Standard	0.7191	%	9.77	9.7		0	0.15
16/06/2003 EPA 415.1 (T) Ox	Check Standard	1.05	0.1 mg/L as C	0	0	1.05	0.85	1.15
23/06/2003 EPA 415.1 (T) Ox	Method Blank	0.09	0.1 mg/L as C	0	0			
23/06/2003 EPA 415.1 (T) Ox	Low Level Standard	0.65	0.1 mg/L as C	0	0	1.3	0.5	1.5
23/06/2003 EPA 415.1 (T) Ox	RPD - Check Standard	6.186	%	10.81	11.5		0	0.15
23/06/2003 EPA 415.1 (T) Ox	RPD - LCS Duplicate	1.958	%	10.62	10.83		0	0.15
30/06/2003 EPA 415.1 (T) Ox	LCS - Recovery	9.75	0.1 mg/L as C	0	0	0.975	0.85	1.15
30/06/2003 EPA 415.1 (T) Ox	RPD - Check Standard	2.469	%	9.84	9.6		0	0.15

30/06/2003	EPA 415.1 (T) Ox	Method Blank	0.08	0.1 mg/L as C	0	0			
30/06/2003	EPA 415.1 (T) Ox	LCS - Recovery	9.96	0.1 mg/L as C	0	0	0.996	0.85	1.15
30/06/2003	EPA 415.1 (T) Ox	RPD - LCS Duplicate	2.131	%	9.96	9.75		0	0.15
30/06/2003	EPA 415.1 (T) Ox	Check Standard	9.6	0.1 mg/L as C	0	0	0.96	0.85	1.15
30/06/2003	EPA 415.1 (T) Ox	Check Standard	9.84	0.1 mg/L as C	0	0	0.984	0.85	1.15
30/06/2003	EPA 415.1 (T) Ox	Check Standard	1.1	0.1 mg/L as C	0	0	1.1	0.85	1.15
30/06/2003	EPA 415.1 (T) Ox	Low Level Standard	0.62	0.1 mg/L as C	0	0	1.24	0.5	1.5
07/07/2003	EPA 415.1 (T) Ox	Check Standard	10.46	0.1 mg/L as C	0	0	1.046	0.85	1.15
07/07/2003	EPA 415.1 (T) Ox	Check Standard	1.12	0.1 mg/L as C	0	0	1.12	0.85	1.15
07/07/2003	EPA 415.1 (T) Ox	LCS - Recovery	10.33	0.1 mg/L as C	0	0	1.033	0.85	1.15
07/07/2003	EPA 415.1 (T) Ox	Check Standard	10.17	0.1 mg/L as C	0	0	1.017	0.85	1.15
07/07/2003	EPA 415.1 (T) Ox	Low Level Standard	0.62	0.1 mg/L as C	0	0	1.24	0.5	1.5
07/07/2003	EPA 415.1 (T) Ox	LCS - Recovery	10.3	0.1 mg/L as C	0	0	1.03	0.85	1.15
07/07/2003	EPA 415.1 (T) Ox	Method Blank	0.09	0.1 mg/L as C	0	0			
07/07/2003	EPA 415.1 (T) Ox	RPD - Check Standard	2.811	%	10.17	10.46		0	0.15
07/07/2003	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.2908	%	10.3	10.33		0	0.15
14/07/2003	EPA 415.1 (T) Ox	Check Standard	10.13	0.1 mg/L as C	0	0	1.013	0.85	1.15
14/07/2003	EPA 415.1 (T) Ox	LCS - Recovery	10.34	0.1 mg/L as C	0	0	1.034	0.85	1.15
14/07/2003	EPA 415.1 (T) Ox	Low Level Standard	0.68	0.1 mg/L as C	0	0	1.36	0.5	1.5
14/07/2003	EPA 415.1 (T) Ox	LCS - Recovery	10.32	0.1 mg/L as C	0	0	1.032	0.85	1.15
14/07/2003	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.1936	%	10.32	10.34		0	0.15
14/07/2003	EPA 415.1 (T) Ox	Method Blank	0.04	0.1 mg/L as C	0	0			
14/07/2003	EPA 415.1 (T) Ox	Check Standard	10.37	0.1 mg/L as C	0	0	1.037	0.85	1.15
14/07/2003	EPA 415.1 (T) Ox	RPD - Check Standard	2.341	%	10.37	10.13		0	0.15
21/07/2003	EPA 415.1 (T) Ox	Check Standard	10.4	0.1 mg/L as C	0	0	1.04	0.85	1.15
21/07/2003	EPA 415.1 (T) Ox	Check Standard	10.32	0.1 mg/L as C	0	0	1.032	0.85	1.15
21/07/2003	EPA 415.1 (T) Ox	Check Standard	1.06	0.1 mg/L as C	0	0	1.06	0.85	1.15
21/07/2003	EPA 415.1 (T) Ox	Method Blank	0.08	0.1 mg/L as C	0	0			
21/07/2003	EPA 415.1 (T) Ox	Low Level Standard	0.68	0.1 mg/L as C	0	0	1.36	0.5	1.5
21/07/2003	EPA 415.1 (T) Ox	Low Level Standard	0.68	0.1 mg/L as C	0	0	1.36	0.5	1.5
21/07/2003	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.5808	%	10.3	10.36		0	0.15
21/07/2003	EPA 415.1 (T) Ox	RPD - Check Standard	0.7722	%	10.32	10.4		0	0.15
21/07/2003	EPA 415.1 (T) Ox	LCS - Recovery	10.3	0.1 mg/L as C	0	0	1.03	0.85	1.15
21/07/2003	EPA 415.1 (T) Ox	LCS - Recovery	10.36	0.1 mg/L as C	0	0	1.036	0.85	1.15
21/07/2003	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.5808	%	10.3	10.36		0	0.15
21/07/2003	EPA 415.1 (T) Ox	RPD - Check Standard	0.7722	%	10.32	10.4		0	0.15
21/07/2003	EPA 415.1 (T) Ox	Method Blank	0.08	0.1 mg/L as C	0	0			

21/07/2003	EPA 415.1 (T) Ox	Check Standard	10.4	0.1 mg/L as C	0	0	1.04	0.85	1.15
21/07/2003	EPA 415.1 (T) Ox	Check Standard	10.32	0.1 mg/L as C	0	0	1.032	0.85	1.15
21/07/2003	EPA 415.1 (T) Ox	Check Standard	1.06	0.1 mg/L as C	0	0	1.06	0.85	1.15
21/07/2003	EPA 415.1 (T) Ox	LCS - Recovery	10.3	0.1 mg/L as C	0	0	1.03	0.85	1.15
21/07/2003	EPA 415.1 (T) Ox	LCS - Recovery	10.36	0.1 mg/L as C	0	0	1.036	0.85	1.15
28/07/2003	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.2041	%	9.79	9.81		0	0.15
28/07/2003	EPA 415.1 (T) Ox	Method Blank	0	0.1 mg/L as C	0	0			
28/07/2003	EPA 415.1 (T) Ox	Check Standard	9.8	0.1 mg/L as C	0	0	0.98	0.85	1.15
28/07/2003	EPA 415.1 (T) Ox	Method Blank	0	0.1 mg/L as C	0	0			
28/07/2003	EPA 415.1 (T) Ox	Low Level Standard	0.45	0.1 mg/L as C	0	0	0.9	0.5	1.5
28/07/2003	EPA 415.1 (T) Ox	LCS - Recovery	9.79	0.1 mg/L as C	0	0	0.979	0.85	1.15
28/07/2003	EPA 415.1 (T) Ox	Check Standard	0.95	0.1 mg/L as C	0	0	0.95	0.85	1.15
28/07/2003	EPA 415.1 (T) Ox	Check Standard	9.7	0.1 mg/L as C	0	0	0.97	0.85	1.15
28/07/2003	EPA 415.1 (T) Ox	Check Standard	9.7	0.1 mg/L as C	0	0	0.97	0.85	1.15
28/07/2003	EPA 415.1 (T) Ox	Check Standard	0.95	0.1 mg/L as C	0	0	0.95	0.85	1.15
28/07/2003	EPA 415.1 (T) Ox	LCS - Recovery	9.81	0.1 mg/L as C	0	0	0.981	0.85	1.15
28/07/2003	EPA 415.1 (T) Ox	Check Standard	9.8	0.1 mg/L as C	0	0	0.98	0.85	1.15
28/07/2003	EPA 415.1 (T) Ox	RPD - Check Standard	1.026	%	9.8	9.7		0	0.15
28/07/2003	EPA 415.1 (T) Ox	LCS - Recovery	9.79	0.1 mg/L as C	0	0	0.979	0.85	1.15
28/07/2003	EPA 415.1 (T) Ox	LCS - Recovery	9.81	0.1 mg/L as C	0	0	0.981	0.85	1.15
28/07/2003	EPA 415.1 (T) Ox	RPD - Check Standard	1.026	%	9.8	9.7		0	0.15
28/07/2003	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.2041	%	9.79	9.81		0	0.15
28/07/2003	EPA 415.1 (T) Ox	Low Level Standard	0.45	0.1 mg/L as C	0	0	0.9	0.5	1.5
04/08/2003	EPA 415.1 (T) Ox	RPD - Check Standard	3.345	%	9.7	10.03		0	0.15
04/08/2003	EPA 415.1 (T) Ox	Check Standard	1.08	0.1 mg/L as C	0	0	1.08	0.85	1.15
04/08/2003	EPA 415.1 (T) Ox	Check Standard	1.04	0.1 mg/L as C	0	0	1.04	0.85	1.15
04/08/2003	EPA 415.1 (T) Ox	Check Standard	9.7	0.1 mg/L as C	0	0	0.97	0.85	1.15
04/08/2003	EPA 415.1 (T) Ox	RPD - Check Standard	3.774	%	1.04	1.08		0	0.15
04/08/2003	EPA 415.1 (T) Ox	Method Blank	0.08	0.1 mg/L as C	0	0			
04/08/2003	EPA 415.1 (T) Ox	LCS - Recovery	9.78	0.1 mg/L as C	0	0	0.978	0.85	1.15
04/08/2003	EPA 415.1 (T) Ox	Check Standard	10.03	0.1 mg/L as C	0	0	1.003	0.85	1.15
04/08/2003	EPA 415.1 (T) Ox	LCS - Recovery	9.67	0.1 mg/L as C	0	0	0.967	0.85	1.15
04/08/2003	EPA 415.1 (T) Ox	Low Level Standard	0.57	0.1 mg/L as C	0	0	1.14	0.5	1.5
04/08/2003	EPA 415.1 (T) Ox	RPD - LCS Duplicate	1.131	%	9.78	9.67		0	0.15
18/08/2003	EPA 415.1 (T) Ox	LCS - Recovery	9.85	0.1 mg/L as C	0	0	0.985	0.85	1.15
18/08/2003	EPA 415.1 (T) Ox	LCS - Recovery	9.75	0.1 mg/L as C	0	0	0.975	0.85	1.15
18/08/2003	EPA 415.1 (T) Ox	Low Level Standard	0.61	0.1 mg/L as C	0	0	1.22	0.5	1.5

18/08/2003	EPA 415.1 (T) Ox	RPD - LCS Duplicate	1.02	%	9.85	9.75	0	0.15	
18/08/2003	EPA 415.1 (T) Ox	Low Level Standard	0.61	0.1 mg/L as C	0	0	1.22	0.5	1.5
18/08/2003	EPA 415.1 (T) Ox	Check Standard	9.85	0.1 mg/L as C	0	0	0.985	0.85	1.15
18/08/2003	EPA 415.1 (T) Ox	Check Standard	10.01	0.1 mg/L as C	0	0	1.001	0.85	1.15
18/08/2003	EPA 415.1 (T) Ox	Check Standard	1.08	0.1 mg/L as C	0	0	1.08	0.85	1.15
18/08/2003	EPA 415.1 (T) Ox	LCS - Recovery	9.85	0.1 mg/L as C	0	0	0.985	0.85	1.15
18/08/2003	EPA 415.1 (T) Ox	Method Blank	0.09	0.1 mg/L as C	0	0			
18/08/2003	EPA 415.1 (T) Ox	RPD - Check Standard	1.611	%	10.01	9.85	0	0.15	
18/08/2003	EPA 415.1 (T) Ox	RPD - LCS Duplicate	1.02	%	9.85	9.75	0	0.15	
18/08/2003	EPA 415.1 (T) Ox	Method Blank	0.09	0.1 mg/L as C	0	0			
18/08/2003	EPA 415.1 (T) Ox	RPD - Check Standard	1.611	%	10.01	9.85	0	0.15	
18/08/2003	EPA 415.1 (T) Ox	Check Standard	1.08	0.1 mg/L as C	0	0	1.08	0.85	1.15
18/08/2003	EPA 415.1 (T) Ox	Check Standard	9.85	0.1 mg/L as C	0	0	0.985	0.85	1.15
18/08/2003	EPA 415.1 (T) Ox	LCS - Recovery	9.75	0.1 mg/L as C	0	0	0.975	0.85	1.15
18/08/2003	EPA 415.1 (T) Ox	Check Standard	10.01	0.1 mg/L as C	0	0	1.001	0.85	1.15
25/08/2003	EPA 415.1 (T) Ox	LCS - Recovery	10.1	0.1 mg/L as C	0	0	1.01	0.85	1.15
25/08/2003	EPA 415.1 (T) Ox	Check Standard	9.89	0.1 mg/L as C	0	0	0.989	0.85	1.15
25/08/2003	EPA 415.1 (T) Ox	RPD - LCS Duplicate	1.496	%	10.1	9.95	0	0.15	
25/08/2003	EPA 415.1 (T) Ox	LCS - Recovery	9.95	0.1 mg/L as C	0	0	0.995	0.85	1.15
25/08/2003	EPA 415.1 (T) Ox	Check Standard	1.14	0.1 mg/L as C	0	0	1.14	0.85	1.15
25/08/2003	EPA 415.1 (T) Ox	RPD - Check Standard	2.002	%	10.09	9.89	0	0.15	
25/08/2003	EPA 415.1 (T) Ox	Low Level Standard	0.66	0.1 mg/L as C	0	0	1.32	0.5	1.5
25/08/2003	EPA 415.1 (T) Ox	Check Standard	10.09	0.1 mg/L as C	0	0	1.009	0.85	1.15
25/08/2003	EPA 415.1 (T) Ox	Method Blank	0.09	0.1 mg/L as C	0	0			
08/09/2003	EPA 415.1 (T) Ox	Method Blank	0.05	0.1 mg/L as C	0	0			
08/09/2003	EPA 415.1 (T) Ox	LCS - Recovery	9.6	0.1 mg/L as C	0	0	0.96	0.85	1.15
08/09/2003	EPA 415.1 (T) Ox	Low Level Standard	0.57	0.1 mg/L as C	0	0	1.14	0.5	1.5
08/09/2003	EPA 415.1 (T) Ox	Check Standard	1.07	0.1 mg/L as C	0	0	1.07	0.85	1.15
08/09/2003	EPA 415.1 (T) Ox	Check Standard	9.49	0.1 mg/L as C	0	0	0.949	0.85	1.15
08/09/2003	EPA 415.1 (T) Ox	RPD - Check Standard	1.361	%	9.49	9.62	0	0.15	
08/09/2003	EPA 415.1 (T) Ox	LCS - Recovery	9.63	0.1 mg/L as C	0	0	0.963	0.85	1.15
08/09/2003	EPA 415.1 (T) Ox	Check Standard	9.62	0.1 mg/L as C	0	0	0.962	0.85	1.15
08/09/2003	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.312	%	9.6	9.63	0	0.15	
16/09/2003	EPA 415.1 (T) Ox	Check Standard	9.84	0.1 mg/L as C	0	0	0.984	0.85	1.15
16/09/2003	EPA 415.1 (T) Ox	RPD - Check Standard	0.4073	%	9.84	9.8	0	0.15	
16/09/2003	EPA 415.1 (T) Ox	RPD - LCS Duplicate	2.603	%	9.86	10.12	0	0.15	
16/09/2003	EPA 415.1 (T) Ox	LCS - Recovery	9.86	0.1 mg/L as C	0	0	0.986	0.85	1.15

16/09/2003	EPA 415.1 (T) Ox	Check Standard	1.05	0.1 mg/L as C	0	0	1.05	0.85	1.15
16/09/2003	EPA 415.1 (T) Ox	Method Blank	0.07	0.1 mg/L as C	0	0			
16/09/2003	EPA 415.1 (T) Ox	Low Level Standard	0.57	0.1 mg/L as C	0	0	1.14	0.5	1.5
16/09/2003	EPA 415.1 (T) Ox	Check Standard	9.8	0.1 mg/L as C	0	0	0.98	0.85	1.15
16/09/2003	EPA 415.1 (T) Ox	LCS - Recovery	10.12	0.1 mg/L as C	0	0	1.012	0.85	1.15
22/09/2003	EPA 415.1 (T) Ox	Low Level Standard	0.51	0.1 mg/L as C	0	0	1.02	0.5	1.5
22/09/2003	EPA 415.1 (T) Ox	LCS - Recovery	9.5	0.1 mg/L as C	0	0	0.95	0.85	1.15
22/09/2003	EPA 415.1 (T) Ox	LCS - Recovery	9.51	0.1 mg/L as C	0	0	0.951	0.85	1.15
22/09/2003	EPA 415.1 (T) Ox	LCS - Recovery	9.5	0.1 mg/L as C	0	0	0.95	0.85	1.15
22/09/2003	EPA 415.1 (T) Ox	Method Blank	0.06	0.1 mg/L as C	0	0			
22/09/2003	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.1052	%	9.51	9.5		0	0.15
22/09/2003	EPA 415.1 (T) Ox	Check Standard	9.46	0.1 mg/L as C	0	0	0.946	0.85	1.15
22/09/2003	EPA 415.1 (T) Ox	Check Standard	9.3	0.1 mg/L as C	0	0	0.93	0.85	1.15
22/09/2003	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.1052	%	9.51	9.5		0	0.15
22/09/2003	EPA 415.1 (T) Ox	Method Blank	0.06	0.1 mg/L as C	0	0			
22/09/2003	EPA 415.1 (T) Ox	Check Standard	9.3	0.1 mg/L as C	0	0	0.93	0.85	1.15
22/09/2003	EPA 415.1 (T) Ox	RPD - Check Standard	1.706	%	9.46	9.3		0	0.15
22/09/2003	EPA 415.1 (T) Ox	RPD - Check Standard	1.706	%	9.46	9.3		0	0.15
22/09/2003	EPA 415.1 (T) Ox	Check Standard	9.46	0.1 mg/L as C	0	0	0.946	0.85	1.15
22/09/2003	EPA 415.1 (T) Ox	LCS - Recovery	9.51	0.1 mg/L as C	0	0	0.951	0.85	1.15
22/09/2003	EPA 415.1 (T) Ox	Low Level Standard	0.51	0.1 mg/L as C	0	0	1.02	0.5	1.5
29/09/2003	EPA 415.1 (T) Ox	Check Standard	9.71	0.1 mg/L as C	0	0	0.971	0.85	1.15
29/09/2003	EPA 415.1 (T) Ox	Method Blank	0.05	0.1 mg/L as C	0	0			
29/09/2003	EPA 415.1 (T) Ox	RPD - Check Standard	1.59	%	9.51	9.36		0	0.15
29/09/2003	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.3163	%	9.5	9.47		0	0.15
29/09/2003	EPA 415.1 (T) Ox	Check Standard	9.36	0.1 mg/L as C	0	0	0.936	0.85	1.15
29/09/2003	EPA 415.1 (T) Ox	Check Standard	9.51	0.1 mg/L as C	0	0	0.951	0.85	1.15
29/09/2003	EPA 415.1 (T) Ox	Check Standard	9.6	0.1 mg/L as C	0	0	0.96	0.85	1.15
29/09/2003	EPA 415.1 (T) Ox	Low Level Standard	0.484	0.1 mg/L as C	0	0	0.968	0.5	1.5
29/09/2003	EPA 415.1 (T) Ox	RPD - Check Standard	1.139	%	9.6	9.71		0	0.15
29/09/2003	EPA 415.1 (T) Ox	LCS - Recovery	9.5	0.1 mg/L as C	0	0	0.95	0.85	1.15
29/09/2003	EPA 415.1 (T) Ox	Low Level Standard	0.496	0.1 mg/L as C	0	0	0.992	0.5	1.5
29/09/2003	EPA 415.1 (T) Ox	LCS - Recovery	9.47	0.1 mg/L as C	0	0	0.947	0.85	1.15
29/09/2003	EPA 415.1 (T) Ox	Method Blank	0.03	0.1 mg/L as C	0	0			
06/10/2003	EPA 415.1 (T) Ox	Check Standard	9.6	0.1 mg/L as C	0	0	0.96	0.85	1.15
06/10/2003	EPA 415.1 (T) Ox	RPD - Check Standard	2.746	%	9.6	9.34		0	0.15
06/10/2003	EPA 415.1 (T) Ox	Method Blank	0.03	0.1 mg/L as C	0	0			

06/10/2003	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.1031	%	9.69	9.7	0	0.15	
06/10/2003	EPA 415.1 (T) Ox	RPD - Check Standard	2.746	%	9.6	9.34	0	0.15	
06/10/2003	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.1031	%	9.69	9.7	0	0.15	
06/10/2003	EPA 415.1 (T) Ox	LCS - Recovery	9.7	0.1 mg/L as C	0	0	0.97	0.85	1.15
06/10/2003	EPA 415.1 (T) Ox	LCS - Recovery	9.69	0.1 mg/L as C	0	0	0.969	0.85	1.15
06/10/2003	EPA 415.1 (T) Ox	LCS - Recovery	9.69	0.1 mg/L as C	0	0	0.969	0.85	1.15
06/10/2003	EPA 415.1 (T) Ox	Check Standard	9.34	0.1 mg/L as C	0	0	0.934	0.85	1.15
06/10/2003	EPA 415.1 (T) Ox	Check Standard	9.6	0.1 mg/L as C	0	0	0.96	0.85	1.15
06/10/2003	EPA 415.1 (T) Ox	Check Standard	9.34	0.1 mg/L as C	0	0	0.934	0.85	1.15
06/10/2003	EPA 415.1 (T) Ox	Method Blank	0.03	0.1 mg/L as C	0	0			
06/10/2003	EPA 415.1 (T) Ox	LCS - Recovery	9.7	0.1 mg/L as C	0	0	0.97	0.85	1.15
14/10/2003	EPA 415.1 (T) Ox	Low Level Standard	0.69	0.1 mg/L as C	0	0	1.38	0.5	1.5
14/10/2003	EPA 415.1 (T) Ox	Low Level Standard	0.69	0.1 mg/L as C	0	0	1.38	0.5	1.5
14/10/2003	EPA 415.1 (T) Ox	LCS - Recovery	10.08	0.1 mg/L as C	0	0	1.008	0.85	1.15
14/10/2003	EPA 415.1 (T) Ox	LCS - Recovery	10.21	0.1 mg/L as C	0	0	1.021	0.85	1.15
14/10/2003	EPA 415.1 (T) Ox	LCS - Recovery	10.08	0.1 mg/L as C	0	0	1.008	0.85	1.15
14/10/2003	EPA 415.1 (T) Ox	LCS - Recovery	10.21	0.1 mg/L as C	0	0	1.021	0.85	1.15
14/10/2003	EPA 415.1 (T) Ox	Check Standard	10.22	0.1 mg/L as C	0	0	1.022	0.85	1.15
14/10/2003	EPA 415.1 (T) Ox	Check Standard	1.14	0.1 mg/L as C	0	0	1.14	0.85	1.15
14/10/2003	EPA 415.1 (T) Ox	Check Standard	10.08	0.1 mg/L as C	0	0	1.008	0.85	1.15
14/10/2003	EPA 415.1 (T) Ox	Check Standard	10.22	0.1 mg/L as C	0	0	1.022	0.85	1.15
14/10/2003	EPA 415.1 (T) Ox	RPD - LCS Duplicate	1.281	%	10.21	10.08	0	0.15	
14/10/2003	EPA 415.1 (T) Ox	RPD - Check Standard	1.379	%	10.08	10.22	0	0.15	
14/10/2003	EPA 415.1 (T) Ox	Method Blank	0.07	0.1 mg/L as C	0	0			
14/10/2003	EPA 415.1 (T) Ox	Check Standard	10.08	0.1 mg/L as C	0	0	1.008	0.85	1.15
14/10/2003	EPA 415.1 (T) Ox	RPD - Check Standard	1.379	%	10.08	10.22	0	0.15	
14/10/2003	EPA 415.1 (T) Ox	Method Blank	0.07	0.1 mg/L as C	0	0			
14/10/2003	EPA 415.1 (T) Ox	RPD - LCS Duplicate	1.281	%	10.21	10.08	0	0.15	
14/10/2003	EPA 415.1 (T) Ox	Check Standard	1.14	0.1 mg/L as C	0	0	1.14	0.85	1.15
20/10/2003	EPA 415.1 (T) Ox	Low Level Standard	0.53	0.1 mg/L as C	0	0	1.06	0.5	1.5
20/10/2003	EPA 415.1 (T) Ox	LCS - Recovery	9.95	0.1 mg/L as C	0	0	0.995	0.85	1.15
20/10/2003	EPA 415.1 (T) Ox	RPD - Check Standard	1.319	%	9.79	9.92	0	0.15	
20/10/2003	EPA 415.1 (T) Ox	LCS - Recovery	9.94	0.1 mg/L as C	0	0	0.994	0.85	1.15
20/10/2003	EPA 415.1 (T) Ox	Method Blank	0.06	0.1 mg/L as C	0	0			
20/10/2003	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.1006	%	9.94	9.95	0	0.15	
20/10/2003	EPA 415.1 (T) Ox	Check Standard	1.04	0.1 mg/L as C	0	0	1.04	0.85	1.15
20/10/2003	EPA 415.1 (T) Ox	Check Standard	1.04	0.1 mg/L as C	0	0	1.04	0.85	1.15

20/10/2003 EPA 415.1 (T) Ox	RPD - Check Standard	0	%	1.04	1.04		0	0.15
20/10/2003 EPA 415.1 (T) Ox	Method Blank	0.06	0.1 mg/L as C	0	0			
20/10/2003 EPA 415.1 (T) Ox	LCS - Recovery	9.95	0.1 mg/L as C	0	0	0.995	0.85	1.15
20/10/2003 EPA 415.1 (T) Ox	Check Standard	9.79	0.1 mg/L as C	0	0	0.979	0.85	1.15
20/10/2003 EPA 415.1 (T) Ox	LCS - Recovery	9.94	0.1 mg/L as C	0	0	0.994	0.85	1.15
20/10/2003 EPA 415.1 (T) Ox	Check Standard	9.92	0.1 mg/L as C	0	0	0.992	0.85	1.15
20/10/2003 EPA 415.1 (T) Ox	Check Standard	9.79	0.1 mg/L as C	0	0	0.979	0.85	1.15
20/10/2003 EPA 415.1 (T) Ox	Check Standard	1.04	0.1 mg/L as C	0	0	1.04	0.85	1.15
20/10/2003 EPA 415.1 (T) Ox	Check Standard	1.04	0.1 mg/L as C	0	0	1.04	0.85	1.15
20/10/2003 EPA 415.1 (T) Ox	Check Standard	9.92	0.1 mg/L as C	0	0	0.992	0.85	1.15
20/10/2003 EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.1006	%	9.94	9.95		0	0.15
20/10/2003 EPA 415.1 (T) Ox	RPD - Check Standard	0	%	1.04	1.04		0	0.15
20/10/2003 EPA 415.1 (T) Ox	RPD - Check Standard	1.319	%	9.79	9.92		0	0.15
20/10/2003 EPA 415.1 (T) Ox	Low Level Standard	0.53	0.1 mg/L as C	0	0	1.06	0.5	1.5
27/10/2003 EPA 415.1 (T) Ox	Check Standard	10.28	0.1 mg/L as C	0	0	1.028	0.85	1.15
27/10/2003 EPA 415.1 (T) Ox	Check Standard	1.14	0.1 mg/L as C	0	0	1.14	0.85	1.15
27/10/2003 EPA 415.1 (T) Ox	RPD - Check Standard	4.47	%	10.75	10.28		0	0.15
27/10/2003 EPA 415.1 (T) Ox	RPD - LCS Duplicate	0	%	10.41	10.41		0	0.15
27/10/2003 EPA 415.1 (T) Ox	Check Standard	10.28	0.1 mg/L as C	0	0	1.028	0.85	1.15
27/10/2003 EPA 415.1 (T) Ox	RPD - Check Standard	4.47	%	10.75	10.28		0	0.15
27/10/2003 EPA 415.1 (T) Ox	RPD - LCS Duplicate	0	%	10.79	10.79		0	0.15
27/10/2003 EPA 415.1 (T) Ox	RPD - Check Standard	4.47	%	10.75	10.28		0	0.15
27/10/2003 EPA 415.1 (T) Ox	RPD - LCS Duplicate	0	%	10.79	10.79		0	0.15
27/10/2003 EPA 415.1 (T) Ox	RPD - LCS Duplicate	0	%	10.41	10.41		0	0.15
27/10/2003 EPA 415.1 (T) Ox	Check Standard	10.38	0.1 mg/L as C	0	0	1.038	0.85	1.15
27/10/2003 EPA 415.1 (T) Ox	Low Level Standard	1.2	0.1 mg/L as C	0	0	1.2	0.5	1.5
27/10/2003 EPA 415.1 (T) Ox	Low Level Standard	1.2	0.1 mg/L as C	0	0	1.2	0.5	1.5
27/10/2003 EPA 415.1 (T) Ox	Low Level Standard	1.2	0.1 mg/L as C	0	0	1.2	0.5	1.5
27/10/2003 EPA 415.1 (T) Ox	Low Level Standard	0.64	0.1 mg/L as C	0	0	1.28	0.5	1.5
27/10/2003 EPA 415.1 (T) Ox	LCS - Recovery	10.41	0.1 mg/L as C	0	0	1.041	0.85	1.15
27/10/2003 EPA 415.1 (T) Ox	LCS - Recovery	10.41	0.1 mg/L as C	0	0	1.041	0.85	1.15
27/10/2003 EPA 415.1 (T) Ox	RPD - Check Standard	0.1929	%	10.38	10.36		0	0.15
27/10/2003 EPA 415.1 (T) Ox	Check Standard	10.36	0.1 mg/L as C	0	0	1.036	0.85	1.15
27/10/2003 EPA 415.1 (T) Ox	Method Blank	0.1	0.1 mg/L as C	0	0			
27/10/2003 EPA 415.1 (T) Ox	Check Standard	10.38	0.1 mg/L as C	0	0	1.038	0.85	1.15
27/10/2003 EPA 415.1 (T) Ox	Check Standard	10.28	0.1 mg/L as C	0	0	1.028	0.85	1.15
27/10/2003 EPA 415.1 (T) Ox	RPD - Check Standard	0.1929	%	10.38	10.36		0	0.15

27/10/2003	EPA 415.1 (T) Ox	Low Level Standard	0.64	0.1 mg/L as C	0	0	1.28	0.5	1.5
27/10/2003	EPA 415.1 (T) Ox	LCS - Recovery	10.41	0.1 mg/L as C	0	0	1.041	0.85	1.15
27/10/2003	EPA 415.1 (T) Ox	LCS - Recovery	10.41	0.1 mg/L as C	0	0	1.041	0.85	1.15
27/10/2003	EPA 415.1 (T) Ox	Method Blank	0.1	0.1 mg/L as C	0	0			
27/10/2003	EPA 415.1 (T) Ox	Method Blank	0.07	0.1 mg/L as C	0	0			
27/10/2003	EPA 415.1 (T) Ox	Method Blank	0.1	0.1 mg/L as C	0	0			
27/10/2003	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0	%	10.79	10.79		0	0.15
27/10/2003	EPA 415.1 (T) Ox	Method Blank	0.07	0.1 mg/L as C	0	0			
27/10/2003	EPA 415.1 (T) Ox	Check Standard	10.36	0.1 mg/L as C	0	0	1.036	0.85	1.15
27/10/2003	EPA 415.1 (T) Ox	Check Standard	1.14	0.1 mg/L as C	0	0	1.14	0.85	1.15
03/11/2003	EPA 415.1 (T) Ox	Check Standard	10.41	0.1 mg/L as C	0	0	1.041	0.85	1.15
03/11/2003	EPA 415.1 (T) Ox	RPD - Check Standard	1.062	%	10.3	10.41		0	0.15
03/11/2003	EPA 415.1 (T) Ox	RPD - Check Standard	4.444	%	1.15	1.1		0	0.15
03/11/2003	EPA 415.1 (T) Ox	Method Blank	0.08	0.1 mg/L as C	0	0			
03/11/2003	EPA 415.1 (T) Ox	Check Standard	1.15	0.1 mg/L as C	0	0	1.15	0.85	1.15
03/11/2003	EPA 415.1 (T) Ox	Check Standard	1.1	0.1 mg/L as C	0	0	1.1	0.85	1.15
03/11/2003	EPA 415.1 (T) Ox	LCS - Recovery	10.24	0.1 mg/L as C	0	0	1.024	0.85	1.15
03/11/2003	EPA 415.1 (T) Ox	Check Standard	10.3	0.1 mg/L as C	0	0	1.03	0.85	1.15
03/11/2003	EPA 415.1 (T) Ox	LCS - Recovery	10.31	0.1 mg/L as C	0	0	1.031	0.85	1.15
03/11/2003	EPA 415.1 (T) Ox	Low Level Standard	0.66	0.1 mg/L as C	0	0	1.32	0.5	1.5
03/11/2003	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.6813	%	10.24	10.31		0	0.15
10/11/2003	EPA 415.1 (T) Ox	Check Standard	1.1	0.1 mg/L as C	0	0	1.1	0.85	1.15
10/11/2003	EPA 415.1 (T) Ox	Check Standard	10.14	0.1 mg/L as C	0	0	1.014	0.85	1.15
10/11/2003	EPA 415.1 (T) Ox	Check Standard	10.32	0.1 mg/L as C	0	0	1.032	0.85	1.15
10/11/2003	EPA 415.1 (T) Ox	Method Blank	0.09	0.1 mg/L as C	0	0			
10/11/2003	EPA 415.1 (T) Ox	Low Level Standard	0.59	0.1 mg/L as C	0	0	1.18	0.5	1.5
10/11/2003	EPA 415.1 (T) Ox	RPD - Check Standard	1.76	%	10.14	10.32		0	0.15
10/11/2003	EPA 415.1 (T) Ox	RPD - LCS Duplicate	1.174	%	10.16	10.28		0	0.15
10/11/2003	EPA 415.1 (T) Ox	LCS - Recovery	10.28	0.1 mg/L as C	0	0	1.028	0.85	1.15
10/11/2003	EPA 415.1 (T) Ox	LCS - Recovery	10.16	0.1 mg/L as C	0	0	1.016	0.85	1.15
24/11/2003	EPA 415.1 (T) Ox	Check Standard	10.41	0.1 mg/L as C	0	0	1.041	0.85	1.15
24/11/2003	EPA 415.1 (T) Ox	Low Level Standard	0.67	0.1 mg/L as C	0	0	1.34	0.5	1.5
24/11/2003	EPA 415.1 (T) Ox	LCS - Recovery	10.38	0.1 mg/L as C	0	0	1.038	0.85	1.15
24/11/2003	EPA 415.1 (T) Ox	Check Standard	1.02	0.1 mg/L as C	0	0	1.02	0.85	1.15
24/11/2003	EPA 415.1 (T) Ox	Check Standard	9.61	0.1 mg/L as C	0	0	0.961	0.85	1.15
24/11/2003	EPA 415.1 (T) Ox	LCS - Recovery	10.34	0.1 mg/L as C	0	0	1.034	0.85	1.15
24/11/2003	EPA 415.1 (T) Ox	Method Blank	0.08	0.1 mg/L as C	0	0			

24/11/2003	EPA 415.1 (T) Ox	Check Standard	1.15	0.1 mg/L as C	0	0	1.15	0.85	1.15
01/12/2003	EPA 415.1 (T) Ox	LCS - Recovery	10.18	0.1 mg/L as C	0	0	1.018	0.85	1.15
01/12/2003	EPA 415.1 (T) Ox	Check Standard	9.61	0.1 mg/L as C	0	0	0.961	0.85	1.15
01/12/2003	EPA 415.1 (T) Ox	Check Standard	10.41	0.1 mg/L as C	0	0	1.041	0.85	1.15
01/12/2003	EPA 415.1 (T) Ox	Low Level Standard	0.59	0.1 mg/L as C	0	0	1.18	0.5	1.5
01/12/2003	EPA 415.1 (T) Ox	Check Standard	10.16	0.1 mg/L as C	0	0	1.016	0.85	1.15
01/12/2003	EPA 415.1 (T) Ox	LCS - Recovery	10.21	0.1 mg/L as C	0	0	1.021	0.85	1.15
01/12/2003	EPA 415.1 (T) Ox	Low Level Standard	0.59	0.1 mg/L as C	0	0	1.18	0.5	1.5
01/12/2003	EPA 415.1 (T) Ox	Check Standard	9.61	0.1 mg/L as C	0	0	0.961	0.85	1.15
01/12/2003	EPA 415.1 (T) Ox	Check Standard	1.15	0.1 mg/L as C	0	0	1.15	0.85	1.15
01/12/2003	EPA 415.1 (T) Ox	Check Standard	1.02	0.1 mg/L as C	0	0	1.02	0.85	1.15
01/12/2003	EPA 415.1 (T) Ox	Low Level Standard	0.67	0.1 mg/L as C	0	0	1.34	0.5	1.5
01/12/2003	EPA 415.1 (T) Ox	Check Standard	10.41	0.1 mg/L as C	0	0	1.041	0.85	1.15
01/12/2003	EPA 415.1 (T) Ox	LCS - Recovery	10.21	0.1 mg/L as C	0	0	1.021	0.85	1.15
01/12/2003	EPA 415.1 (T) Ox	Check Standard	1.06	0.1 mg/L as C	0	0	1.06	0.85	1.15
01/12/2003	EPA 415.1 (T) Ox	Check Standard	1.07	0.1 mg/L as C	0	0	1.07	0.85	1.15
01/12/2003	EPA 415.1 (T) Ox	Check Standard	10.16	0.1 mg/L as C	0	0	1.016	0.85	1.15
01/12/2003	EPA 415.1 (T) Ox	Check Standard	10.23	0.1 mg/L as C	0	0	1.023	0.85	1.15
01/12/2003	EPA 415.1 (T) Ox	Check Standard	1.06	0.1 mg/L as C	0	0	1.06	0.85	1.15
01/12/2003	EPA 415.1 (T) Ox	Check Standard	1.07	0.1 mg/L as C	0	0	1.07	0.85	1.15
01/12/2003	EPA 415.1 (T) Ox	LCS - Recovery	10.38	0.1 mg/L as C	0	0	1.038	0.85	1.15
01/12/2003	EPA 415.1 (T) Ox	Check Standard	10.23	0.1 mg/L as C	0	0	1.023	0.85	1.15
01/12/2003	EPA 415.1 (T) Ox	Method Blank	0.08	0.1 mg/L as C	0	0			
01/12/2003	EPA 415.1 (T) Ox	Method Blank	0.08	0.1 mg/L as C	0	0			
01/12/2003	EPA 415.1 (T) Ox	LCS - Recovery	10.34	0.1 mg/L as C	0	0	1.034	0.85	1.15
01/12/2003	EPA 415.1 (T) Ox	LCS - Recovery	10.18	0.1 mg/L as C	0	0	1.018	0.85	1.15
01/12/2003	EPA 415.1 (T) Ox	Check Standard	1.15	0.1 mg/L as C	0	0	1.15	0.85	1.15
01/12/2003	EPA 415.1 (T) Ox	RPD - Check Standard	0.939	%	1.06	1.07		0	0.15
01/12/2003	EPA 415.1 (T) Ox	RPD - Check Standard	0.939	%	1.06	1.07		0	0.15
01/12/2003	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.2943	%	10.18	10.21		0	0.15
01/12/2003	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.2943	%	10.18	10.21		0	0.15
01/12/2003	EPA 415.1 (T) Ox	Check Standard	1.02	0.1 mg/L as C	0	0	1.02	0.85	1.15
01/12/2003	EPA 415.1 (T) Ox	Method Blank	0.08	0.1 mg/L as C	0	0			
01/12/2003	EPA 415.1 (T) Ox	Method Blank	0.08	0.1 mg/L as C	0	0			
01/12/2003	EPA 415.1 (T) Ox	LCS - Recovery	10.34	0.1 mg/L as C	0	0	1.034	0.85	1.15
01/12/2003	EPA 415.1 (T) Ox	RPD - Check Standard	0.6866	%	10.16	10.23		0	0.15
01/12/2003	EPA 415.1 (T) Ox	LCS - Recovery	10.38	0.1 mg/L as C	0	0	1.038	0.85	1.15

01/12/2003	EPA 415.1 (T) Ox	Low Level Standard	0.67	0.1 mg/L as C	0	0	1.34	0.5	1.5
01/12/2003	EPA 415.1 (T) Ox	RPD - Check Standard	0.6866	%	10.16	10.23		0	0.15
08/12/2003	EPA 415.1 (T) Ox	RPD - Check Standard	1.976	%	10.22	10.02		0	0.15
08/12/2003	EPA 415.1 (T) Ox	Check Standard	10.02	0.1 mg/L as C	0	0	1.002	0.85	1.15
08/12/2003	EPA 415.1 (T) Ox	Check Standard	10.22	0.1 mg/L as C	0	0	1.022	0.85	1.15
08/12/2003	EPA 415.1 (T) Ox	Check Standard	1.08	0.1 mg/L as C	0	0	1.08	0.85	1.15
08/12/2003	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.2978	%	10.09	10.06		0	0.15
08/12/2003	EPA 415.1 (T) Ox	Method Blank	0.07	0.1 mg/L as C	0	0			
08/12/2003	EPA 415.1 (T) Ox	Low Level Standard	0.58	0.1 mg/L as C	0	0	1.16	0.5	1.5
08/12/2003	EPA 415.1 (T) Ox	LCS - Recovery	10.06	0.1 mg/L as C	0	0	1.006	0.85	1.15
08/12/2003	EPA 415.1 (T) Ox	LCS - Recovery	10.09	0.1 mg/L as C	0	0	1.009	0.85	1.15
15/12/2003	EPA 415.1 (T) Ox	RPD - Check Standard	0.1969	%	10.15	10.17		0	0.15
15/12/2003	EPA 415.1 (T) Ox	RPD - Check Standard	1.766	%	10.1	10.28		0	0.15
15/12/2003	EPA 415.1 (T) Ox	Check Standard	10.28	0.1 mg/L as C	0	0	1.028	0.85	1.15
15/12/2003	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0	%	10.16	10.16		0	0.15
15/12/2003	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.5911	%	10.12	10.18		0	0.15
15/12/2003	EPA 415.1 (T) Ox	Low Level Standard	0.62	0.1 mg/L as C	0	0	1.24	0.5	1.5
15/12/2003	EPA 415.1 (T) Ox	Check Standard	1.1	0.1 mg/L as C	0	0	1.1	0.85	1.15
15/12/2003	EPA 415.1 (T) Ox	Method Blank	0.08	0.1 mg/L as C	0	0			
15/12/2003	EPA 415.1 (T) Ox	Check Standard	10.1	0.1 mg/L as C	0	0	1.01	0.85	1.15
15/12/2003	EPA 415.1 (T) Ox	LCS - Recovery	10.16	0.1 mg/L as C	0	0	1.016	0.85	1.15
15/12/2003	EPA 415.1 (T) Ox	Method Blank	0.09	0.1 mg/L as C	0	0			
15/12/2003	EPA 415.1 (T) Ox	Check Standard	10.17	0.1 mg/L as C	0	0	1.017	0.85	1.15
15/12/2003	EPA 415.1 (T) Ox	LCS - Recovery	10.12	0.1 mg/L as C	0	0	1.012	0.85	1.15
15/12/2003	EPA 415.1 (T) Ox	LCS - Recovery	10.18	0.1 mg/L as C	0	0	1.018	0.85	1.15
15/12/2003	EPA 415.1 (T) Ox	Low Level Standard	0.64	0.1 mg/L as C	0	0	1.28	0.5	1.5
15/12/2003	EPA 415.1 (T) Ox	Check Standard	10.15	0.1 mg/L as C	0	0	1.015	0.85	1.15
15/12/2003	EPA 415.1 (T) Ox	LCS - Recovery	10.16	0.1 mg/L as C	0	0	1.016	0.85	1.15
22/12/2003	EPA 415.1 (T) Ox	RPD - Check Standard	0.7937	%	10.12	10.04		0	0.15
22/12/2003	EPA 415.1 (T) Ox	Check Standard	1.01	0.1 mg/L as C	0	0	1.01	0.85	1.15
22/12/2003	EPA 415.1 (T) Ox	Check Standard	10.12	0.1 mg/L as C	0	0	1.012	0.85	1.15
22/12/2003	EPA 415.1 (T) Ox	Check Standard	10.04	0.1 mg/L as C	0	0	1.004	0.85	1.15
22/12/2003	EPA 415.1 (T) Ox	Method Blank	0.08	0.1 mg/L as C	0	0			
22/12/2003	EPA 415.1 (T) Ox	Check Standard	1.09	0.1 mg/L as C	0	0	1.09	0.85	1.15
22/12/2003	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.09877	%	10.13	10.12		0	0.15
22/12/2003	EPA 415.1 (T) Ox	RPD - Check Standard	7.619	%	1.09	1.01		0	0.15
22/12/2003	EPA 415.1 (T) Ox	LCS - Recovery	10.12	0.1 mg/L as C	0	0	1.012	0.85	1.15

22/12/2003	EPA 415.1 (T) Ox	LCS - Recovery	10.13	0.1 mg/L as C	0	0	1.013	0.85	1.15
22/12/2003	EPA 415.1 (T) Ox	Low Level Standard	0.58	0.1 mg/L as C	0	0	1.16	0.5	1.5
29/12/2003	EPA 415.1 (T) Ox	LCS - Recovery	10.12	0.1 mg/L as C	0	0	1.012	0.85	1.15
29/12/2003	EPA 415.1 (T) Ox	LCS - Recovery	10.13	0.1 mg/L as C	0	0	1.013	0.85	1.15
29/12/2003	EPA 415.1 (T) Ox	Low Level Standard	0.58	0.1 mg/L as C	0	0	1.16	0.5	1.5
29/12/2003	EPA 415.1 (T) Ox	Method Blank	0.08	0.1 mg/L as C	0	0			
29/12/2003	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.09877	%	10.13	10.12		0	0.15
29/12/2003	EPA 415.1 (T) Ox	Check Standard	1.09	0.1 mg/L as C	0	0	1.09	0.85	1.15
29/12/2003	EPA 415.1 (T) Ox	Check Standard	1.01	0.1 mg/L as C	0	0	1.01	0.85	1.15
29/12/2003	EPA 415.1 (T) Ox	Check Standard	10.12	0.1 mg/L as C	0	0	1.012	0.85	1.15
29/12/2003	EPA 415.1 (T) Ox	Check Standard	10.04	0.1 mg/L as C	0	0	1.004	0.85	1.15
29/12/2003	EPA 415.1 (T) Ox	RPD - Check Standard	0.7937	%	10.12	10.04		0	0.15
29/12/2003	EPA 415.1 (T) Ox	RPD - Check Standard	7.619	%	1.09	1.01		0	0.15
07/01/2004	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.09857	%	10.15	10.14		0	0.15
07/01/2004	EPA 415.1 (T) Ox	Low Level Standard	0.6	0.1 mg/L as C	0	0	1.2	0.5	1.5
07/01/2004	EPA 415.1 (T) Ox	RPD - Check Standard	2.537	%	10.12	10.38		0	0.15
07/01/2004	EPA 415.1 (T) Ox	LCS - Recovery	10.15	0.1 mg/L as C	0	0	1.015	0.85	1.15
07/01/2004	EPA 415.1 (T) Ox	Method Blank	0.09	0.1 mg/L as C	0	0			
07/01/2004	EPA 415.1 (T) Ox	Check Standard	10.38	0.1 mg/L as C	0	0	1.038	0.85	1.15
07/01/2004	EPA 415.1 (T) Ox	Check Standard	10.12	0.1 mg/L as C	0	0	1.012	0.85	1.15
07/01/2004	EPA 415.1 (T) Ox	Check Standard	1.09	0.1 mg/L as C	0	0	1.09	0.85	1.15
07/01/2004	EPA 415.1 (T) Ox	LCS - Recovery	10.14	0.1 mg/L as C	0	0	1.014	0.85	1.15
12/01/2004	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.3895	%	10.25	10.29		0	0.15
12/01/2004	EPA 415.1 (T) Ox	Check Standard	10.2	0.1 mg/L as C	0	0	1.02	0.85	1.15
12/01/2004	EPA 415.1 (T) Ox	Check Standard	10.24	0.1 mg/L as C	0	0	1.024	0.85	1.15
12/01/2004	EPA 415.1 (T) Ox	Check Standard	1.11	0.1 mg/L as C	0	0	1.11	0.85	1.15
12/01/2004	EPA 415.1 (T) Ox	Check Standard	1.06	0.1 mg/L as C	0	0	1.06	0.85	1.15
12/01/2004	EPA 415.1 (T) Ox	RPD - Check Standard	0.3914	%	10.24	10.2		0	0.15
12/01/2004	EPA 415.1 (T) Ox	Low Level Standard	0.62	0.1 mg/L as C	0	0	1.24	0.5	1.5
12/01/2004	EPA 415.1 (T) Ox	LCS - Recovery	10.29	0.1 mg/L as C	0	0	1.029	0.85	1.15
12/01/2004	EPA 415.1 (T) Ox	LCS - Recovery	10.25	0.1 mg/L as C	0	0	1.025	0.85	1.15
12/01/2004	EPA 415.1 (T) Ox	Method Blank	0.08	0.1 mg/L as C	0	0			
12/01/2004	EPA 415.1 (T) Ox	RPD - Check Standard	4.608	%	1.06	1.11		0	0.15
20/01/2004	EPA 415.1 (T) Ox	RPD - Check Standard	1.105	%	9.9	10.01		0	0.15
20/01/2004	EPA 415.1 (T) Ox	Check Standard	10.01	0.1 mg/L as C	0	0	1.001	0.85	1.15
20/01/2004	EPA 415.1 (T) Ox	Check Standard	9.9	0.1 mg/L as C	0	0	0.99	0.85	1.15
20/01/2004	EPA 415.1 (T) Ox	Check Standard	1.1	0.1 mg/L as C	0	0	1.1	0.85	1.15

20/01/2004	EPA 415.1 (T) Ox	Check Standard	1.08	0.1 mg/L as C	0	0	1.08	0.85	1.15
20/01/2004	EPA 415.1 (T) Ox	Low Level Standard	0.59	0.1 mg/L as C	0	0	1.18	0.5	1.5
20/01/2004	EPA 415.1 (T) Ox	LCS - Recovery	9.87	0.1 mg/L as C	0	0	0.987	0.85	1.15
20/01/2004	EPA 415.1 (T) Ox	LCS - Recovery	9.94	0.1 mg/L as C	0	0	0.994	0.85	1.15
20/01/2004	EPA 415.1 (T) Ox	RPD - Check Standard	1.835	%	1.08	1.1		0	0.15
20/01/2004	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.7067	%	9.94	9.87		0	0.15
20/01/2004	EPA 415.1 (T) Ox	Method Blank	0.07	0.1 mg/L as C	0	0			
26/01/2004	EPA 415.1 (T) Ox	LCS - Recovery	9.79	0.1 mg/L as C	0	0	0.979	0.85	1.15
26/01/2004	EPA 415.1 (T) Ox	RPD - Check Standard	0.4866	%	10.3	10.25		0	0.15
26/01/2004	EPA 415.1 (T) Ox	RPD - LCS Duplicate	1.167	%	10.22	10.34		0	0.15
26/01/2004	EPA 415.1 (T) Ox	RPD - Check Standard	0	%	1.02	1.02		0	0.15
26/01/2004	EPA 415.1 (T) Ox	RPD - Check Standard	1.02	%	9.85	9.75		0	0.15
26/01/2004	EPA 415.1 (T) Ox	RPD - LCS Duplicate	1.117	%	9.79	9.9		0	0.15
26/01/2004	EPA 415.1 (T) Ox	Low Level Standard	0.53	0.1 mg/L as C	0	0	1.06	0.5	1.5
26/01/2004	EPA 415.1 (T) Ox	LCS - Recovery	9.9	0.1 mg/L as C	0	0	0.99	0.85	1.15
26/01/2004	EPA 415.1 (T) Ox	LCS - Recovery	10.34	0.1 mg/L as C	0	0	1.034	0.85	1.15
26/01/2004	EPA 415.1 (T) Ox	Check Standard	1.02	0.1 mg/L as C	0	0	1.02	0.85	1.15
26/01/2004	EPA 415.1 (T) Ox	Check Standard	1.02	0.1 mg/L as C	0	0	1.02	0.85	1.15
26/01/2004	EPA 415.1 (T) Ox	Method Blank	0.07	0.1 mg/L as C	0	0			
26/01/2004	EPA 415.1 (T) Ox	Check Standard	9.85	0.1 mg/L as C	0	0	0.985	0.85	1.15
26/01/2004	EPA 415.1 (T) Ox	Check Standard	10.3	0.1 mg/L as C	0	0	1.03	0.85	1.15
26/01/2004	EPA 415.1 (T) Ox	Check Standard	1.12	0.1 mg/L as C	0	0	1.12	0.85	1.15
26/01/2004	EPA 415.1 (T) Ox	Check Standard	10.25	0.1 mg/L as C	0	0	1.025	0.85	1.15
26/01/2004	EPA 415.1 (T) Ox	LCS - Recovery	10.22	0.1 mg/L as C	0	0	1.022	0.85	1.15
26/01/2004	EPA 415.1 (T) Ox	Low Level Standard	0.59	0.1 mg/L as C	0	0	1.18	0.5	1.5
26/01/2004	EPA 415.1 (T) Ox	Method Blank	0.06	0.1 mg/L as C	0	0			
26/01/2004	EPA 415.1 (T) Ox	Check Standard	9.75	0.1 mg/L as C	0	0	0.975	0.85	1.15
02/02/2004	EPA 415.1 (T) Ox	RPD - LCS Duplicate	2.382	%	9.77	9.54		0	0.15
02/02/2004	EPA 415.1 (T) Ox	Method Blank	0.09	0.1 mg/L as C	0	0			
02/02/2004	EPA 415.1 (T) Ox	LCS - Recovery	9.77	0.1 mg/L as C	0	0	0.977	0.85	1.15
02/02/2004	EPA 415.1 (T) Ox	Low Level Standard	0.57	0.1 mg/L as C	0	0	1.14	0.5	1.5
02/02/2004	EPA 415.1 (T) Ox	RPD - Check Standard	4.036	%	9.71	10.11		0	0.15
02/02/2004	EPA 415.1 (T) Ox	Check Standard	1.05	0.1 mg/L as C	0	0	1.05	0.85	1.15
02/02/2004	EPA 415.1 (T) Ox	Check Standard	10.11	0.1 mg/L as C	0	0	1.011	0.85	1.15
02/02/2004	EPA 415.1 (T) Ox	Check Standard	9.71	0.1 mg/L as C	0	0	0.971	0.85	1.15
02/02/2004	EPA 415.1 (T) Ox	LCS - Recovery	9.54	0.1 mg/L as C	0	0	0.954	0.85	1.15
09/02/2004	EPA 415.1 (T) Ox	RPD - LCS Duplicate	1.007	%	9.88	9.98		0	0.15

09/02/2004	EPA 415.1 (T) Ox	Check Standard	9.95	0.1 mg/L as C	0	0	0.995	0.85	1.15
09/02/2004	EPA 415.1 (T) Ox	Check Standard	1.04	0.1 mg/L as C	0	0	1.04	0.85	1.15
09/02/2004	EPA 415.1 (T) Ox	Check Standard	9.76	0.1 mg/L as C	0	0	0.976	0.85	1.15
09/02/2004	EPA 415.1 (T) Ox	LCS - Recovery	9.98	0.1 mg/L as C	0	0	0.998	0.85	1.15
09/02/2004	EPA 415.1 (T) Ox	Method Blank	0.07	0.1 mg/L as C	0	0			
09/02/2004	EPA 415.1 (T) Ox	Low Level Standard	0.54	0.1 mg/L as C	0	0	1.08	0.5	1.5
09/02/2004	EPA 415.1 (T) Ox	LCS - Recovery	9.88	0.1 mg/L as C	0	0	0.988	0.85	1.15
09/02/2004	EPA 415.1 (T) Ox	RPD - Check Standard	1.928	%	9.76	9.95		0	0.15
17/02/2004	EPA 415.1 (T) Ox	Check Standard	10.14	0.1 mg/L as C	0	0	1.014	0.85	1.15
17/02/2004	EPA 415.1 (T) Ox	Check Standard	1.15	0.1 mg/L as C	0	0	1.15	0.85	1.15
17/02/2004	EPA 415.1 (T) Ox	Check Standard	10.23	0.1 mg/L as C	0	0	1.023	0.85	1.15
17/02/2004	EPA 415.1 (T) Ox	Low Level Standard	0.64	0.1 mg/L as C	0	0	1.28	0.5	1.5
17/02/2004	EPA 415.1 (T) Ox	Check Standard	10.23	0.1 mg/L as C	0	0	1.023	0.85	1.15
17/02/2004	EPA 415.1 (T) Ox	RPD - Check Standard	0.8837	%	10.14	10.23		0	0.15
17/02/2004	EPA 415.1 (T) Ox	Check Standard	1.15	0.1 mg/L as C	0	0	1.15	0.85	1.15
17/02/2004	EPA 415.1 (T) Ox	Low Level Standard	0.64	0.1 mg/L as C	0	0	1.28	0.5	1.5
17/02/2004	EPA 415.1 (T) Ox	LCS - Recovery	10.17	0.1 mg/L as C	0	0	1.017	0.85	1.15
17/02/2004	EPA 415.1 (T) Ox	LCS - Recovery	10.07	0.1 mg/L as C	0	0	1.007	0.85	1.15
17/02/2004	EPA 415.1 (T) Ox	Check Standard	10.14	0.1 mg/L as C	0	0	1.014	0.85	1.15
17/02/2004	EPA 415.1 (T) Ox	RPD - Check Standard	0.8837	%	10.14	10.23		0	0.15
17/02/2004	EPA 415.1 (T) Ox	LCS - Recovery	10.17	0.1 mg/L as C	0	0	1.017	0.85	1.15
17/02/2004	EPA 415.1 (T) Ox	LCS - Recovery	10.07	0.1 mg/L as C	0	0	1.007	0.85	1.15
17/02/2004	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.9881	%	10.07	10.17		0	0.15
17/02/2004	EPA 415.1 (T) Ox	Method Blank	0.08	0.1 mg/L as C	0	0			
17/02/2004	EPA 415.1 (T) Ox	Method Blank	0.08	0.1 mg/L as C	0	0			
17/02/2004	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.9881	%	10.07	10.17		0	0.15
23/02/2004	EPA 415.1 (T) Ox	LCS - Recovery	9.67	0.1 mg/L as C	0	0	0.967	0.85	1.15
23/02/2004	EPA 415.1 (T) Ox	Low Level Standard	0.54	0.1 mg/L as C	0	0	1.08	0.5	1.5
23/02/2004	EPA 415.1 (T) Ox	LCS - Recovery	9.58	0.1 mg/L as C	0	0	0.958	0.85	1.15
23/02/2004	EPA 415.1 (T) Ox	RPD - Check Standard	5.637	%	9.85	9.31		0	0.15
23/02/2004	EPA 415.1 (T) Ox	RPD - Check Standard	10.1	%	1.04	0.94		0	0.15
23/02/2004	EPA 415.1 (T) Ox	RPD - LCS Duplicate	1.211	%	9.85	9.97		0	0.15
23/02/2004	EPA 415.1 (T) Ox	Method Blank	0.08	0.1 mg/L as C	0	0			
23/02/2004	EPA 415.1 (T) Ox	Check Standard	9.31	0.1 mg/L as C	0	0	0.931	0.85	1.15
23/02/2004	EPA 415.1 (T) Ox	RPD - Check Standard	2.105	%	0.96	0.94		0	0.15
23/02/2004	EPA 415.1 (T) Ox	LCS - Recovery	9.85	0.1 mg/L as C	0	0	0.985	0.85	1.15
23/02/2004	EPA 415.1 (T) Ox	LCS - Recovery	9.97	0.1 mg/L as C	0	0	0.997	0.85	1.15

23/02/2004	EPA 415.1 (T) Ox	Low Level Standard	0.58	0.1 mg/L as C	0	0	1.16	0.5	1.5
23/02/2004	EPA 415.1 (T) Ox	RPD - Check Standard	0	%	9.31	9.31		0	0.15
23/02/2004	EPA 415.1 (T) Ox	Check Standard	1.04	0.1 mg/L as C	0	0	1.04	0.85	1.15
23/02/2004	EPA 415.1 (T) Ox	Check Standard	9.31	0.1 mg/L as C	0	0	0.931	0.85	1.15
23/02/2004	EPA 415.1 (T) Ox	Check Standard	0.94	0.1 mg/L as C	0	0	0.94	0.85	1.15
23/02/2004	EPA 415.1 (T) Ox	Check Standard	0.96	0.1 mg/L as C	0	0	0.96	0.85	1.15
23/02/2004	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.9351	%	9.58	9.67		0	0.15
23/02/2004	EPA 415.1 (T) Ox	Check Standard	0.94	0.1 mg/L as C	0	0	0.94	0.85	1.15
23/02/2004	EPA 415.1 (T) Ox	Check Standard	9.85	0.1 mg/L as C	0	0	0.985	0.85	1.15
23/02/2004	EPA 415.1 (T) Ox	Method Blank	0.08	0.1 mg/L as C	0	0			
23/02/2004	EPA 415.1 (T) Ox	Check Standard	9.31	0.1 mg/L as C	0	0	0.931	0.85	1.15
01/03/2004	EPA 415.1 (T) Ox	Check Standard	9.36	0.1 mg/L as C	0	0	0.936	0.85	1.15
01/03/2004	EPA 415.1 (T) Ox	Check Standard	9.36	0.1 mg/L as C	0	0	0.936	0.85	1.15
01/03/2004	EPA 415.1 (T) Ox	Check Standard	9.64	0.1 mg/L as C	0	0	0.964	0.85	1.15
01/03/2004	EPA 415.1 (T) Ox	Check Standard	1.04	0.1 mg/L as C	0	0	1.04	0.85	1.15
01/03/2004	EPA 415.1 (T) Ox	Check Standard	1.04	0.1 mg/L as C	0	0	1.04	0.85	1.15
01/03/2004	EPA 415.1 (T) Ox	Check Standard	9.64	0.1 mg/L as C	0	0	0.964	0.85	1.15
01/03/2004	EPA 415.1 (T) Ox	Check Standard	1.04	0.1 mg/L as C	0	0	1.04	0.85	1.15
01/03/2004	EPA 415.1 (T) Ox	Check Standard	1.06	0.1 mg/L as C	0	0	1.06	0.85	1.15
01/03/2004	EPA 415.1 (T) Ox	Check Standard	9.64	0.1 mg/L as C	0	0	0.964	0.85	1.15
01/03/2004	EPA 415.1 (T) Ox	Check Standard	9.36	0.1 mg/L as C	0	0	0.936	0.85	1.15
01/03/2004	EPA 415.1 (T) Ox	Check Standard	1.04	0.1 mg/L as C	0	0	1.04	0.85	1.15
01/03/2004	EPA 415.1 (T) Ox	Check Standard	1.06	0.1 mg/L as C	0	0	1.06	0.85	1.15
01/03/2004	EPA 415.1 (T) Ox	Check Standard	9.64	0.1 mg/L as C	0	0	0.964	0.85	1.15
01/03/2004	EPA 415.1 (T) Ox	Check Standard	1.06	0.1 mg/L as C	0	0	1.06	0.85	1.15
01/03/2004	EPA 415.1 (T) Ox	Check Standard	9.36	0.1 mg/L as C	0	0	0.936	0.85	1.15
01/03/2004	EPA 415.1 (T) Ox	RPD - Check Standard	2.947	%	9.36	9.64		0	0.15
01/03/2004	EPA 415.1 (T) Ox	Check Standard	1.06	0.1 mg/L as C	0	0	1.06	0.85	1.15
01/03/2004	EPA 415.1 (T) Ox	LCS - Recovery	9.42	0.1 mg/L as C	0	0	0.942	0.85	1.15
01/03/2004	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.5322	%	9.37	9.42		0	0.15
01/03/2004	EPA 415.1 (T) Ox	Low Level Standard	0.58	0.1 mg/L as C	0	0	1.16	0.5	1.5
01/03/2004	EPA 415.1 (T) Ox	RPD - Check Standard	2.947	%	9.36	9.64		0	0.15
01/03/2004	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.5322	%	9.37	9.42		0	0.15
01/03/2004	EPA 415.1 (T) Ox	RPD - Check Standard	1.905	%	1.06	1.04		0	0.15
01/03/2004	EPA 415.1 (T) Ox	RPD - Check Standard	1.905	%	1.06	1.04		0	0.15
01/03/2004	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.5322	%	9.37	9.42		0	0.15
01/03/2004	EPA 415.1 (T) Ox	RPD - Check Standard	2.947	%	9.36	9.64		0	0.15

01/03/2004	EPA 415.1 (T) Ox	RPD - Check Standard	2.947	%	9.36	9.64	0	0.15	
01/03/2004	EPA 415.1 (T) Ox	LCS - Recovery	9.37	0.1 mg/L as C	0	0	0.937	0.85	1.15
01/03/2004	EPA 415.1 (T) Ox	RPD - Check Standard	1.905	%	1.06	1.04	0	0.15	
01/03/2004	EPA 415.1 (T) Ox	Low Level Standard	0.58	0.1 mg/L as C	0	0	1.16	0.5	1.5
01/03/2004	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.5322	%	9.37	9.42	0	0.15	
01/03/2004	EPA 415.1 (T) Ox	LCS - Recovery	9.42	0.1 mg/L as C	0	0	0.942	0.85	1.15
01/03/2004	EPA 415.1 (T) Ox	LCS - Recovery	9.37	0.1 mg/L as C	0	0	0.937	0.85	1.15
01/03/2004	EPA 415.1 (T) Ox	LCS - Recovery	9.37	0.1 mg/L as C	0	0	0.937	0.85	1.15
01/03/2004	EPA 415.1 (T) Ox	RPD - Check Standard	1.905	%	1.06	1.04	0	0.15	
01/03/2004	EPA 415.1 (T) Ox	Method Blank	0.09	0.1 mg/L as C	0	0			
01/03/2004	EPA 415.1 (T) Ox	Low Level Standard	0.58	0.1 mg/L as C	0	0	1.16	0.5	1.5
01/03/2004	EPA 415.1 (T) Ox	Low Level Standard	0.58	0.1 mg/L as C	0	0	1.16	0.5	1.5
01/03/2004	EPA 415.1 (T) Ox	Method Blank	0.09	0.1 mg/L as C	0	0			
01/03/2004	EPA 415.1 (T) Ox	Method Blank	0.09	0.1 mg/L as C	0	0			
01/03/2004	EPA 415.1 (T) Ox	Method Blank	0.09	0.1 mg/L as C	0	0			
01/03/2004	EPA 415.1 (T) Ox	LCS - Recovery	9.37	0.1 mg/L as C	0	0	0.937	0.85	1.15
01/03/2004	EPA 415.1 (T) Ox	LCS - Recovery	9.42	0.1 mg/L as C	0	0	0.942	0.85	1.15
01/03/2004	EPA 415.1 (T) Ox	LCS - Recovery	9.42	0.1 mg/L as C	0	0	0.942	0.85	1.15
08/03/2004	EPA 415.1 (T) Ox	Low Level Standard	0.57	0.1 mg/L as C	0	0	1.14	0.5	1.5
08/03/2004	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.2112	%	9.48	9.46	0	0.15	
08/03/2004	EPA 415.1 (T) Ox	RPD - Check Standard	4.608	%	1.06	1.11	0	0.15	
08/03/2004	EPA 415.1 (T) Ox	RPD - Check Standard	1.145	%	9.55	9.66	0	0.15	
08/03/2004	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.2112	%	9.48	9.46	0	0.15	
08/03/2004	EPA 415.1 (T) Ox	LCS - Recovery	9.46	0.1 mg/L as C	0	0	0.946	0.85	1.15
08/03/2004	EPA 415.1 (T) Ox	LCS - Recovery	9.48	0.1 mg/L as C	0	0	0.948	0.85	1.15
08/03/2004	EPA 415.1 (T) Ox	Low Level Standard	0.57	0.1 mg/L as C	0	0	1.14	0.5	1.5
08/03/2004	EPA 415.1 (T) Ox	LCS - Recovery	9.46	0.1 mg/L as C	0	0	0.946	0.85	1.15
08/03/2004	EPA 415.1 (T) Ox	LCS - Recovery	9.48	0.1 mg/L as C	0	0	0.948	0.85	1.15
08/03/2004	EPA 415.1 (T) Ox	Low Level Standard	0.57	0.1 mg/L as C	0	0	1.14	0.5	1.5
08/03/2004	EPA 415.1 (T) Ox	LCS - Recovery	9.46	0.1 mg/L as C	0	0	0.946	0.85	1.15
08/03/2004	EPA 415.1 (T) Ox	RPD - Check Standard	1.145	%	9.55	9.66	0	0.15	
08/03/2004	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.2112	%	9.48	9.46	0	0.15	
08/03/2004	EPA 415.1 (T) Ox	LCS - Recovery	9.48	0.1 mg/L as C	0	0	0.948	0.85	1.15
08/03/2004	EPA 415.1 (T) Ox	RPD - Check Standard	1.145	%	9.55	9.66	0	0.15	
08/03/2004	EPA 415.1 (T) Ox	RPD - Check Standard	4.608	%	1.06	1.11	0	0.15	
08/03/2004	EPA 415.1 (T) Ox	Check Standard	9.66	0.1 mg/L as C	0	0	0.966	0.85	1.15
08/03/2004	EPA 415.1 (T) Ox	RPD - Check Standard	4.608	%	1.06	1.11	0	0.15	

08/03/2004	EPA 415.1 (T) Ox	Check Standard	9.66	0.1 mg/L as C	0	0	0.966	0.85	1.15
08/03/2004	EPA 415.1 (T) Ox	Check Standard	9.55	0.1 mg/L as C	0	0	0.955	0.85	1.15
08/03/2004	EPA 415.1 (T) Ox	Check Standard	1.11	0.1 mg/L as C	0	0	1.11	0.85	1.15
08/03/2004	EPA 415.1 (T) Ox	Check Standard	1.06	0.1 mg/L as C	0	0	1.06	0.85	1.15
08/03/2004	EPA 415.1 (T) Ox	Check Standard	9.66	0.1 mg/L as C	0	0	0.966	0.85	1.15
08/03/2004	EPA 415.1 (T) Ox	Method Blank	0.06	0.1 mg/L as C	0	0			
08/03/2004	EPA 415.1 (T) Ox	Method Blank	0.06	0.1 mg/L as C	0	0			
08/03/2004	EPA 415.1 (T) Ox	Method Blank	0.06	0.1 mg/L as C	0	0			
08/03/2004	EPA 415.1 (T) Ox	Check Standard	9.55	0.1 mg/L as C	0	0	0.955	0.85	1.15
08/03/2004	EPA 415.1 (T) Ox	Check Standard	1.11	0.1 mg/L as C	0	0	1.11	0.85	1.15
08/03/2004	EPA 415.1 (T) Ox	Check Standard	1.06	0.1 mg/L as C	0	0	1.06	0.85	1.15
08/03/2004	EPA 415.1 (T) Ox	Check Standard	9.55	0.1 mg/L as C	0	0	0.955	0.85	1.15
08/03/2004	EPA 415.1 (T) Ox	Check Standard	1.06	0.1 mg/L as C	0	0	1.06	0.85	1.15
08/03/2004	EPA 415.1 (T) Ox	Check Standard	1.11	0.1 mg/L as C	0	0	1.11	0.85	1.15
15/03/2004	EPA 415.1 (T) Ox	Method Blank	0.08	0.1 mg/L as C	0	0			
15/03/2004	EPA 415.1 (T) Ox	Method Blank	0.08	0.1 mg/L as C	0	0			
15/03/2004	EPA 415.1 (T) Ox	LCS - Recovery	9.43	0.1 mg/L as C	0	0	0.943	0.85	1.15
15/03/2004	EPA 415.1 (T) Ox	LCS - Recovery	9.48	0.1 mg/L as C	0	0	0.948	0.85	1.15
15/03/2004	EPA 415.1 (T) Ox	Low Level Standard	0.56	0.1 mg/L as C	0	0	1.12	0.5	1.5
15/03/2004	EPA 415.1 (T) Ox	LCS - Recovery	9.43	0.1 mg/L as C	0	0	0.943	0.85	1.15
15/03/2004	EPA 415.1 (T) Ox	LCS - Recovery	9.48	0.1 mg/L as C	0	0	0.948	0.85	1.15
15/03/2004	EPA 415.1 (T) Ox	Check Standard	9.73	0.1 mg/L as C	0	0	0.973	0.85	1.15
15/03/2004	EPA 415.1 (T) Ox	Check Standard	1.03	0.1 mg/L as C	0	0	1.03	0.85	1.15
15/03/2004	EPA 415.1 (T) Ox	Low Level Standard	0.56	0.1 mg/L as C	0	0	1.12	0.5	1.5
15/03/2004	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.5288	%	9.48	9.43		0	0.15
15/03/2004	EPA 415.1 (T) Ox	RPD - Check Standard	0.9662	%	1.03	1.04		0	0.15
15/03/2004	EPA 415.1 (T) Ox	RPD - Check Standard	1.972	%	9.54	9.73		0	0.15
15/03/2004	EPA 415.1 (T) Ox	Check Standard	9.73	0.1 mg/L as C	0	0	0.973	0.85	1.15
15/03/2004	EPA 415.1 (T) Ox	RPD - Check Standard	0.9662	%	1.03	1.04		0	0.15
15/03/2004	EPA 415.1 (T) Ox	Check Standard	1.04	0.1 mg/L as C	0	0	1.04	0.85	1.15
15/03/2004	EPA 415.1 (T) Ox	Check Standard	9.54	0.1 mg/L as C	0	0	0.954	0.85	1.15
15/03/2004	EPA 415.1 (T) Ox	RPD - Check Standard	1.972	%	9.54	9.73		0	0.15
15/03/2004	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.5288	%	9.48	9.43		0	0.15
15/03/2004	EPA 415.1 (T) Ox	Check Standard	1.04	0.1 mg/L as C	0	0	1.04	0.85	1.15
15/03/2004	EPA 415.1 (T) Ox	Check Standard	1.03	0.1 mg/L as C	0	0	1.03	0.85	1.15
15/03/2004	EPA 415.1 (T) Ox	Check Standard	9.54	0.1 mg/L as C	0	0	0.954	0.85	1.15
22/03/2004	EPA 415.1 (T) Ox	LCS - Recovery	9.48	0.1 mg/L as C	0	0	0.948	0.85	1.15

22/03/2004	EPA 415.1 (T) Ox	LCS - Recovery	9.24	0.1 mg/L as C	0	0	0.924	0.85	1.15
22/03/2004	EPA 415.1 (T) Ox	LCS - Recovery	9.3	0.1 mg/L as C	0	0	0.93	0.85	1.15
22/03/2004	EPA 415.1 (T) Ox	Low Level Standard	0.54	0.1 mg/L as C	0	0	1.08	0.5	1.5
22/03/2004	EPA 415.1 (T) Ox	Check Standard	9.73	0.1 mg/L as C	0	0	0.973	0.85	1.15
22/03/2004	EPA 415.1 (T) Ox	Method Blank	0.08	0.1 mg/L as C	0	0			
22/03/2004	EPA 415.1 (T) Ox	Method Blank	0.08	0.1 mg/L as C	0	0			
22/03/2004	EPA 415.1 (T) Ox	Check Standard	1.06	0.1 mg/L as C	0	0	1.06	0.85	1.15
22/03/2004	EPA 415.1 (T) Ox	Check Standard	9.54	0.1 mg/L as C	0	0	0.954	0.85	1.15
22/03/2004	EPA 415.1 (T) Ox	Check Standard	9.73	0.1 mg/L as C	0	0	0.973	0.85	1.15
22/03/2004	EPA 415.1 (T) Ox	Check Standard	1.03	0.1 mg/L as C	0	0	1.03	0.85	1.15
22/03/2004	EPA 415.1 (T) Ox	Method Blank	0.05	0.1 mg/L as C	0	0			
22/03/2004	EPA 415.1 (T) Ox	RPD - Check Standard	1.972	%	9.54	9.73		0	0.15
22/03/2004	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.5288	%	9.48	9.43		0	0.15
22/03/2004	EPA 415.1 (T) Ox	Check Standard	1.04	0.1 mg/L as C	0	0	1.04	0.85	1.15
22/03/2004	EPA 415.1 (T) Ox	Check Standard	1.04	0.1 mg/L as C	0	0	1.04	0.85	1.15
22/03/2004	EPA 415.1 (T) Ox	RPD - Check Standard	5.825	%	1	1.06		0	0.15
22/03/2004	EPA 415.1 (T) Ox	RPD - Check Standard	1.972	%	9.54	9.73		0	0.15
22/03/2004	EPA 415.1 (T) Ox	Check Standard	9.54	0.1 mg/L as C	0	0	0.954	0.85	1.15
22/03/2004	EPA 415.1 (T) Ox	Check Standard	1.03	0.1 mg/L as C	0	0	1.03	0.85	1.15
22/03/2004	EPA 415.1 (T) Ox	RPD - Check Standard	1.598	%	9.31	9.46		0	0.15
22/03/2004	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.6472	%	9.24	9.3		0	0.15
22/03/2004	EPA 415.1 (T) Ox	Check Standard	9.46	0.1 mg/L as C	0	0	0.946	0.85	1.15
22/03/2004	EPA 415.1 (T) Ox	RPD - Check Standard	0.9662	%	1.03	1.04		0	0.15
22/03/2004	EPA 415.1 (T) Ox	RPD - Check Standard	0.9662	%	1.03	1.04		0	0.15
22/03/2004	EPA 415.1 (T) Ox	LCS - Recovery	9.43	0.1 mg/L as C	0	0	0.943	0.85	1.15
22/03/2004	EPA 415.1 (T) Ox	Low Level Standard	0.56	0.1 mg/L as C	0	0	1.12	0.5	1.5
22/03/2004	EPA 415.1 (T) Ox	LCS - Recovery	9.43	0.1 mg/L as C	0	0	0.943	0.85	1.15
22/03/2004	EPA 415.1 (T) Ox	LCS - Recovery	9.48	0.1 mg/L as C	0	0	0.948	0.85	1.15
22/03/2004	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.5288	%	9.48	9.43		0	0.15
22/03/2004	EPA 415.1 (T) Ox	Check Standard	1	0.1 mg/L as C	0	0	1	0.85	1.15
22/03/2004	EPA 415.1 (T) Ox	Low Level Standard	0.56	0.1 mg/L as C	0	0	1.12	0.5	1.5
22/03/2004	EPA 415.1 (T) Ox	Check Standard	9.31	0.1 mg/L as C	0	0	0.931	0.85	1.15
29/03/2004	EPA 415.1 (T) Ox	Check Standard	1.03	0.1 mg/L as C	0	0	1.03	0.85	1.15
29/03/2004	EPA 415.1 (T) Ox	Check Standard	9.52	0.1 mg/L as C	0	0	0.952	0.85	1.15
29/03/2004	EPA 415.1 (T) Ox	Check Standard	1.03	0.1 mg/L as C	0	0	1.03	0.85	1.15
29/03/2004	EPA 415.1 (T) Ox	Check Standard	1.02	0.1 mg/L as C	0	0	1.02	0.85	1.15
29/03/2004	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.5271	%	9.51	9.46		0	0.15

29/03/2004	EPA 415.1 (T) Ox	Check Standard	9.52	0.1 mg/L as C	0	0	0.952	0.85	1.15
29/03/2004	EPA 415.1 (T) Ox	Method Blank	0.06	0.1 mg/L as C	0	0			
29/03/2004	EPA 415.1 (T) Ox	Check Standard	1.02	0.1 mg/L as C	0	0	1.02	0.85	1.15
29/03/2004	EPA 415.1 (T) Ox	Check Standard	9.51	0.1 mg/L as C	0	0	0.951	0.85	1.15
29/03/2004	EPA 415.1 (T) Ox	RPD - Check Standard	0.9756	%	1.03	1.02		0	0.15
29/03/2004	EPA 415.1 (T) Ox	RPD - Check Standard	0.1051	%	9.51	9.52		0	0.15
29/03/2004	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.5271	%	9.51	9.46		0	0.15
29/03/2004	EPA 415.1 (T) Ox	RPD - Check Standard	0.9756	%	1.03	1.02		0	0.15
29/03/2004	EPA 415.1 (T) Ox	RPD - Check Standard	0.1051	%	9.51	9.52		0	0.15
29/03/2004	EPA 415.1 (T) Ox	Low Level Standard	0.6	0.1 mg/L as C	0	0	1.2	0.5	1.5
29/03/2004	EPA 415.1 (T) Ox	Check Standard	9.51	0.1 mg/L as C	0	0	0.951	0.85	1.15
29/03/2004	EPA 415.1 (T) Ox	LCS - Recovery	9.46	0.1 mg/L as C	0	0	0.946	0.85	1.15
29/03/2004	EPA 415.1 (T) Ox	Method Blank	0.06	0.1 mg/L as C	0	0			
29/03/2004	EPA 415.1 (T) Ox	LCS - Recovery	9.51	0.1 mg/L as C	0	0	0.951	0.85	1.15
29/03/2004	EPA 415.1 (T) Ox	Low Level Standard	0.6	0.1 mg/L as C	0	0	1.2	0.5	1.5
29/03/2004	EPA 415.1 (T) Ox	LCS - Recovery	9.46	0.1 mg/L as C	0	0	0.946	0.85	1.15
29/03/2004	EPA 415.1 (T) Ox	LCS - Recovery	9.51	0.1 mg/L as C	0	0	0.951	0.85	1.15
05/04/2004	EPA 415.1 (T) Ox	Low Level Standard	0.57	0.1 mg/L as C	0	0	1.14	0.5	1.5
05/04/2004	EPA 415.1 (T) Ox	RPD - LCS Duplicate	1.245	%	9.7	9.58		0	0.15
05/04/2004	EPA 415.1 (T) Ox	Check Standard	1.05	0.1 mg/L as C	0	0	1.05	0.85	1.15
05/04/2004	EPA 415.1 (T) Ox	RPD - Check Standard	7.339	%	1.05	1.13		0	0.15
05/04/2004	EPA 415.1 (T) Ox	Check Standard	9.74	0.1 mg/L as C	0	0	0.974	0.85	1.15
05/04/2004	EPA 415.1 (T) Ox	Check Standard	9.78	0.1 mg/L as C	0	0	0.978	0.85	1.15
05/04/2004	EPA 415.1 (T) Ox	Check Standard	1.13	0.1 mg/L as C	0	0	1.13	0.85	1.15
05/04/2004	EPA 415.1 (T) Ox	RPD - Check Standard	7.339	%	1.05	1.13		0	0.15
05/04/2004	EPA 415.1 (T) Ox	LCS - Recovery	9.58	0.1 mg/L as C	0	0	0.958	0.85	1.15
05/04/2004	EPA 415.1 (T) Ox	Check Standard	1.05	0.1 mg/L as C	0	0	1.05	0.85	1.15
05/04/2004	EPA 415.1 (T) Ox	LCS - Recovery	9.7	0.1 mg/L as C	0	0	0.97	0.85	1.15
05/04/2004	EPA 415.1 (T) Ox	LCS - Recovery	9.58	0.1 mg/L as C	0	0	0.958	0.85	1.15
05/04/2004	EPA 415.1 (T) Ox	RPD - Check Standard	0.4098	%	9.78	9.74		0	0.15
05/04/2004	EPA 415.1 (T) Ox	RPD - Check Standard	0.4098	%	9.78	9.74		0	0.15
05/04/2004	EPA 415.1 (T) Ox	RPD - LCS Duplicate	1.245	%	9.7	9.58		0	0.15
05/04/2004	EPA 415.1 (T) Ox	Low Level Standard	0.57	0.1 mg/L as C	0	0	1.14	0.5	1.5
05/04/2004	EPA 415.1 (T) Ox	Check Standard	9.78	0.1 mg/L as C	0	0	0.978	0.85	1.15
05/04/2004	EPA 415.1 (T) Ox	Method Blank	0.09	0.1 mg/L as C	0	0			
05/04/2004	EPA 415.1 (T) Ox	Check Standard	9.74	0.1 mg/L as C	0	0	0.974	0.85	1.15
05/04/2004	EPA 415.1 (T) Ox	Check Standard	1.13	0.1 mg/L as C	0	0	1.13	0.85	1.15

05/04/2004	EPA 415.1 (T) Ox	Method Blank	0.09	0.1 mg/L as C	0	0			
05/04/2004	EPA 415.1 (T) Ox	LCS - Recovery	9.7	0.1 mg/L as C	0	0	0.97	0.85	1.15
12/04/2004	EPA 415.1 (T) Ox	RPD - Check Standard	1.488	%	9.48	9.34		0	0.15
12/04/2004	EPA 415.1 (T) Ox	RPD - Check Standard	3.96	%	1.03	0.99		0	0.15
12/04/2004	EPA 415.1 (T) Ox	Method Blank	0.09	0.1 mg/L as C	0	0			
12/04/2004	EPA 415.1 (T) Ox	Check Standard	1.03	0.1 mg/L as C	0	0	1.03	0.85	1.15
12/04/2004	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0	%	9.45	9.45		0	0.15
12/04/2004	EPA 415.1 (T) Ox	Low Level Standard	0.57	0.1 mg/L as C	0	0	1.14	0.5	1.5
12/04/2004	EPA 415.1 (T) Ox	Check Standard	0.99	0.1 mg/L as C	0	0	0.99	0.85	1.15
12/04/2004	EPA 415.1 (T) Ox	LCS - Recovery	9.45	0.1 mg/L as C	0	0	0.945	0.85	1.15
12/04/2004	EPA 415.1 (T) Ox	LCS - Recovery	9.45	0.1 mg/L as C	0	0	0.945	0.85	1.15
12/04/2004	EPA 415.1 (T) Ox	Check Standard	9.48	0.1 mg/L as C	0	0	0.948	0.85	1.15
12/04/2004	EPA 415.1 (T) Ox	Check Standard	9.34	0.1 mg/L as C	0	0	0.934	0.85	1.15
19/04/2004	EPA 415.1 (T) Ox	LCS - Recovery	9.69	0.1 mg/L as C	0	0	0.969	0.85	1.15
19/04/2004	EPA 415.1 (T) Ox	Method Blank	0.09	0.1 mg/L as C	0	0			
19/04/2004	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.3101	%	9.69	9.66		0	0.15
19/04/2004	EPA 415.1 (T) Ox	RPD - Check Standard	0.4124	%	9.72	9.68		0	0.15
19/04/2004	EPA 415.1 (T) Ox	Check Standard	1.04	0.1 mg/L as C	0	0	1.04	0.85	1.15
19/04/2004	EPA 415.1 (T) Ox	LCS - Recovery	9.66	0.1 mg/L as C	0	0	0.966	0.85	1.15
19/04/2004	EPA 415.1 (T) Ox	Low Level Standard	0.55	0.1 mg/L as C	0	0	1.1	0.5	1.5
19/04/2004	EPA 415.1 (T) Ox	Check Standard	9.68	0.1 mg/L as C	0	0	0.968	0.85	1.15
19/04/2004	EPA 415.1 (T) Ox	Check Standard	9.72	0.1 mg/L as C	0	0	0.972	0.85	1.15
19/04/2004	EPA 415.1 (T) Ox	Check Standard	1.07	0.1 mg/L as C	0	0	1.07	0.85	1.15
19/04/2004	EPA 415.1 (T) Ox	RPD - Check Standard	2.844	%	1.04	1.07		0	0.15
26/04/2004	EPA 415.1 (T) Ox	Check Standard	1.08	0.1 mg/L as C	0	0	1.08	0.85	1.15
26/04/2004	EPA 415.1 (T) Ox	LCS - Recovery	9.72	0.1 mg/L as C	0	0	0.972	0.85	1.15
26/04/2004	EPA 415.1 (T) Ox	LCS - Recovery	9.68	0.1 mg/L as C	0	0	0.968	0.85	1.15
26/04/2004	EPA 415.1 (T) Ox	Low Level Standard	0.58	0.1 mg/L as C	0	0	1.16	0.5	1.5
26/04/2004	EPA 415.1 (T) Ox	Check Standard	9.76	0.1 mg/L as C	0	0	0.976	0.85	1.15
26/04/2004	EPA 415.1 (T) Ox	Check Standard	9.76	0.1 mg/L as C	0	0	0.976	0.85	1.15
26/04/2004	EPA 415.1 (T) Ox	Check Standard	1.08	0.1 mg/L as C	0	0	1.08	0.85	1.15
26/04/2004	EPA 415.1 (T) Ox	Check Standard	1.06	0.1 mg/L as C	0	0	1.06	0.85	1.15
26/04/2004	EPA 415.1 (T) Ox	Low Level Standard	0.58	0.1 mg/L as C	0	0	1.16	0.5	1.5
26/04/2004	EPA 415.1 (T) Ox	Check Standard	9.76	0.1 mg/L as C	0	0	0.976	0.85	1.15
26/04/2004	EPA 415.1 (T) Ox	Check Standard	1.06	0.1 mg/L as C	0	0	1.06	0.85	1.15
26/04/2004	EPA 415.1 (T) Ox	Method Blank	0.09	0.1 mg/L as C	0	0			
26/04/2004	EPA 415.1 (T) Ox	Method Blank	0.09	0.1 mg/L as C	0	0			

26/04/2004	EPA 415.1 (T) Ox	Check Standard	9.76	0.1 mg/L as C	0	0	0.976	0.85	1.15
26/04/2004	EPA 415.1 (T) Ox	LCS - Recovery	9.68	0.1 mg/L as C	0	0	0.968	0.85	1.15
26/04/2004	EPA 415.1 (T) Ox	LCS - Recovery	9.72	0.1 mg/L as C	0	0	0.972	0.85	1.15
03/05/2004	EPA 415.1 (T) Ox	LCS - Recovery	9.73	0.1 mg/L as C	0	0	0.973	0.85	1.15
03/05/2004	EPA 415.1 (T) Ox	Check Standard	9.72	0.1 mg/L as C	0	0	0.972	0.85	1.15
03/05/2004	EPA 415.1 (T) Ox	Check Standard	9.64	0.1 mg/L as C	0	0	0.964	0.85	1.15
03/05/2004	EPA 415.1 (T) Ox	Check Standard	9.64	0.1 mg/L as C	0	0	0.964	0.85	1.15
03/05/2004	EPA 415.1 (T) Ox	LCS - Recovery	9.73	0.1 mg/L as C	0	0	0.973	0.85	1.15
03/05/2004	EPA 415.1 (T) Ox	RPD - Check Standard	0.8264	%	9.64	9.72	0	0	0.15
03/05/2004	EPA 415.1 (T) Ox	LCS - Recovery	9.73	0.1 mg/L as C	0	0	0.973	0.85	1.15
03/05/2004	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.722	%	9.73	9.66	0	0	0.15
03/05/2004	EPA 415.1 (T) Ox	RPD - Check Standard	0.8264	%	9.64	9.72	0	0	0.15
03/05/2004	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.722	%	9.73	9.66	0	0	0.15
03/05/2004	EPA 415.1 (T) Ox	Check Standard	9.72	0.1 mg/L as C	0	0	0.972	0.85	1.15
03/05/2004	EPA 415.1 (T) Ox	LCS - Recovery	9.66	0.1 mg/L as C	0	0	0.966	0.85	1.15
03/05/2004	EPA 415.1 (T) Ox	Low Level Standard	0.63	0.1 mg/L as C	0	0	1.26	0.5	1.5
03/05/2004	EPA 415.1 (T) Ox	LCS - Recovery	9.66	0.1 mg/L as C	0	0	0.966	0.85	1.15
03/05/2004	EPA 415.1 (T) Ox	Low Level Standard	0.63	0.1 mg/L as C	0	0	1.26	0.5	1.5
03/05/2004	EPA 415.1 (T) Ox	Low Level Standard	0.63	0.1 mg/L as C	0	0	1.26	0.5	1.5
03/05/2004	EPA 415.1 (T) Ox	LCS - Recovery	9.66	0.1 mg/L as C	0	0	0.966	0.85	1.15
03/05/2004	EPA 415.1 (T) Ox	RPD - Check Standard	0.8264	%	9.64	9.72	0	0	0.15
03/05/2004	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.722	%	9.73	9.66	0	0	0.15
03/05/2004	EPA 415.1 (T) Ox	Check Standard	9.72	0.1 mg/L as C	0	0	0.972	0.85	1.15
03/05/2004	EPA 415.1 (T) Ox	Method Blank	0.1	0.1 mg/L as C	0	0			
03/05/2004	EPA 415.1 (T) Ox	Check Standard	9.64	0.1 mg/L as C	0	0	0.964	0.85	1.15
03/05/2004	EPA 415.1 (T) Ox	Method Blank	0.1	0.1 mg/L as C	0	0			
03/05/2004	EPA 415.1 (T) Ox	Method Blank	0.1	0.1 mg/L as C	0	0			
10/05/2004	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.3107	%	9.67	9.64	0	0	0.15
10/05/2004	EPA 415.1 (T) Ox	Low Level Standard	0.6	0.1 mg/L as C	0	0	1.2	0.5	1.5
10/05/2004	EPA 415.1 (T) Ox	RPD - Check Standard	2.363	%	9.85	9.62	0	0	0.15
10/05/2004	EPA 415.1 (T) Ox	LCS - Recovery	9.67	0.1 mg/L as C	0	0	0.967	0.85	1.15
10/05/2004	EPA 415.1 (T) Ox	Check Standard	9.62	0.1 mg/L as C	0	0	0.962	0.85	1.15
10/05/2004	EPA 415.1 (T) Ox	Check Standard	1.09	0.1 mg/L as C	0	0	1.09	0.85	1.15
10/05/2004	EPA 415.1 (T) Ox	Check Standard	9.85	0.1 mg/L as C	0	0	0.985	0.85	1.15
10/05/2004	EPA 415.1 (T) Ox	Method Blank	0.08	0.1 mg/L as C	0	0			
10/05/2004	EPA 415.1 (T) Ox	LCS - Recovery	9.64	0.1 mg/L as C	0	0	0.964	0.85	1.15
17/05/2004	EPA 415.1 (T) Ox	LCS - Recovery	9.62	0.1 mg/L as C	0	0	0.962	0.85	1.15

17/05/2004	EPA 415.1 (T) Ox	Check Standard	1.06	0.1 mg/L as C	0	0	1.06	0.85	1.15
17/05/2004	EPA 415.1 (T) Ox	Low Level Standard	0.65	0.1 mg/L as C	0	0	1.3	0.5	1.5
17/05/2004	EPA 415.1 (T) Ox	LCS - Recovery	9.68	0.1 mg/L as C	0	0	0.968	0.85	1.15
17/05/2004	EPA 415.1 (T) Ox	LCS - Recovery	9.62	0.1 mg/L as C	0	0	0.962	0.85	1.15
17/05/2004	EPA 415.1 (T) Ox	LCS - Recovery	9.68	0.1 mg/L as C	0	0	0.968	0.85	1.15
17/05/2004	EPA 415.1 (T) Ox	Check Standard	9.67	0.1 mg/L as C	0	0	0.967	0.85	1.15
17/05/2004	EPA 415.1 (T) Ox	Check Standard	9.74	0.1 mg/L as C	0	0	0.974	0.85	1.15
17/05/2004	EPA 415.1 (T) Ox	RPD - Check Standard	0.7213	%	9.74	9.67		0	0.15
17/05/2004	EPA 415.1 (T) Ox	Low Level Standard	0.65	0.1 mg/L as C	0	0	1.3	0.5	1.5
17/05/2004	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.6218	%	9.62	9.68		0	0.15
17/05/2004	EPA 415.1 (T) Ox	Check Standard	1.06	0.1 mg/L as C	0	0	1.06	0.85	1.15
17/05/2004	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.6218	%	9.62	9.68		0	0.15
17/05/2004	EPA 415.1 (T) Ox	RPD - Check Standard	0.7213	%	9.74	9.67		0	0.15
17/05/2004	EPA 415.1 (T) Ox	Check Standard	9.67	0.1 mg/L as C	0	0	0.967	0.85	1.15
17/05/2004	EPA 415.1 (T) Ox	Check Standard	9.74	0.1 mg/L as C	0	0	0.974	0.85	1.15
17/05/2004	EPA 415.1 (T) Ox	Method Blank	0.09	0.1 mg/L as C	0	0			
17/05/2004	EPA 415.1 (T) Ox	Method Blank	0.09	0.1 mg/L as C	0	0			
24/05/2004	EPA 415.1 (T) Ox	LCS - Recovery	9.8	0.1 mg/L as C	0	0	0.98	0.85	1.15
24/05/2004	EPA 415.1 (T) Ox	LCS - Recovery	9.84	0.1 mg/L as C	0	0	0.984	0.85	1.15
24/05/2004	EPA 415.1 (T) Ox	RPD - Check Standard	3.299	%	9.86	9.54		0	0.15
24/05/2004	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.4073	%	9.8	9.84		0	0.15
24/05/2004	EPA 415.1 (T) Ox	Method Blank	0.09	0.1 mg/L as C	0	0			
24/05/2004	EPA 415.1 (T) Ox	Low Level Standard	0.63	0.1 mg/L as C	0	0	1.26	0.5	1.5
24/05/2004	EPA 415.1 (T) Ox	Check Standard	1.13	0.1 mg/L as C	0	0	1.13	0.85	1.15
24/05/2004	EPA 415.1 (T) Ox	Check Standard	9.86	0.1 mg/L as C	0	0	0.986	0.85	1.15
24/05/2004	EPA 415.1 (T) Ox	Check Standard	9.54	0.1 mg/L as C	0	0	0.954	0.85	1.15
01/06/2004	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.9014	%	9.94	10.03		0	0.15
01/06/2004	EPA 415.1 (T) Ox	Low Level Standard	0.71	0.1 mg/L as C	0	0	1.42	0.5	1.5
01/06/2004	EPA 415.1 (T) Ox	LCS - Recovery	9.94	0.1 mg/L as C	0	0	0.994	0.85	1.15
01/06/2004	EPA 415.1 (T) Ox	LCS - Recovery	10.03	0.1 mg/L as C	0	0	1.003	0.85	1.15
01/06/2004	EPA 415.1 (T) Ox	LCS - Recovery	9.94	0.1 mg/L as C	0	0	0.994	0.85	1.15
01/06/2004	EPA 415.1 (T) Ox	Check Standard	10.1	0.1 mg/L as C	0	0	1.01	0.85	1.15
01/06/2004	EPA 415.1 (T) Ox	Check Standard	1.15	0.1 mg/L as C	0	0	1.15	0.85	1.15
01/06/2004	EPA 415.1 (T) Ox	LCS - Recovery	10.03	0.1 mg/L as C	0	0	1.003	0.85	1.15
01/06/2004	EPA 415.1 (T) Ox	RPD - Check Standard	4.04	%	10.1	9.7		0	0.15
01/06/2004	EPA 415.1 (T) Ox	Low Level Standard	0.71	0.1 mg/L as C	0	0	1.42	0.5	1.5
01/06/2004	EPA 415.1 (T) Ox	RPD - Check Standard	4.04	%	10.1	9.7		0	0.15

01/06/2004	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.9014	%	9.94	10.03	0	0.15
01/06/2004	EPA 415.1 (T) Ox	Check Standard	10.1	0.1 mg/L as C	0	0	1.01	0.85
01/06/2004	EPA 415.1 (T) Ox	Check Standard	1.15	0.1 mg/L as C	0	0	1.15	0.85
01/06/2004	EPA 415.1 (T) Ox	Method Blank	0.1	0.1 mg/L as C	0	0		
01/06/2004	EPA 415.1 (T) Ox	Method Blank	0.1	0.1 mg/L as C	0	0		
01/06/2004	EPA 415.1 (T) Ox	Check Standard	9.7	0.1 mg/L as C	0	0	0.97	0.85
01/06/2004	EPA 415.1 (T) Ox	Check Standard	9.7	0.1 mg/L as C	0	0	0.97	0.85
07/06/2004	EPA 415.1 (T) Ox	Check Standard	9.85	0.1 mg/L as C	0	0	0.985	0.85
07/06/2004	EPA 415.1 (T) Ox	Check Standard	9.95	0.1 mg/L as C	0	0	0.995	0.85
07/06/2004	EPA 415.1 (T) Ox	Check Standard	1.04	0.1 mg/L as C	0	0	1.04	0.85
07/06/2004	EPA 415.1 (T) Ox	Check Standard	9.85	0.1 mg/L as C	0	0	0.985	0.85
07/06/2004	EPA 415.1 (T) Ox	LCS - Recovery	9.85	0.1 mg/L as C	0	0	0.985	0.85
07/06/2004	EPA 415.1 (T) Ox	LCS - Recovery	9.85	0.1 mg/L as C	0	0	0.985	0.85
07/06/2004	EPA 415.1 (T) Ox	LCS - Recovery	9.91	0.1 mg/L as C	0	0	0.991	0.85
07/06/2004	EPA 415.1 (T) Ox	Low Level Standard	0.62	0.1 mg/L as C	0	0	1.24	0.5
07/06/2004	EPA 415.1 (T) Ox	RPD - Check Standard	1.01	%	9.95	9.85	0	0.15
07/06/2004	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.6073	%	9.85	9.91	0	0.15
07/06/2004	EPA 415.1 (T) Ox	RPD - Check Standard	1.01	%	9.95	9.85	0	0.15
07/06/2004	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.6073	%	9.85	9.91	0	0.15
07/06/2004	EPA 415.1 (T) Ox	Method Blank	0.07	0.1 mg/L as C	0	0		
07/06/2004	EPA 415.1 (T) Ox	Method Blank	0.07	0.1 mg/L as C	0	0		
07/06/2004	EPA 415.1 (T) Ox	LCS - Recovery	9.91	0.1 mg/L as C	0	0	0.991	0.85
07/06/2004	EPA 415.1 (T) Ox	Check Standard	1.04	0.1 mg/L as C	0	0	1.04	0.85
07/06/2004	EPA 415.1 (T) Ox	Low Level Standard	0.62	0.1 mg/L as C	0	0	1.24	0.5
07/06/2004	EPA 415.1 (T) Ox	Check Standard	9.95	0.1 mg/L as C	0	0	0.995	0.85
14/06/2004	EPA 415.1 (T) Ox	LCS - Recovery	10.08	0.1 mg/L as C	0	0	1.008	0.85
14/06/2004	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0	%	10.08	10.08	0	0.15
14/06/2004	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0	%	10.08	10.08	0	0.15
14/06/2004	EPA 415.1 (T) Ox	Check Standard	1.09	0.1 mg/L as C	0	0	1.09	0.85
14/06/2004	EPA 415.1 (T) Ox	Check Standard	10.24	0.1 mg/L as C	0	0	1.024	0.85
14/06/2004	EPA 415.1 (T) Ox	Check Standard	10	0.1 mg/L as C	0	0	1	0.85
14/06/2004	EPA 415.1 (T) Ox	Check Standard	1.09	0.1 mg/L as C	0	0	1.09	0.85
14/06/2004	EPA 415.1 (T) Ox	LCS - Recovery	10.08	0.1 mg/L as C	0	0	1.008	0.85
14/06/2004	EPA 415.1 (T) Ox	RPD - Check Standard	2.372	%	10.24	10	0	0.15
14/06/2004	EPA 415.1 (T) Ox	Low Level Standard	0.61	0.1 mg/L as C	0	0	1.22	0.5
14/06/2004	EPA 415.1 (T) Ox	LCS - Recovery	10.08	0.1 mg/L as C	0	0	1.008	0.85
14/06/2004	EPA 415.1 (T) Ox	LCS - Recovery	10.08	0.1 mg/L as C	0	0	1.008	0.85

14/06/2004	EPA 415.1 (T) Ox	Check Standard	10.24	0.1 mg/L as C	0	0	1.024	0.85	1.15
14/06/2004	EPA 415.1 (T) Ox	Check Standard	10	0.1 mg/L as C	0	0	1	0.85	1.15
14/06/2004	EPA 415.1 (T) Ox	Method Blank	0.1	0.1 mg/L as C	0	0			
14/06/2004	EPA 415.1 (T) Ox	Method Blank	0.1	0.1 mg/L as C	0	0			
14/06/2004	EPA 415.1 (T) Ox	Low Level Standard	0.61	0.1 mg/L as C	0	0	1.22	0.5	1.5
14/06/2004	EPA 415.1 (T) Ox	RPD - Check Standard	2.372	%	10.24	10		0	0.15
21/06/2004	EPA 415.1 (T) Ox	Method Blank	0.07	0.1 mg/L as C	0	0			
21/06/2004	EPA 415.1 (T) Ox	LCS - Recovery	10	0.1 mg/L as C	0	0	1	0.85	1.15
21/06/2004	EPA 415.1 (T) Ox	LCS - Recovery	10	0.1 mg/L as C	0	0	1	0.85	1.15
21/06/2004	EPA 415.1 (T) Ox	Low Level Standard	0.62	0.1 mg/L as C	0	0	1.24	0.5	1.5
21/06/2004	EPA 415.1 (T) Ox	RPD - Check Standard	4.784	%	10.06	9.59		0	0.15
21/06/2004	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0	%	10	10		0	0.15
21/06/2004	EPA 415.1 (T) Ox	Check Standard	1.06	0.1 mg/L as C	0	0	1.06	0.85	1.15
21/06/2004	EPA 415.1 (T) Ox	Check Standard	9.59	0.1 mg/L as C	0	0	0.959	0.85	1.15
21/06/2004	EPA 415.1 (T) Ox	Check Standard	10.06	0.1 mg/L as C	0	0	1.006	0.85	1.15
28/06/2004	EPA 415.1 (T) Ox	Method Blank	0.07	0.1 mg/L as C	0	0			
28/06/2004	EPA 415.1 (T) Ox	RPD - LCS Duplicate	1.436	%	10.52	10.37		0	0.15
28/06/2004	EPA 415.1 (T) Ox	Check Standard	1.08	0.1 mg/L as C	0	0	1.08	0.85	1.15
28/06/2004	EPA 415.1 (T) Ox	RPD - Check Standard	0.6747	%	10.34	10.41		0	0.15
28/06/2004	EPA 415.1 (T) Ox	LCS - Recovery	10.37	0.1 mg/L as C	0	0	1.037	0.85	1.15
28/06/2004	EPA 415.1 (T) Ox	Method Blank	0.07	0.1 mg/L as C	0	0			
28/06/2004	EPA 415.1 (T) Ox	RPD - LCS Duplicate	1.436	%	10.52	10.37		0	0.15
28/06/2004	EPA 415.1 (T) Ox	RPD - Check Standard	0.6747	%	10.34	10.41		0	0.15
28/06/2004	EPA 415.1 (T) Ox	Check Standard	10.41	0.1 mg/L as C	0	0	1.041	0.85	1.15
28/06/2004	EPA 415.1 (T) Ox	Check Standard	10.34	0.1 mg/L as C	0	0	1.034	0.85	1.15
28/06/2004	EPA 415.1 (T) Ox	Low Level Standard	0.6	0.1 mg/L as C	0	0	1.2	0.5	1.5
28/06/2004	EPA 415.1 (T) Ox	Check Standard	10.41	0.1 mg/L as C	0	0	1.041	0.85	1.15
28/06/2004	EPA 415.1 (T) Ox	Check Standard	1.08	0.1 mg/L as C	0	0	1.08	0.85	1.15
28/06/2004	EPA 415.1 (T) Ox	Check Standard	10.34	0.1 mg/L as C	0	0	1.034	0.85	1.15
28/06/2004	EPA 415.1 (T) Ox	LCS - Recovery	10.37	0.1 mg/L as C	0	0	1.037	0.85	1.15
28/06/2004	EPA 415.1 (T) Ox	Low Level Standard	0.6	0.1 mg/L as C	0	0	1.2	0.5	1.5
06/07/2004	EPA 415.1 (T) Ox	RPD - LCS Duplicate	1.118	%	9.89	9.78		0	0.15
06/07/2004	EPA 415.1 (T) Ox	Check Standard	9.88	0.1 mg/L as C	0	0	0.988	0.85	1.15
06/07/2004	EPA 415.1 (T) Ox	LCS - Recovery	9.78	0.1 mg/L as C	0	0	0.978	0.85	1.15
06/07/2004	EPA 415.1 (T) Ox	LCS - Recovery	9.89	0.1 mg/L as C	0	0	0.989	0.85	1.15
06/07/2004	EPA 415.1 (T) Ox	RPD - Check Standard	5.32	%	9.88	10.42		0	0.15
06/07/2004	EPA 415.1 (T) Ox	Low Level Standard	0.58	0.1 mg/L as C	0	0	1.16	0.5	1.5

06/07/2004	EPA 415.1 (T) Ox	Check Standard	10.42	0.1 mg/L as C	0	0	1.042	0.85	1.15
06/07/2004	EPA 415.1 (T) Ox	Check Standard	1.03	0.1 mg/L as C	0	0	1.03	0.85	1.15
06/07/2004	EPA 415.1 (T) Ox	Method Blank	0.08	0.1 mg/L as C	0	0			
12/07/2004	EPA 415.1 (T) Ox	Check Standard	9.83	0.1 mg/L as C	0	0	0.983	0.85	1.15
12/07/2004	EPA 415.1 (T) Ox	RPD - Check Standard	3.006	%	10.13	9.83		0	0.15
12/07/2004	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.5074	%	9.83	9.88		0	0.15
12/07/2004	EPA 415.1 (T) Ox	Method Blank	0.06	0.1 mg/L as C	0	0			
12/07/2004	EPA 415.1 (T) Ox	Check Standard	10.13	0.1 mg/L as C	0	0	1.013	0.85	1.15
12/07/2004	EPA 415.1 (T) Ox	Check Standard	1.07	0.1 mg/L as C	0	0	1.07	0.85	1.15
12/07/2004	EPA 415.1 (T) Ox	LCS - Recovery	9.83	0.1 mg/L as C	0	0	0.983	0.85	1.15
12/07/2004	EPA 415.1 (T) Ox	LCS - Recovery	9.88	0.1 mg/L as C	0	0	0.988	0.85	1.15
12/07/2004	EPA 415.1 (T) Ox	Low Level Standard	0.61	0.1 mg/L as C	0	0	1.22	0.5	1.5
20/07/2004	EPA 415.1 (T) Ox	Check Standard	1.14	0.1 mg/L as C	0	0	1.14	0.85	1.15
20/07/2004	EPA 415.1 (T) Ox	Low Level Standard	0.6	0.1 mg/L as C	0	0	1.2	0.5	1.5
20/07/2004	EPA 415.1 (T) Ox	LCS - Recovery	9.91	0.1 mg/L as C	0	0	0.991	0.85	1.15
20/07/2004	EPA 415.1 (T) Ox	LCS - Recovery	9.92	0.1 mg/L as C	0	0	0.992	0.85	1.15
20/07/2004	EPA 415.1 (T) Ox	RPD - Check Standard	0.4057	%	9.88	9.84		0	0.15
20/07/2004	EPA 415.1 (T) Ox	Check Standard	9.84	0.1 mg/L as C	0	0	0.984	0.85	1.15
20/07/2004	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.1009	%	9.92	9.91		0	0.15
20/07/2004	EPA 415.1 (T) Ox	Check Standard	9.88	0.1 mg/L as C	0	0	0.988	0.85	1.15
20/07/2004	EPA 415.1 (T) Ox	Method Blank	0.1	0.1 mg/L as C	0	0			
26/07/2004	EPA 415.1 (T) Ox	RPD - Check Standard	0.6006	%	10.02	9.96		0	0.15
26/07/2004	EPA 415.1 (T) Ox	Check Standard	9.96	0.1 mg/L as C	0	0	0.996	0.85	1.15
26/07/2004	EPA 415.1 (T) Ox	RPD - LCS Duplicate	1.413	%	9.98	9.84		0	0.15
26/07/2004	EPA 415.1 (T) Ox	LCS - Recovery	9.98	0.1 mg/L as C	0	0	0.998	0.85	1.15
26/07/2004	EPA 415.1 (T) Ox	Check Standard	10.02	0.1 mg/L as C	0	0	1.002	0.85	1.15
26/07/2004	EPA 415.1 (T) Ox	LCS - Recovery	9.84	0.1 mg/L as C	0	0	0.984	0.85	1.15
26/07/2004	EPA 415.1 (T) Ox	Method Blank	0.05	0.1 mg/L as C	0	0			
26/07/2004	EPA 415.1 (T) Ox	Low Level Standard	0.58	0.1 mg/L as C	0	0	1.16	0.5	1.5
03/08/2004	EPA 415.1 (T) Ox	Check Standard	9.96	0.1 mg/L as C	0	0	0.996	0.85	1.15
03/08/2004	EPA 415.1 (T) Ox	Check Standard	9.86	0.1 mg/L as C	0	0	0.986	0.85	1.15
03/08/2004	EPA 415.1 (T) Ox	LCS - Recovery	9.72	0.1 mg/L as C	0	0	0.972	0.85	1.15
03/08/2004	EPA 415.1 (T) Ox	RPD - Check Standard	1.009	%	9.86	9.96		0	0.15
03/08/2004	EPA 415.1 (T) Ox	Low Level Standard	0.62	0.1 mg/L as C	0	0	1.24	0.5	1.5
03/08/2004	EPA 415.1 (T) Ox	LCS - Recovery	9.84	0.1 mg/L as C	0	0	0.984	0.85	1.15
03/08/2004	EPA 415.1 (T) Ox	Method Blank	0	0.1 mg/L as C	0	0			
03/08/2004	EPA 415.1 (T) Ox	RPD - LCS Duplicate	1.227	%	9.72	9.84		0	0.15

10/08/2004	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.9756	%	10.3	10.2	0	0.15
10/08/2004	EPA 415.1 (T) Ox	RPD - Check Standard	1.961	%	10.1	10.3	0	0.15
10/08/2004	EPA 415.1 (T) Ox	Method Blank	0	0.1 mg/L as C	0	0		
10/08/2004	EPA 415.1 (T) Ox	LCS - Recovery	10.2	0.1 mg/L as C	0	0	1.02	0.85
10/08/2004	EPA 415.1 (T) Ox	Check Standard	10.1	0.1 mg/L as C	0	0	1.01	0.85
10/08/2004	EPA 415.1 (T) Ox	Check Standard	10.3	0.1 mg/L as C	0	0	1.03	0.85
10/08/2004	EPA 415.1 (T) Ox	LCS - Recovery	10.3	0.1 mg/L as C	0	0	1.03	0.85
17/08/2004	EPA 415.1 (T) Ox	LCS - Recovery	9.99	0.1 mg/L as C	0	0	0.999	0.85
17/08/2004	EPA 415.1 (T) Ox	Check Standard	1.12	0.1 mg/L as C	0	0	1.12	0.85
17/08/2004	EPA 415.1 (T) Ox	Method Blank	0.1	0.1 mg/L as C	0	0		
17/08/2004	EPA 415.1 (T) Ox	RPD - Check Standard	5.505	%	1.12	1.06	0	0.15
17/08/2004	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.996	%	9.99	10.09	0	0.15
17/08/2004	EPA 415.1 (T) Ox	RPD - Check Standard	2.457	%	10.05	10.3	0	0.15
17/08/2004	EPA 415.1 (T) Ox	Check Standard	1.06	0.1 mg/L as C	0	0	1.06	0.85
17/08/2004	EPA 415.1 (T) Ox	LCS - Recovery	10.09	0.1 mg/L as C	0	0	1.009	0.85
17/08/2004	EPA 415.1 (T) Ox	Check Standard	10.3	0.1 mg/L as C	0	0	1.03	0.85
17/08/2004	EPA 415.1 (T) Ox	Check Standard	10.05	0.1 mg/L as C	0	0	1.005	0.85
17/08/2004	EPA 415.1 (T) Ox	Low Level Standard	0.63	0.1 mg/L as C	0	0	1.26	0.5
24/08/2004	EPA 415.1 (T) Ox	RPD - Check Standard	0.8016	%	10.02	9.94	0	0.15
24/08/2004	EPA 415.1 (T) Ox	Check Standard	9.94	0.1 mg/L as C	0	0	0.994	0.85
24/08/2004	EPA 415.1 (T) Ox	RPD - Check Standard	0.905	%	1.1	1.11	0	0.15
24/08/2004	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.3023	%	9.94	9.91	0	0.15
24/08/2004	EPA 415.1 (T) Ox	Check Standard	1.1	0.1 mg/L as C	0	0	1.1	0.85
24/08/2004	EPA 415.1 (T) Ox	LCS - Recovery	9.94	0.1 mg/L as C	0	0	0.994	0.85
24/08/2004	EPA 415.1 (T) Ox	Check Standard	1.11	0.1 mg/L as C	0	0	1.11	0.85
24/08/2004	EPA 415.1 (T) Ox	Check Standard	10.02	0.1 mg/L as C	0	0	1.002	0.85
24/08/2004	EPA 415.1 (T) Ox	Low Level Standard	0.65	0.1 mg/L as C	0	0	1.3	0.5
24/08/2004	EPA 415.1 (T) Ox	LCS - Recovery	9.91	0.1 mg/L as C	0	0	0.991	0.85
24/08/2004	EPA 415.1 (T) Ox	Method Blank	0.08	0.1 mg/L as C	0	0		
31/08/2004	EPA 415.1 (T) Ox	Check Standard	1.06	0.1 mg/L as C	0	0	1.06	0.85
31/08/2004	EPA 415.1 (T) Ox	Method Blank	0.08	0.1 mg/L as C	0	0		
31/08/2004	EPA 415.1 (T) Ox	Check Standard	9.94	0.1 mg/L as C	0	0	0.994	0.85
31/08/2004	EPA 415.1 (T) Ox	LCS - Recovery	10.05	0.1 mg/L as C	0	0	1.005	0.85
31/08/2004	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.299	%	10.05	10.02	0	0.15
31/08/2004	EPA 415.1 (T) Ox	LCS - Recovery	10.02	0.1 mg/L as C	0	0	1.002	0.85
31/08/2004	EPA 415.1 (T) Ox	Low Level Standard	0.62	0.1 mg/L as C	0	0	1.24	0.5
31/08/2004	EPA 415.1 (T) Ox	RPD - Check Standard	9.761	%	9.94	10.96	0	0.15

13/09/2004	EPA 415.1 (T) Ox	Low Level Standard	0.56	0.1 mg/L as C	0	0	1.12	0.5	1.5
13/09/2004	EPA 415.1 (T) Ox	RPD - Check Standard	1.322	%	9.77	9.9		0	0.15
13/09/2004	EPA 415.1 (T) Ox	RPD - LCS Duplicate	1.13	%	9.79	9.68		0	0.15
13/09/2004	EPA 415.1 (T) Ox	Check Standard	9.9	0.1 mg/L as C	0	0	0.99	0.85	1.15
13/09/2004	EPA 415.1 (T) Ox	LCS - Recovery	9.68	0.1 mg/L as C	0	0	0.968	0.85	1.15
13/09/2004	EPA 415.1 (T) Ox	RPD - Check Standard	3.922	%	1	1.04		0	0.15
13/09/2004	EPA 415.1 (T) Ox	Check Standard	1	0.1 mg/L as C	0	0	1	0.85	1.15
13/09/2004	EPA 415.1 (T) Ox	Check Standard	1.04	0.1 mg/L as C	0	0	1.04	0.85	1.15
13/09/2004	EPA 415.1 (T) Ox	Check Standard	9.77	0.1 mg/L as C	0	0	0.977	0.85	1.15
13/09/2004	EPA 415.1 (T) Ox	Method Blank	0.07	0.1 mg/L as C	0	0			
13/09/2004	EPA 415.1 (T) Ox	LCS - Recovery	9.79	0.1 mg/L as C	0	0	0.979	0.85	1.15
20/09/2004	EPA 415.1 (T) Ox	LCS - Recovery	9.57	0.1 mg/L as C	0	0	0.957	0.85	1.15
20/09/2004	EPA 415.1 (T) Ox	Check Standard	1	0.1 mg/L as C	0	0	1	0.85	1.15
20/09/2004	EPA 415.1 (T) Ox	Low Level Standard	0.56	0.1 mg/L as C	0	0	1.12	0.5	1.5
20/09/2004	EPA 415.1 (T) Ox	LCS - Recovery	9.7	0.1 mg/L as C	0	0	0.97	0.85	1.15
20/09/2004	EPA 415.1 (T) Ox	Check Standard	9.96	0.1 mg/L as C	0	0	0.996	0.85	1.15
20/09/2004	EPA 415.1 (T) Ox	Check Standard	9.66	0.1 mg/L as C	0	0	0.966	0.85	1.15
20/09/2004	EPA 415.1 (T) Ox	RPD - Check Standard	3.058	%	9.66	9.96		0	0.15
20/09/2004	EPA 415.1 (T) Ox	RPD - LCS Duplicate	1.349	%	9.57	9.7		0	0.15
20/09/2004	EPA 415.1 (T) Ox	Method Blank	0.09	0.1 mg/L as C	0	0			
27/09/2004	EPA 415.1 (T) Ox	Method Blank	0.09	0.1 mg/L as C	0	0			
27/09/2004	EPA 415.1 (T) Ox	Low Level Standard	0.55	0.1 mg/L as C	0	0	1.1	0.5	1.5
27/09/2004	EPA 415.1 (T) Ox	LCS - Recovery	9.92	0.1 mg/L as C	0	0	0.992	0.85	1.15
27/09/2004	EPA 415.1 (T) Ox	RPD - LCS Duplicate	1.421	%	9.78	9.92		0	0.15
27/09/2004	EPA 415.1 (T) Ox	RPD - Check Standard	3.236	%	9.73	10.05		0	0.15
27/09/2004	EPA 415.1 (T) Ox	Check Standard	10.05	0.1 mg/L as C	0	0	1.005	0.85	1.15
27/09/2004	EPA 415.1 (T) Ox	Check Standard	9.73	0.1 mg/L as C	0	0	0.973	0.85	1.15
27/09/2004	EPA 415.1 (T) Ox	Check Standard	0.98	0.1 mg/L as C	0	0	0.98	0.85	1.15
27/09/2004	EPA 415.1 (T) Ox	LCS - Recovery	9.78	0.1 mg/L as C	0	0	0.978	0.85	1.15
04/10/2004	EPA 415.1 (T) Ox	Check Standard	9.93	0.1 mg/L as C	0	0	0.993	0.85	1.15
04/10/2004	EPA 415.1 (T) Ox	RPD - Check Standard	2.757	%	9.66	9.93		0	0.15
04/10/2004	EPA 415.1 (T) Ox	Check Standard	9.66	0.1 mg/L as C	0	0	0.966	0.85	1.15
04/10/2004	EPA 415.1 (T) Ox	Method Blank	0.08	0.1 mg/L as C	0	0			
04/10/2004	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.9322	%	9.7	9.61		0	0.15
04/10/2004	EPA 415.1 (T) Ox	Low Level Standard	0.54	0.1 mg/L as C	0	0	1.08	0.5	1.5
04/10/2004	EPA 415.1 (T) Ox	LCS - Recovery	9.61	0.1 mg/L as C	0	0	0.961	0.85	1.15
04/10/2004	EPA 415.1 (T) Ox	LCS - Recovery	9.7	0.1 mg/L as C	0	0	0.97	0.85	1.15

04/10/2004	EPA 415.1 (T) Ox	Check Standard	0.97	0.1 mg/L as C	0	0	0.97	0.85	1.15
18/10/2004	EPA 415.1 (T) Ox	Check Standard	9.6	0.1 mg/L as C	0	0	0.96	0.85	1.15
18/10/2004	EPA 415.1 (T) Ox	Method Blank	0.03	0.1 mg/L as C	0	0			
18/10/2004	EPA 415.1 (T) Ox	RPD - LCS Duplicate	5.037	%	9.87	10.38		0	0.15
18/10/2004	EPA 415.1 (T) Ox	RPD - Check Standard	0.8299	%	9.68	9.6		0	0.15
18/10/2004	EPA 415.1 (T) Ox	Check Standard	0.95	0.1 mg/L as C	0	0	0.95	0.85	1.15
18/10/2004	EPA 415.1 (T) Ox	Check Standard	9.68	0.1 mg/L as C	0	0	0.968	0.85	1.15
18/10/2004	EPA 415.1 (T) Ox	LCS - Recovery	10.38	0.1 mg/L as C	0	0	1.038	0.85	1.15
18/10/2004	EPA 415.1 (T) Ox	LCS - Recovery	9.87	0.1 mg/L as C	0	0	0.987	0.85	1.15
18/10/2004	EPA 415.1 (T) Ox	Low Level Standard	0.56	0.1 mg/L as C	0	0	1.12	0.5	1.5
25/10/2004	EPA 415.1 (T) Ox	Low Level Standard	0.62	0.1 mg/L as C	0	0	1.24	0.5	1.5
25/10/2004	EPA 415.1 (T) Ox	RPD - Check Standard	2.817	%	1.08	1.05		0	0.15
25/10/2004	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.1992	%	10.03	10.05		0	0.15
25/10/2004	EPA 415.1 (T) Ox	RPD - Check Standard	1.603	%	10.06	9.9		0	0.15
25/10/2004	EPA 415.1 (T) Ox	RPD - Check Standard	2.817	%	1.08	1.05		0	0.15
25/10/2004	EPA 415.1 (T) Ox	Low Level Standard	0.62	0.1 mg/L as C	0	0	1.24	0.5	1.5
25/10/2004	EPA 415.1 (T) Ox	RPD - Check Standard	1.603	%	10.06	9.9		0	0.15
25/10/2004	EPA 415.1 (T) Ox	LCS - Recovery	10.03	0.1 mg/L as C	0	0	1.003	0.85	1.15
25/10/2004	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.1992	%	10.03	10.05		0	0.15
25/10/2004	EPA 415.1 (T) Ox	Check Standard	10.06	0.1 mg/L as C	0	0	1.006	0.85	1.15
25/10/2004	EPA 415.1 (T) Ox	Check Standard	1.08	0.1 mg/L as C	0	0	1.08	0.85	1.15
25/10/2004	EPA 415.1 (T) Ox	Check Standard	1.05	0.1 mg/L as C	0	0	1.05	0.85	1.15
25/10/2004	EPA 415.1 (T) Ox	Check Standard	10.06	0.1 mg/L as C	0	0	1.006	0.85	1.15
25/10/2004	EPA 415.1 (T) Ox	Check Standard	9.9	0.1 mg/L as C	0	0	0.99	0.85	1.15
25/10/2004	EPA 415.1 (T) Ox	Check Standard	1.08	0.1 mg/L as C	0	0	1.08	0.85	1.15
25/10/2004	EPA 415.1 (T) Ox	LCS - Recovery	10.05	0.1 mg/L as C	0	0	1.005	0.85	1.15
25/10/2004	EPA 415.1 (T) Ox	Check Standard	1.05	0.1 mg/L as C	0	0	1.05	0.85	1.15
25/10/2004	EPA 415.1 (T) Ox	Method Blank	0.05	0.1 mg/L as C	0	0			
25/10/2004	EPA 415.1 (T) Ox	Check Standard	9.9	0.1 mg/L as C	0	0	0.99	0.85	1.15
25/10/2004	EPA 415.1 (T) Ox	LCS - Recovery	10.03	0.1 mg/L as C	0	0	1.003	0.85	1.15
25/10/2004	EPA 415.1 (T) Ox	LCS - Recovery	10.05	0.1 mg/L as C	0	0	1.005	0.85	1.15
25/10/2004	EPA 415.1 (T) Ox	Method Blank	0.05	0.1 mg/L as C	0	0			
01/11/2004	EPA 415.1 (T) Ox	LCS - Recovery	9.78	0.1 mg/L as C	0	0	0.978	0.85	1.15
01/11/2004	EPA 415.1 (T) Ox	LCS - Recovery	9.84	0.1 mg/L as C	0	0	0.984	0.85	1.15
01/11/2004	EPA 415.1 (T) Ox	Low Level Standard	0.48	0.1 mg/L as C	0	0	0.96	0.5	1.5
01/11/2004	EPA 415.1 (T) Ox	RPD - Check Standard	2.737	%	9.73	10		0	0.15
01/11/2004	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.6116	%	9.84	9.78		0	0.15

01/11/2004	EPA 415.1 (T) Ox	Check Standard	10	0.1 mg/L as C	0	0	1	0.85	1.15
01/11/2004	EPA 415.1 (T) Ox	Check Standard	9.73	0.1 mg/L as C	0	0	0.973	0.85	1.15
01/11/2004	EPA 415.1 (T) Ox	Check Standard	1.03	0.1 mg/L as C	0	0	1.03	0.85	1.15
01/11/2004	EPA 415.1 (T) Ox	Check Standard	0.96	0.1 mg/L as C	0	0	0.96	0.85	1.15
01/11/2004	EPA 415.1 (T) Ox	Method Blank	0.001	0.1 mg/L as C	0	0			
01/11/2004	EPA 415.1 (T) Ox	RPD - Check Standard	7.035	%	0.96	1.03		0	0.15
06/12/2004	EPA 415.1 (T) Ox	RPD - LCS Duplicate	1.009	%	9.96	9.86		0	0.15
06/12/2004	EPA 415.1 (T) Ox	RPD - Check Standard	2.469	%	10	10.25		0	0.15
06/12/2004	EPA 415.1 (T) Ox	LCS - Recovery	9.86	0.1 mg/L as C	0	0	0.986	0.85	1.15
06/12/2004	EPA 415.1 (T) Ox	LCS - Recovery	9.96	0.1 mg/L as C	0	0	0.996	0.85	1.15
06/12/2004	EPA 415.1 (T) Ox	Method Blank	0	0.1 mg/L as C	0	0			
06/12/2004	EPA 415.1 (T) Ox	Check Standard	10	0.1 mg/L as C	0	0	1	0.85	1.15
06/12/2004	EPA 415.1 (T) Ox	Low Level Standard	0.56	0.1 mg/L as C	0	0	1.12	0.5	1.5
06/12/2004	EPA 415.1 (T) Ox	Check Standard	10.25	0.1 mg/L as C	0	0	1.025	0.85	1.15
03/01/2005	EPA 415.1 (T) Ox	Low Level Standard	0.55	0.1 mg/L as C	0	0	1.1	0.5	1.5
03/01/2005	EPA 415.1 (T) Ox	RPD - Check Standard	2.247	%	9.68	9.9		0	0.15
03/01/2005	EPA 415.1 (T) Ox	RPD - LCS Duplicate	1.833	%	9.91	9.73		0	0.15
03/01/2005	EPA 415.1 (T) Ox	Check Standard	9.68	0.1 mg/L as C	0	0	0.968	0.85	1.15
03/01/2005	EPA 415.1 (T) Ox	Check Standard	9.9	0.1 mg/L as C	0	0	0.99	0.85	1.15
03/01/2005	EPA 415.1 (T) Ox	LCS - Recovery	9.73	0.1 mg/L as C	0	0	0.973	0.85	1.15
03/01/2005	EPA 415.1 (T) Ox	LCS - Recovery	9.91	0.1 mg/L as C	0	0	0.991	0.85	1.15
03/01/2005	EPA 415.1 (T) Ox	Method Blank	0.01	0.1 mg/L as C	0	0			
07/02/2005	EPA 415.1 (T) Ox	RPD - LCS Duplicate	1.923	%	10.5	10.3		0	0.15
07/02/2005	EPA 415.1 (T) Ox	Check Standard	10.4	0.1 mg/L as C	0	0	1.04	0.85	1.15
07/02/2005	EPA 415.1 (T) Ox	LCS - Recovery	10.3	0.1 mg/L as C	0	0	1.03	0.85	1.15
07/02/2005	EPA 415.1 (T) Ox	Check Standard	10.4	0.1 mg/L as C	0	0	1.04	0.85	1.15
07/02/2005	EPA 415.1 (T) Ox	RPD - Check Standard	0	%	10.4	10.4		0	0.15
07/02/2005	EPA 415.1 (T) Ox	Method Blank	0.1	0.1 mg/L as C	0	0			
16/02/2005	EPA 415.1 (T) Ox	LCS - Recovery	10.4	0.1 mg/L as C	0	0	1.04	0.85	1.15
16/02/2005	EPA 415.1 (T) Ox	RPD - Check Standard	2.927	%	10.4	10.1		0	0.15
16/02/2005	EPA 415.1 (T) Ox	Check Standard	10.4	0.1 mg/L as C	0	0	1.04	0.85	1.15
16/02/2005	EPA 415.1 (T) Ox	LCS - Recovery	10.3	0.1 mg/L as C	0	0	1.03	0.85	1.15
16/02/2005	EPA 415.1 (T) Ox	LCS - Recovery	10.3	0.1 mg/L as C	0	0	1.03	0.85	1.15
16/02/2005	EPA 415.1 (T) Ox	RPD - Check Standard	2.927	%	10.4	10.1		0	0.15
16/02/2005	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.9662	%	10.4	10.3		0	0.15
16/02/2005	EPA 415.1 (T) Ox	LCS - Recovery	10.4	0.1 mg/L as C	0	0	1.04	0.85	1.15
16/02/2005	EPA 415.1 (T) Ox	Check Standard	10.1	0.1 mg/L as C	0	0	1.01	0.85	1.15

16/02/2005	EPA 415.1 (T) Ox	Check Standard	10.1	0.1 mg/L as C	0	0	1.01	0.85	1.15
16/02/2005	EPA 415.1 (T) Ox	Method Blank	0.1	0.1 mg/L as C	0	0			
16/02/2005	EPA 415.1 (T) Ox	Method Blank	0.1	0.1 mg/L as C	0	0			
16/02/2005	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.9662	%	10.4	10.3		0	0.15
16/02/2005	EPA 415.1 (T) Ox	Check Standard	10.4	0.1 mg/L as C	0	0	1.04	0.85	1.15
08/03/2005	EPA 415.1 (T) Ox	Check Standard	10	0.1 mg/L as C	0	0	1	0.85	1.15
08/03/2005	EPA 415.1 (T) Ox	Method Blank	0.1	0.1 mg/L as C	0	0			
08/03/2005	EPA 415.1 (T) Ox	Check Standard	9.9	0.1 mg/L as C	0	0	0.99	0.85	1.15
08/03/2005	EPA 415.1 (T) Ox	RPD - Check Standard	1.005	%	10	9.9		0	0.15
08/03/2005	EPA 415.1 (T) Ox	LCS - Recovery	10	0.1 mg/L as C	0	0	1	0.85	1.15
08/03/2005	EPA 415.1 (T) Ox	LCS - Recovery	9.9	0.1 mg/L as C	0	0	0.99	0.85	1.15
08/03/2005	EPA 415.1 (T) Ox	RPD - LCS Duplicate	1.005	%	10	9.9		0	0.15
16/03/2005	EPA 415.1 (T) Ox	LCS - Recovery	9.8	0.1 mg/L as C	0	0	0.98	0.85	1.15
16/03/2005	EPA 415.1 (T) Ox	LCS - Recovery	9.7	0.1 mg/L as C	0	0	0.97	0.85	1.15
16/03/2005	EPA 415.1 (T) Ox	Check Standard	9.9	0.1 mg/L as C	0	0	0.99	0.85	1.15
16/03/2005	EPA 415.1 (T) Ox	RPD - Check Standard	3.077	%	9.9	9.6		0	0.15
16/03/2005	EPA 415.1 (T) Ox	Check Standard	9.6	0.1 mg/L as C	0	0	0.96	0.85	1.15
16/03/2005	EPA 415.1 (T) Ox	RPD - LCS Duplicate	1.026	%	9.8	9.7		0	0.15
16/03/2005	EPA 415.1 (T) Ox	Method Blank	0	0.1 mg/L as C	0	0			
04/04/2005	EPA 415.1 (T) Ox	Check Standard	10	0.1 mg/L as C	0	0	1	0.85	1.15
04/04/2005	EPA 415.1 (T) Ox	RPD - Check Standard	0.995	%	10	10.1		0	0.15
04/04/2005	EPA 415.1 (T) Ox	LCS - Recovery	9.6	0.1 mg/L as C	0	0	0.96	0.85	1.15
04/04/2005	EPA 415.1 (T) Ox	Check Standard	10.1	0.1 mg/L as C	0	0	1.01	0.85	1.15
04/04/2005	EPA 415.1 (T) Ox	LCS - Recovery	9.5	0.1 mg/L as C	0	0	0.95	0.85	1.15
04/04/2005	EPA 415.1 (T) Ox	Method Blank	0	0.1 mg/L as C	0	0			
04/04/2005	EPA 415.1 (T) Ox	RPD - LCS Duplicate	1.047	%	9.5	9.6		0	0.15
28/04/2005	EPA 415.1 (T) Ox	Method Blank	0	0.5 mg/L as C	0	0			
28/04/2005	EPA 415.1 (T) Ox	RPD - LCS Duplicate	1.41	%	10	9.86		0	0.15
28/04/2005	EPA 415.1 (T) Ox	RPD - Check Standard	0.7125	%	9.79	9.86		0	0.15
28/04/2005	EPA 415.1 (T) Ox	RPD - LCS Duplicate	1.41	%	10	9.86		0	0.15
28/04/2005	EPA 415.1 (T) Ox	RPD - Check Standard	0.7125	%	9.79	9.86		0	0.15
28/04/2005	EPA 415.1 (T) Ox	Method Blank	0	0.5 mg/L as C	0	0			
28/04/2005	EPA 415.1 (T) Ox	Check Standard	9.79	0.5 mg/L as C	0	0	0.979	0.85	1.15
28/04/2005	EPA 415.1 (T) Ox	Check Standard	9.86	0.5 mg/L as C	0	0	0.986	0.85	1.15
28/04/2005	EPA 415.1 (T) Ox	Check Standard	9.79	0.5 mg/L as C	0	0	0.979	0.85	1.15
28/04/2005	EPA 415.1 (T) Ox	Check Standard	9.86	0.5 mg/L as C	0	0	0.986	0.85	1.15
28/04/2005	EPA 415.1 (T) Ox	LCS - Recovery	10	0.5 mg/L as C	0	0	1	0.85	1.15

28/04/2005	EPA 415.1 (T) Ox	LCS - Recovery	9.86	0.5 mg/L as C	0	0	0.986	0.85	1.15
28/04/2005	EPA 415.1 (T) Ox	LCS - Recovery	10	0.5 mg/L as C	0	0	1	0.85	1.15
28/04/2005	EPA 415.1 (T) Ox	LCS - Recovery	9.86	0.5 mg/L as C	0	0	0.986	0.85	1.15
02/05/2005	EPA 415.1 (T) Ox	LCS - Recovery	9.5	0.5 mg/L as C	0	0	0.95	0.85	1.15
02/05/2005	EPA 415.1 (T) Ox	Check Standard	10	0.5 mg/L as C	0	0	1	0.85	1.15
02/05/2005	EPA 415.1 (T) Ox	Check Standard	9.7	0.5 mg/L as C	0	0	0.97	0.85	1.15
02/05/2005	EPA 415.1 (T) Ox	LCS - Recovery	9.8	0.5 mg/L as C	0	0	0.98	0.85	1.15
02/05/2005	EPA 415.1 (T) Ox	RPD - Check Standard	3.046	%	10	9.7		0	0.15
02/05/2005	EPA 415.1 (T) Ox	RPD - LCS Duplicate	3.109	%	9.8	9.5		0	0.15
02/05/2005	EPA 415.1 (T) Ox	Method Blank	0	0.5 mg/L as C	0	0			
18/05/2005	EPA 415.1 (T) Ox	Check Standard	9.3	0.5 mg/L as C	0	0	0.93	0.85	1.15
18/05/2005	EPA 415.1 (T) Ox	Check Standard	9.3	0.5 mg/L as C	0	0	0.93	0.85	1.15
18/05/2005	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0	%	9.5	9.5		0	0.15
18/05/2005	EPA 415.1 (T) Ox	Check Standard	9.3	0.5 mg/L as C	0	0	0.93	0.85	1.15
18/05/2005	EPA 415.1 (T) Ox	RPD - Check Standard	3.279	%	9	9.3		0	0.15
18/05/2005	EPA 415.1 (T) Ox	Check Standard	9	0.5 mg/L as C	0	0	0.9	0.85	1.15
18/05/2005	EPA 415.1 (T) Ox	RPD - Check Standard	3.279	%	9	9.3		0	0.15
18/05/2005	EPA 415.1 (T) Ox	Check Standard	9	0.5 mg/L as C	0	0	0.9	0.85	1.15
18/05/2005	EPA 415.1 (T) Ox	Check Standard	9	0.5 mg/L as C	0	0	0.9	0.85	1.15
18/05/2005	EPA 415.1 (T) Ox	Check Standard	9	0.5 mg/L as C	0	0	0.9	0.85	1.15
18/05/2005	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0	%	9.5	9.5		0	0.15
18/05/2005	EPA 415.1 (T) Ox	Method Blank	0	0.5 mg/L as C	0	0			
18/05/2005	EPA 415.1 (T) Ox	LCS - Recovery	9.5	0.5 mg/L as C	0	0	0.95	0.85	1.15
18/05/2005	EPA 415.1 (T) Ox	LCS - Recovery	9.5	0.5 mg/L as C	0	0	0.95	0.85	1.15
18/05/2005	EPA 415.1 (T) Ox	Method Blank	0	0.5 mg/L as C	0	0			
18/05/2005	EPA 415.1 (T) Ox	Check Standard	9.3	0.5 mg/L as C	0	0	0.93	0.85	1.15
18/05/2005	EPA 415.1 (T) Ox	LCS - Recovery	9.5	0.5 mg/L as C	0	0	0.95	0.85	1.15
18/05/2005	EPA 415.1 (T) Ox	RPD - Check Standard	3.279	%	9	9.3		0	0.15
18/05/2005	EPA 415.1 (T) Ox	RPD - Check Standard	3.279	%	9	9.3		0	0.15
18/05/2005	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0	%	9.5	9.5		0	0.15
18/05/2005	EPA 415.1 (T) Ox	LCS - Recovery	9.5	0.5 mg/L as C	0	0	0.95	0.85	1.15
18/05/2005	EPA 415.1 (T) Ox	LCS - Recovery	9.5	0.5 mg/L as C	0	0	0.95	0.85	1.15
18/05/2005	EPA 415.1 (T) Ox	Method Blank	0	0.5 mg/L as C	0	0			
18/05/2005	EPA 415.1 (T) Ox	LCS - Recovery	9.5	0.5 mg/L as C	0	0	0.95	0.85	1.15
18/05/2005	EPA 415.1 (T) Ox	Method Blank	0	0.5 mg/L as C	0	0			
18/05/2005	EPA 415.1 (T) Ox	LCS - Recovery	9.5	0.5 mg/L as C	0	0	0.95	0.85	1.15
18/05/2005	EPA 415.1 (T) Ox	LCS - Recovery	9.5	0.5 mg/L as C	0	0	0.95	0.85	1.15

18/05/2005	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0	%	9.5	9.5		0	0.15
06/06/2005	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.9142	%	9.8	9.89		0	0.15
06/06/2005	EPA 415.1 (T) Ox	LCS - Recovery	9.8	0.5 mg/L as C	0	0	0.98	0.85	1.15
06/06/2005	EPA 415.1 (T) Ox	LCS - Recovery	9.89	0.5 mg/L as C	0	0	0.989	0.85	1.15
06/06/2005	EPA 415.1 (T) Ox	RPD - Check Standard	0.9188	%	9.75	9.84		0	0.15
06/06/2005	EPA 415.1 (T) Ox	Method Blank	0	0.5 mg/L as C	0	0			
06/06/2005	EPA 415.1 (T) Ox	Check Standard	9.84	0.5 mg/L as C	0	0	0.984	0.85	1.15
06/06/2005	EPA 415.1 (T) Ox	Check Standard	9.75	0.5 mg/L as C	0	0	0.975	0.85	1.15
05/07/2005	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.9479	%	10.6	10.5		0	0.15
05/07/2005	EPA 415.1 (T) Ox	RPD - Check Standard	0.9662	%	10.3	10.4		0	0.15
05/07/2005	EPA 415.1 (T) Ox	Check Standard	10.4	0.5 mg/L as C	0	0	1.04	0.85	1.15
05/07/2005	EPA 415.1 (T) Ox	Method Blank	0	0.5 mg/L as C	0	0			
05/07/2005	EPA 415.1 (T) Ox	Check Standard	10.3	0.5 mg/L as C	0	0	1.03	0.85	1.15
01/08/2005	EPA 415.1 (T) Ox	LCS - Recovery	10.2	0.5 mg/L as C	0	0	1.02	0.85	1.15
01/08/2005	EPA 415.1 (T) Ox	RPD - Check Standard	0.9756	%	10.3	10.2		0	0.15
01/08/2005	EPA 415.1 (T) Ox	Check Standard	10.2	0.5 mg/L as C	0	0	1.02	0.85	1.15
01/08/2005	EPA 415.1 (T) Ox	Check Standard	10.3	0.5 mg/L as C	0	0	1.03	0.85	1.15
01/08/2005	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.9756	%	10.2	10.3		0	0.15
01/08/2005	EPA 415.1 (T) Ox	Method Blank	0	0.5 mg/L as C	0	0			
01/08/2005	EPA 415.1 (T) Ox	Check Standard	10.2	0.5 mg/L as C	0	0	1.02	0.85	1.15
01/08/2005	EPA 415.1 (T) Ox	RPD - Check Standard	0.9756	%	10.3	10.2		0	0.15
01/08/2005	EPA 415.1 (T) Ox	Method Blank	0	0.5 mg/L as C	0	0			
01/08/2005	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.9756	%	10.2	10.3		0	0.15
01/08/2005	EPA 415.1 (T) Ox	Check Standard	10.3	0.5 mg/L as C	0	0	1.03	0.85	1.15
01/08/2005	EPA 415.1 (T) Ox	LCS - Recovery	10.2	0.5 mg/L as C	0	0	1.02	0.85	1.15
01/08/2005	EPA 415.1 (T) Ox	LCS - Recovery	10.3	0.5 mg/L as C	0	0	1.03	0.85	1.15
01/08/2005	EPA 415.1 (T) Ox	LCS - Recovery	10.3	0.5 mg/L as C	0	0	1.03	0.85	1.15
25/08/2005	EPA 415.1 (T) Ox	Check Standard	9.97	0.5 mg/L as C	0	0	0.997	0.85	1.15
25/08/2005	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.6079	%	9.84	9.9		0	0.15
25/08/2005	EPA 415.1 (T) Ox	Check Standard	9.97	0.5 mg/L as C	0	0	0.997	0.85	1.15
25/08/2005	EPA 415.1 (T) Ox	RPD - Check Standard	4.094	%	9.97	9.57		0	0.15
25/08/2005	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.1022	%	9.79	9.78		0	0.15
25/08/2005	EPA 415.1 (T) Ox	RPD - Check Standard	0.995	%	10.1	10		0	0.15
25/08/2005	EPA 415.1 (T) Ox	RPD - Check Standard	4.094	%	9.97	9.57		0	0.15
25/08/2005	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.6079	%	9.84	9.9		0	0.15
25/08/2005	EPA 415.1 (T) Ox	Method Blank	0	0.5 mg/L as C	0	0			
25/08/2005	EPA 415.1 (T) Ox	Check Standard	9.57	0.5 mg/L as C	0	0	0.957	0.85	1.15

25/08/2005	EPA 415.1 (T) Ox	LCS - Recovery	9.9	0.5 mg/L as C	0	0	0.99	0.85	1.15
25/08/2005	EPA 415.1 (T) Ox	Check Standard	9.57	0.5 mg/L as C	0	0	0.957	0.85	1.15
25/08/2005	EPA 415.1 (T) Ox	Check Standard	10.1	0.5 mg/L as C	0	0	1.01	0.85	1.15
25/08/2005	EPA 415.1 (T) Ox	Check Standard	10	0.5 mg/L as C	0	0	1	0.85	1.15
25/08/2005	EPA 415.1 (T) Ox	Method Blank	0	0.5 mg/L as C	0	0			
25/08/2005	EPA 415.1 (T) Ox	Method Blank	0	0.5 mg/L as C	0	0			
25/08/2005	EPA 415.1 (T) Ox	LCS - Recovery	9.84	0.5 mg/L as C	0	0	0.984	0.85	1.15
25/08/2005	EPA 415.1 (T) Ox	LCS - Recovery	9.9	0.5 mg/L as C	0	0	0.99	0.85	1.15
25/08/2005	EPA 415.1 (T) Ox	LCS - Recovery	9.79	0.5 mg/L as C	0	0	0.979	0.85	1.15
25/08/2005	EPA 415.1 (T) Ox	LCS - Recovery	9.78	0.5 mg/L as C	0	0	0.978	0.85	1.15
25/08/2005	EPA 415.1 (T) Ox	LCS - Recovery	9.84	0.5 mg/L as C	0	0	0.984	0.85	1.15
06/09/2005	EPA 415.1 (T) Ox	Check Standard	9.91	0.5 mg/L as C	0	0	0.991	0.85	1.15
06/09/2005	EPA 415.1 (T) Ox	LCS - Recovery	9.71	0.5 mg/L as C	0	0	0.971	0.85	1.15
06/09/2005	EPA 415.1 (T) Ox	LCS - Recovery	9.8	0.5 mg/L as C	0	0	0.98	0.85	1.15
06/09/2005	EPA 415.1 (T) Ox	RPD - Check Standard	1.936	%	9.72	9.91		0	0.15
06/09/2005	EPA 415.1 (T) Ox	Check Standard	9.72	0.5 mg/L as C	0	0	0.972	0.85	1.15
06/09/2005	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.9226	%	9.8	9.71		0	0.15
06/09/2005	EPA 415.1 (T) Ox	LCS - Recovery	9.71	0.5 mg/L as C	0	0	0.971	0.85	1.15
06/09/2005	EPA 415.1 (T) Ox	Method Blank	0	0.5 mg/L as C	0	0			
06/09/2005	EPA 415.1 (T) Ox	Check Standard	9.72	0.5 mg/L as C	0	0	0.972	0.85	1.15
06/09/2005	EPA 415.1 (T) Ox	Check Standard	9.91	0.5 mg/L as C	0	0	0.991	0.85	1.15
06/09/2005	EPA 415.1 (T) Ox	Method Blank	0	0.5 mg/L as C	0	0			
06/09/2005	EPA 415.1 (T) Ox	RPD - Check Standard	1.936	%	9.72	9.91		0	0.15
06/09/2005	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.9226	%	9.8	9.71		0	0.15
06/09/2005	EPA 415.1 (T) Ox	LCS - Recovery	9.8	0.5 mg/L as C	0	0	0.98	0.85	1.15
03/10/2005	EPA 415.1 (T) Ox	Method Blank	0	0.5 mg/L as C	0	0			
03/10/2005	EPA 415.1 (T) Ox	Check Standard	9.4	0.5 mg/L as C	0	0	0.94	0.85	1.15
03/10/2005	EPA 415.1 (T) Ox	LCS - Recovery	9.41	0.5 mg/L as C	0	0	0.941	0.85	1.15
03/10/2005	EPA 415.1 (T) Ox	RPD - LCS Duplicate	1.79	%	9.58	9.41		0	0.15
03/10/2005	EPA 415.1 (T) Ox	LCS - Recovery	9.58	0.5 mg/L as C	0	0	0.958	0.85	1.15
03/10/2005	EPA 415.1 (T) Ox	RPD - Check Standard	2.001	%	9.59	9.4		0	0.15
03/10/2005	EPA 415.1 (T) Ox	Check Standard	9.59	0.5 mg/L as C	0	0	0.959	0.85	1.15
07/11/2005	EPA 415.1 (T) Ox	LCS - Recovery	9.59	0.5 mg/L as C	0	0	0.959	0.85	1.15
07/11/2005	EPA 415.1 (T) Ox	LCS - Recovery	9.66	0.5 mg/L as C	0	0	0.966	0.85	1.15
07/11/2005	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.7273	%	9.59	9.66		0	0.15
07/11/2005	EPA 415.1 (T) Ox	Method Blank	0	0.5 mg/L as C	0	0			
07/11/2005	EPA 415.1 (T) Ox	Check Standard	9.81	0.5 mg/L as C	0	0	0.981	0.85	1.15

07/11/2005	EPA 415.1 (T) Ox	RPD - Check Standard	0.6135	%	9.81	9.75	0	0.15	
07/11/2005	EPA 415.1 (T) Ox	Check Standard	9.75	0.5 mg/L as C	0	0	0.975	0.85	1.15
22/11/2005	EPA 415.1 (T) Ox	Method Blank	0	0.5 mg/L as C	0	0			
22/11/2005	EPA 415.1 (T) Ox	LCS - Recovery	9.58	0.5 mg/L as C	0	0	0.958	0.85	1.15
22/11/2005	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.4167	%	9.62	9.58	0	0.15	
22/11/2005	EPA 415.1 (T) Ox	RPD - Check Standard	0.8214	%	9.7	9.78	0	0.15	
22/11/2005	EPA 415.1 (T) Ox	LCS - Recovery	9.62	0.5 mg/L as C	0	0	0.962	0.85	1.15
22/11/2005	EPA 415.1 (T) Ox	Check Standard	9.7	0.5 mg/L as C	0	0	0.97	0.85	1.15
22/11/2005	EPA 415.1 (T) Ox	Check Standard	9.78	0.5 mg/L as C	0	0	0.978	0.85	1.15
22/11/2005	EPA 415.1 (T) Ox	Check Standard	9.7	0.5 mg/L as C	0	0	0.97	0.85	1.15
22/11/2005	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.4167	%	9.62	9.58	0	0.15	
22/11/2005	EPA 415.1 (T) Ox	Check Standard	9.78	0.5 mg/L as C	0	0	0.978	0.85	1.15
22/11/2005	EPA 415.1 (T) Ox	LCS - Recovery	9.58	0.5 mg/L as C	0	0	0.958	0.85	1.15
22/11/2005	EPA 415.1 (T) Ox	LCS - Recovery	9.62	0.5 mg/L as C	0	0	0.962	0.85	1.15
22/11/2005	EPA 415.1 (T) Ox	Method Blank	0	0.5 mg/L as C	0	0			
22/11/2005	EPA 415.1 (T) Ox	RPD - Check Standard	0.8214	%	9.7	9.78	0	0.15	
05/12/2005	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.5147	%	9.74	9.69	0	0.15	
05/12/2005	EPA 415.1 (T) Ox	Method Blank	0	0.5 mg/L as C	0	0			
05/12/2005	EPA 415.1 (T) Ox	LCS - Recovery	9.74	0.5 mg/L as C	0	0	0.974	0.85	1.15
05/12/2005	EPA 415.1 (T) Ox	LCS - Recovery	9.69	0.5 mg/L as C	0	0	0.969	0.85	1.15
05/12/2005	EPA 415.1 (T) Ox	Check Standard	9.94	0.5 mg/L as C	0	0	0.994	0.85	1.15
05/12/2005	EPA 415.1 (T) Ox	RPD - Check Standard	0.6054	%	9.94	9.88	0	0.15	
05/12/2005	EPA 415.1 (T) Ox	Check Standard	9.88	0.5 mg/L as C	0	0	0.988	0.85	1.15
20/12/2005	EPA 415.1 (T) Ox	Check Standard	9.84	0.5 mg/L as C	0	0	0.984	0.85	1.15
20/12/2005	EPA 415.1 (T) Ox	LCS - Recovery	9.79	0.5 mg/L as C	0	0	0.979	0.85	1.15
20/12/2005	EPA 415.1 (T) Ox	LCS - Recovery	9.81	0.5 mg/L as C	0	0	0.981	0.85	1.15
20/12/2005	EPA 415.1 (T) Ox	Method Blank	0	0.5 mg/L as C	0	0			
20/12/2005	EPA 415.1 (T) Ox	Check Standard	9.91	0.5 mg/L as C	0	0	0.991	0.85	1.15
20/12/2005	EPA 415.1 (T) Ox	RPD - LCS Duplicate	0.2041	%	9.79	9.81	0	0.15	
20/12/2005	EPA 415.1 (T) Ox	RPD - Check Standard	0.7089	%	9.91	9.84	0	0.15	
03/01/2006	EPA 415.3-1 (T) Ox	RPD - Check Standard	2	%	9.9	10.1	0	0.15	
03/01/2006	EPA 415.3-1 (T) Ox	Check Standard	10.1	0.5 mg/L as C	0	0	1.01	0.85	1.15
03/01/2006	EPA 415.3-1 (T) Ox	LCS - Recovery	9.85	0.5 mg/L as C	0	0	0.985	0.85	1.15
03/01/2006	EPA 415.3-1 (T) Ox	RPD - LCS Duplicate	0.4053	%	9.89	9.85	0	0.15	
03/01/2006	EPA 415.3-1 (T) Ox	Method Blank	0	0.5 mg/L as C	0	0			
03/01/2006	EPA 415.3-1 (T) Ox	Check Standard	9.9	0.5 mg/L as C	0	0	0.99	0.85	1.15
03/01/2006	EPA 415.3-1 (T) Ox	LCS - Recovery	9.89	0.5 mg/L as C	0	0	0.989	0.85	1.15

17/01/2006	EPA 415.3-1 (T) Ox	RPD - Check Standard	0.1026	%	9.74	9.75	0	0.15	
17/01/2006	EPA 415.3-1 (T) Ox	Check Standard	9.75	0.5 mg/L as C	0	0	0.975	0.85	1.15
17/01/2006	EPA 415.3-1 (T) Ox	Check Standard	9.74	0.5 mg/L as C	0	0	0.974	0.85	1.15
17/01/2006	EPA 415.3-1 (T) Ox	Method Blank	0	0.5 mg/L as C	0	0			
17/01/2006	EPA 415.3-1 (T) Ox	RPD - LCS Duplicate	2.096	%	9.44	9.64	0	0.15	
17/01/2006	EPA 415.3-1 (T) Ox	LCS - Recovery	9.44	0.5 mg/L as C	0	0	0.944	0.85	1.15
17/01/2006	EPA 415.3-1 (T) Ox	LCS - Recovery	9.64	0.5 mg/L as C	0	0	0.964	0.85	1.15
06/02/2006	EPA 415.3-1 (T) Ox	LCS - Recovery	10.1	0.5 mg/L as C	0	0	1.01	0.85	1.15
06/02/2006	EPA 415.3-1 (T) Ox	RPD - Check Standard	7.422	%	10.2	9.47	0	0.15	
06/02/2006	EPA 415.3-1 (T) Ox	RPD - LCS Duplicate	1.798	%	10.1	9.92	0	0.15	
06/02/2006	EPA 415.3-1 (T) Ox	Check Standard	10.2	0.5 mg/L as C	0	0	1.02	0.85	1.15
06/02/2006	EPA 415.3-1 (T) Ox	Method Blank	0	0.5 mg/L as C	0	0			
06/02/2006	EPA 415.3-1 (T) Ox	Check Standard	9.47	0.5 mg/L as C	0	0	0.947	0.85	1.15
06/02/2006	EPA 415.3-1 (T) Ox	LCS - Recovery	9.92	0.5 mg/L as C	0	0	0.992	0.85	1.15
23/02/2006	EPA 415.3-1 (T) Ox	Method Blank	0	0.5 mg/L as C	0	0			
23/02/2006	EPA 415.3-1 (T) Ox	Check Standard	9.37	0.5 mg/L as C	0	0	0.937	0.85	1.15
23/02/2006	EPA 415.3-1 (T) Ox	Check Standard	9.45	0.5 mg/L as C	0	0	0.945	0.85	1.15
23/02/2006	EPA 415.3-1 (T) Ox	Method Blank	0	0.5 mg/L as C	0	0			
23/02/2006	EPA 415.3-1 (T) Ox	LCS - Recovery	9.3	0.5 mg/L as C	0	0	0.93	0.85	1.15
23/02/2006	EPA 415.3-1 (T) Ox	Check Standard	9.45	0.5 mg/L as C	0	0	0.945	0.85	1.15
23/02/2006	EPA 415.3-1 (T) Ox	RPD - LCS Duplicate	0.3231	%	9.27	9.3	0	0.15	
23/02/2006	EPA 415.3-1 (T) Ox	Check Standard	9.37	0.5 mg/L as C	0	0	0.937	0.85	1.15
23/02/2006	EPA 415.3-1 (T) Ox	RPD - LCS Duplicate	0.3231	%	9.27	9.3	0	0.15	
23/02/2006	EPA 415.3-1 (T) Ox	RPD - Check Standard	0.8502	%	9.37	9.45	0	0.15	
23/02/2006	EPA 415.3-1 (T) Ox	RPD - Check Standard	0.8502	%	9.37	9.45	0	0.15	
23/02/2006	EPA 415.3-1 (T) Ox	LCS - Recovery	9.3	0.5 mg/L as C	0	0	0.93	0.85	1.15
23/02/2006	EPA 415.3-1 (T) Ox	LCS - Recovery	9.27	0.5 mg/L as C	0	0	0.927	0.85	1.15
23/02/2006	EPA 415.3-1 (T) Ox	LCS - Recovery	9.27	0.5 mg/L as C	0	0	0.927	0.85	1.15
06/03/2006	EPA 415.3-1 (T) Ox	LCS - Recovery	9.84	0.5 mg/L as C	0	0	0.984	0.85	1.15
06/03/2006	EPA 415.3-1 (T) Ox	Method Blank	0	0.5 mg/L as C	0	0			
06/03/2006	EPA 415.3-1 (T) Ox	Method Blank	0	0.5 mg/L as C	0	0			
06/03/2006	EPA 415.3-1 (T) Ox	RPD - LCS Duplicate	0	%	9.84	9.84	0	0.15	
06/03/2006	EPA 415.3-1 (T) Ox	Method Blank	0	0.5 mg/L as C	0	0			
06/03/2006	EPA 415.3-1 (T) Ox	RPD - Check Standard	3.694	%	9.83	10.2	0	0.15	
06/03/2006	EPA 415.3-1 (T) Ox	LCS - Recovery	9.84	0.5 mg/L as C	0	0	0.984	0.85	1.15
06/03/2006	EPA 415.3-1 (T) Ox	RPD - Check Standard	3.694	%	9.83	10.2	0	0.15	
06/03/2006	EPA 415.3-1 (T) Ox	LCS - Recovery	9.84	0.5 mg/L as C	0	0	0.984	0.85	1.15

06/03/2006	EPA 415.3-1 (T) Ox	LCS - Recovery	9.84	0.5 mg/L as C	0	0	0.984	0.85	1.15
06/03/2006	EPA 415.3-1 (T) Ox	LCS - Recovery	9.84	0.5 mg/L as C	0	0	0.984	0.85	1.15
06/03/2006	EPA 415.3-1 (T) Ox	LCS - Recovery	9.84	0.5 mg/L as C	0	0	0.984	0.85	1.15
06/03/2006	EPA 415.3-1 (T) Ox	RPD - LCS Duplicate	0	%	9.84	9.84		0	0.15
06/03/2006	EPA 415.3-1 (T) Ox	Check Standard	10.2	0.5 mg/L as C	0	0	1.02	0.85	1.15
06/03/2006	EPA 415.3-1 (T) Ox	Check Standard	9.83	0.5 mg/L as C	0	0	0.983	0.85	1.15
06/03/2006	EPA 415.3-1 (T) Ox	RPD - Check Standard	3.694	%	9.83	10.2		0	0.15
06/03/2006	EPA 415.3-1 (T) Ox	RPD - LCS Duplicate	0	%	9.84	9.84		0	0.15
06/03/2006	EPA 415.3-1 (T) Ox	Check Standard	10.2	0.5 mg/L as C	0	0	1.02	0.85	1.15
06/03/2006	EPA 415.3-1 (T) Ox	Check Standard	9.83	0.5 mg/L as C	0	0	0.983	0.85	1.15
06/03/2006	EPA 415.3-1 (T) Ox	Check Standard	10.2	0.5 mg/L as C	0	0	1.02	0.85	1.15
06/03/2006	EPA 415.3-1 (T) Ox	Check Standard	9.83	0.5 mg/L as C	0	0	0.983	0.85	1.15
23/03/2006	EPA 415.3-1 (T) Ox	LCS - Recovery	10	0.5 mg/L as C	0	0	1	0.85	1.15
23/03/2006	EPA 415.3-1 (T) Ox	RPD - Check Standard	0.8247	%	9.74	9.66		0	0.15
23/03/2006	EPA 415.3-1 (T) Ox	RPD - Check Standard	0.8247	%	9.74	9.66		0	0.15
23/03/2006	EPA 415.3-1 (T) Ox	RPD - Check Standard	0.8247	%	9.74	9.66		0	0.15
23/03/2006	EPA 415.3-1 (T) Ox	RPD - LCS Duplicate	1.613	%	10	9.84		0	0.15
23/03/2006	EPA 415.3-1 (T) Ox	LCS - Recovery	9.84	0.5 mg/L as C	0	0	0.984	0.85	1.15
23/03/2006	EPA 415.3-1 (T) Ox	Check Standard	9.66	0.5 mg/L as C	0	0	0.966	0.85	1.15
23/03/2006	EPA 415.3-1 (T) Ox	LCS - Recovery	9.84	0.5 mg/L as C	0	0	0.984	0.85	1.15
23/03/2006	EPA 415.3-1 (T) Ox	RPD - LCS Duplicate	1.613	%	10	9.84		0	0.15
23/03/2006	EPA 415.3-1 (T) Ox	Check Standard	9.74	0.5 mg/L as C	0	0	0.974	0.85	1.15
23/03/2006	EPA 415.3-1 (T) Ox	Check Standard	9.66	0.5 mg/L as C	0	0	0.966	0.85	1.15
23/03/2006	EPA 415.3-1 (T) Ox	LCS - Recovery	9.84	0.5 mg/L as C	0	0	0.984	0.85	1.15
23/03/2006	EPA 415.3-1 (T) Ox	LCS - Recovery	10	0.5 mg/L as C	0	0	1	0.85	1.15
23/03/2006	EPA 415.3-1 (T) Ox	Check Standard	9.74	0.5 mg/L as C	0	0	0.974	0.85	1.15
23/03/2006	EPA 415.3-1 (T) Ox	RPD - LCS Duplicate	1.613	%	10	9.84		0	0.15
23/03/2006	EPA 415.3-1 (T) Ox	LCS - Recovery	10	0.5 mg/L as C	0	0	1	0.85	1.15
23/03/2006	EPA 415.3-1 (T) Ox	Check Standard	9.66	0.5 mg/L as C	0	0	0.966	0.85	1.15
23/03/2006	EPA 415.3-1 (T) Ox	Check Standard	9.74	0.5 mg/L as C	0	0	0.974	0.85	1.15
23/03/2006	EPA 415.3-1 (T) Ox	Method Blank	0	0.5 mg/L as C	0	0			
23/03/2006	EPA 415.3-1 (T) Ox	Method Blank	0	0.5 mg/L as C	0	0			
23/03/2006	EPA 415.3-1 (T) Ox	Method Blank	0	0.5 mg/L as C	0	0			
06/04/2006	EPA 415.3-1 (T) Ox	LCS - Recovery	9.71	0.5 mg/L as C	0	0	0.971	0.85	1.15
06/04/2006	EPA 415.3-1 (T) Ox	LCS - Recovery	9.98	0.5 mg/L as C	0	0	0.998	0.85	1.15
06/04/2006	EPA 415.3-1 (T) Ox	Check Standard	9.96	0.5 mg/L as C	0	0	0.996	0.85	1.15
06/04/2006	EPA 415.3-1 (T) Ox	Check Standard	10.1	0.5 mg/L as C	0	0	1.01	0.85	1.15

06/04/2006	EPA 415.3-1 (T) Ox	RPD - Check Standard	1.396	%	10.1	9.96	0	0.15
06/04/2006	EPA 415.3-1 (T) Ox	RPD - LCS Duplicate	2.743	%	9.98	9.71	0	0.15
06/04/2006	EPA 415.3-1 (T) Ox	Method Blank	0	0.5 mg/L as C	0	0		
20/04/2006	EPA 415.3-1 (T) Ox	Method Blank	0	0.5 mg/L as C	0	0		
20/04/2006	EPA 415.3-1 (T) Ox	RPD - LCS Duplicate	0	%	9.6	9.6	0	0.15
20/04/2006	EPA 415.3-1 (T) Ox	RPD - Check Standard	3.015	%	9.8	10.1	0	0.15
20/04/2006	EPA 415.3-1 (T) Ox	Check Standard	10.1	0.5 mg/L as C	0	0	1.01	0.85
20/04/2006	EPA 415.3-1 (T) Ox	Check Standard	9.8	0.5 mg/L as C	0	0	0.98	0.85
20/04/2006	EPA 415.3-1 (T) Ox	Check Standard	10.1	0.5 mg/L as C	0	0	1.01	0.85
20/04/2006	EPA 415.3-1 (T) Ox	LCS - Recovery	9.6	0.5 mg/L as C	0	0	0.96	0.85
20/04/2006	EPA 415.3-1 (T) Ox	RPD - Check Standard	3.015	%	9.8	10.1	0	0.15
20/04/2006	EPA 415.3-1 (T) Ox	Method Blank	0	0.5 mg/L as C	0	0		
20/04/2006	EPA 415.3-1 (T) Ox	Check Standard	9.8	0.5 mg/L as C	0	0	0.98	0.85
20/04/2006	EPA 415.3-1 (T) Ox	RPD - LCS Duplicate	0	%	9.6	9.6	0	0.15
20/04/2006	EPA 415.3-1 (T) Ox	LCS - Recovery	9.6	0.5 mg/L as C	0	0	0.96	0.85
20/04/2006	EPA 415.3-1 (T) Ox	LCS - Recovery	9.6	0.5 mg/L as C	0	0	0.96	0.85
20/04/2006	EPA 415.3-1 (T) Ox	LCS - Recovery	9.6	0.5 mg/L as C	0	0	0.96	0.85
22/05/2006	EPA 415.3-1 (T) Ox	LCS - Recovery	9.7	0.5 mg/L as C	0	0	0.97	0.85
22/05/2006	EPA 415.3-1 (T) Ox	LCS - Recovery	9.67	0.5 mg/L as C	0	0	0.967	0.85
22/05/2006	EPA 415.3-1 (T) Ox	LCS - Recovery	9.67	0.5 mg/L as C	0	0	0.967	0.85
22/05/2006	EPA 415.3-1 (T) Ox	Check Standard	9.6	0.5 mg/L as C	0	0	0.96	0.85
22/05/2006	EPA 415.3-1 (T) Ox	Check Standard	9.6	0.5 mg/L as C	0	0	0.96	0.85
22/05/2006	EPA 415.3-1 (T) Ox	Check Standard	9.6	0.5 mg/L as C	0	0	0.96	0.85
22/05/2006	EPA 415.3-1 (T) Ox	LCS - Recovery	9.7	0.5 mg/L as C	0	0	0.97	0.85
22/05/2006	EPA 415.3-1 (T) Ox	RPD - Check Standard	0	%	9.6	9.6	0	0.15
22/05/2006	EPA 415.3-1 (T) Ox	Check Standard	9.6	0.5 mg/L as C	0	0	0.96	0.85
22/05/2006	EPA 415.3-1 (T) Ox	Method Blank	0	0.5 mg/L as C	0	0		
22/05/2006	EPA 415.3-1 (T) Ox	Method Blank	0	0.5 mg/L as C	0	0		
22/05/2006	EPA 415.3-1 (T) Ox	RPD - LCS Duplicate	0.3098	%	9.7	9.67	0	0.15
22/05/2006	EPA 415.3-1 (T) Ox	RPD - LCS Duplicate	0.3098	%	9.7	9.67	0	0.15
22/05/2006	EPA 415.3-1 (T) Ox	RPD - Check Standard	0	%	9.6	9.6	0	0.15
20/06/2006	EPA 415.3-1 (T) Ox	Method Blank	0	0.5 mg/L as C	0	0		
20/06/2006	EPA 415.3-1 (T) Ox	LCS - Recovery	10.2	0.5 mg/L as C	0	0	1.02	0.85
20/06/2006	EPA 415.3-1 (T) Ox	LCS - Recovery	10.2	0.5 mg/L as C	0	0	1.02	0.85
20/06/2006	EPA 415.3-1 (T) Ox	Check Standard	10.2	0.5 mg/L as C	0	0	1.02	0.85
20/06/2006	EPA 415.3-1 (T) Ox	Check Standard	9.93	0.5 mg/L as C	0	0	0.993	0.85
20/06/2006	EPA 415.3-1 (T) Ox	RPD - LCS Duplicate	0	%	10.2	10.2	0	0.15

20/06/2006	EPA 415.3-1 (T) Ox	RPD - Check Standard	2.683	%	9.93	10.2	0	0.15	
03/07/2006	EPA 415.3-1 (T) Ox	RPD - LCS Duplicate	2.062	%	9.6	9.8	0	0.15	
03/07/2006	EPA 415.3-1 (T) Ox	Check Standard	9.4	0.5 mg/L as C	0	0	0.94	0.85	1.15
03/07/2006	EPA 415.3-1 (T) Ox	Method Blank	0	0.5 mg/L as C	0	0			
03/07/2006	EPA 415.3-1 (T) Ox	LCS - Recovery	9.6	0.5 mg/L as C	0	0	0.96	0.85	1.15
03/07/2006	EPA 415.3-1 (T) Ox	LCS - Recovery	9.8	0.5 mg/L as C	0	0	0.98	0.85	1.15
03/07/2006	EPA 415.3-1 (T) Ox	Check Standard	10	0.5 mg/L as C	0	0	1	0.85	1.15
03/07/2006	EPA 415.3-1 (T) Ox	RPD - Check Standard	6.186	%	9.4	10	0	0.15	
24/08/2006	EPA 415.3-1 (T) Ox	LCS - Recovery	10.06	0.5 mg/L as C	0	0	1.006	0.85	1.15
24/08/2006	EPA 415.3-1 (T) Ox	LCS - Recovery	9.86	0.5 mg/L as C	0	0	0.986	0.85	1.15
24/08/2006	EPA 415.3-1 (T) Ox	Method Blank	0	0.5 mg/L as C	0	0			
24/08/2006	EPA 415.3-1 (T) Ox	Method Blank	0	0.5 mg/L as C	0	0			
24/08/2006	EPA 415.3-1 (T) Ox	Method Blank	0	0.5 mg/L as C	0	0			
24/08/2006	EPA 415.3-1 (T) Ox	LCS - Recovery	9.86	0.5 mg/L as C	0	0	0.986	0.85	1.15
24/08/2006	EPA 415.3-1 (T) Ox	Check Standard	9.46	0.5 mg/L as C	0	0	0.946	0.85	1.15
24/08/2006	EPA 415.3-1 (T) Ox	RPD - LCS Duplicate	2.008	%	10.06	9.86	0	0.15	
24/08/2006	EPA 415.3-1 (T) Ox	RPD - Check Standard	5.35	%	9.98	9.46	0	0.15	
24/08/2006	EPA 415.3-1 (T) Ox	RPD - LCS Duplicate	2.008	%	10.06	9.86	0	0.15	
24/08/2006	EPA 415.3-1 (T) Ox	RPD - Check Standard	5.35	%	9.98	9.46	0	0.15	
24/08/2006	EPA 415.3-1 (T) Ox	LCS - Recovery	10.06	0.5 mg/L as C	0	0	1.006	0.85	1.15
24/08/2006	EPA 415.3-1 (T) Ox	Check Standard	9.98	0.5 mg/L as C	0	0	0.998	0.85	1.15
24/08/2006	EPA 415.3-1 (T) Ox	LCS - Recovery	9.86	0.5 mg/L as C	0	0	0.986	0.85	1.15
24/08/2006	EPA 415.3-1 (T) Ox	Check Standard	9.98	0.5 mg/L as C	0	0	0.998	0.85	1.15
24/08/2006	EPA 415.3-1 (T) Ox	Check Standard	9.46	0.5 mg/L as C	0	0	0.946	0.85	1.15
24/08/2006	EPA 415.3-1 (T) Ox	RPD - LCS Duplicate	2.008	%	10.06	9.86	0	0.15	
24/08/2006	EPA 415.3-1 (T) Ox	RPD - Check Standard	5.35	%	9.98	9.46	0	0.15	
24/08/2006	EPA 415.3-1 (T) Ox	LCS - Recovery	10.06	0.5 mg/L as C	0	0	1.006	0.85	1.15
24/08/2006	EPA 415.3-1 (T) Ox	Check Standard	9.98	0.5 mg/L as C	0	0	0.998	0.85	1.15
24/08/2006	EPA 415.3-1 (T) Ox	Check Standard	9.46	0.5 mg/L as C	0	0	0.946	0.85	1.15
05/09/2006	EPA 415.3-1 (T) Ox	Method Blank	0	0.5 mg/L as C	0	0			
05/09/2006	EPA 415.3-1 (T) Ox	RPD - LCS Duplicate	1.98	%	10.2	10	0	0.15	
05/09/2006	EPA 415.3-1 (T) Ox	LCS - Recovery	10.2	0.5 mg/L as C	0	0	1.02	0.85	1.15
05/09/2006	EPA 415.3-1 (T) Ox	LCS - Recovery	10.2	0.5 mg/L as C	0	0	1.02	0.85	1.15
05/09/2006	EPA 415.3-1 (T) Ox	LCS - Recovery	10	0.5 mg/L as C	0	0	1	0.85	1.15
05/09/2006	EPA 415.3-1 (T) Ox	Check Standard	9.18	0.5 mg/L as C	0	0	0.918	0.85	1.15
05/09/2006	EPA 415.3-1 (T) Ox	LCS - Recovery	10.2	0.5 mg/L as C	0	0	1.02	0.85	1.15
05/09/2006	EPA 415.3-1 (T) Ox	Method Blank	0	0.5 mg/L as C	0	0			

05/09/2006	EPA 415.3-1 (T) Ox	LCS - Recovery	10.2	0.5 mg/L as C	0	0	1.02	0.85	1.15
05/09/2006	EPA 415.3-1 (T) Ox	LCS - Recovery	10	0.5 mg/L as C	0	0	1	0.85	1.15
05/09/2006	EPA 415.3-1 (T) Ox	LCS - Recovery	10	0.5 mg/L as C	0	0	1	0.85	1.15
05/09/2006	EPA 415.3-1 (T) Ox	Check Standard	9.18	0.5 mg/L as C	0	0	0.918	0.85	1.15
05/09/2006	EPA 415.3-1 (T) Ox	Method Blank	0	0.5 mg/L as C	0	0			
05/09/2006	EPA 415.3-1 (T) Ox	LCS - Recovery	10	0.5 mg/L as C	0	0	1	0.85	1.15
05/09/2006	EPA 415.3-1 (T) Ox	RPD - LCS Duplicate	1.98	%	10.2	10		0	0.15
05/09/2006	EPA 415.3-1 (T) Ox	Check Standard	9.18	0.5 mg/L as C	0	0	0.918	0.85	1.15
05/09/2006	EPA 415.3-1 (T) Ox	Check Standard	10.4	0.5 mg/L as C	0	0	1.04	0.85	1.15
05/09/2006	EPA 415.3-1 (T) Ox	Check Standard	10.4	0.5 mg/L as C	0	0	1.04	0.85	1.15
05/09/2006	EPA 415.3-1 (T) Ox	RPD - LCS Duplicate	1.98	%	10.2	10		0	0.15
05/09/2006	EPA 415.3-1 (T) Ox	Check Standard	10.4	0.5 mg/L as C	0	0	1.04	0.85	1.15
05/09/2006	EPA 415.3-1 (T) Ox	Method Blank	0	0.5 mg/L as C	0	0			
05/09/2006	EPA 415.3-1 (T) Ox	RPD - LCS Duplicate	1.98	%	10.2	10		0	0.15
05/09/2006	EPA 415.3-1 (T) Ox	Check Standard	10.4	0.5 mg/L as C	0	0	1.04	0.85	1.15
05/09/2006	EPA 415.3-1 (T) Ox	Check Standard	9.18	0.5 mg/L as C	0	0	0.918	0.85	1.15
21/09/2006	EPA 415.3-1 (T) Ox	LCS - Recovery	9.98	0.5 mg/L as C	0	0	0.998	0.85	1.15
21/09/2006	EPA 415.3-1 (T) Ox	Check Standard	9.88	0.5 mg/L as C	0	0	0.988	0.85	1.15
21/09/2006	EPA 415.3-1 (T) Ox	LCS - Recovery	10	0.5 mg/L as C	0	0	1	0.85	1.15
21/09/2006	EPA 415.3-1 (T) Ox	RPD - LCS Duplicate	0.2002	%	9.98	10		0	0.15
21/09/2006	EPA 415.3-1 (T) Ox	Method Blank	0	0.5 mg/L as C	0	0			
21/09/2006	EPA 415.3-1 (T) Ox	RPD - Check Standard	10.21	%	9.88	8.92		0	0.15
21/09/2006	EPA 415.3-1 (T) Ox	Check Standard	8.92	0.5 mg/L as C	0	0	0.892	0.85	1.15
02/10/2006	EPA 415.3-1 (T) Ox	RPD - LCS Duplicate	2.281	%	10.2	9.97		0	0.15
02/10/2006	EPA 415.3-1 (T) Ox	RPD - LCS Duplicate	0.5043	%	9.94	9.89		0	0.15
02/10/2006	EPA 415.3-1 (T) Ox	RPD - Check Standard	7.234	%	9.86	10.6		0	0.15
02/10/2006	EPA 415.3-1 (T) Ox	RPD - LCS Duplicate	2.281	%	10.2	9.97		0	0.15
02/10/2006	EPA 415.3-1 (T) Ox	RPD - LCS Duplicate	0.5043	%	9.94	9.89		0	0.15
02/10/2006	EPA 415.3-1 (T) Ox	RPD - Check Standard	7.234	%	9.86	10.6		0	0.15
02/10/2006	EPA 415.3-1 (T) Ox	LCS - Recovery	9.89	0.5 mg/L as C	0	0	0.989	0.85	1.15
02/10/2006	EPA 415.3-1 (T) Ox	LCS - Recovery	9.94	0.5 mg/L as C	0	0	0.994	0.85	1.15
02/10/2006	EPA 415.3-1 (T) Ox	LCS - Recovery	10.2	0.5 mg/L as C	0	0	1.02	0.85	1.15
02/10/2006	EPA 415.3-1 (T) Ox	LCS - Recovery	10.2	0.5 mg/L as C	0	0	1.02	0.85	1.15
02/10/2006	EPA 415.3-1 (T) Ox	LCS - Recovery	9.94	0.5 mg/L as C	0	0	0.994	0.85	1.15
02/10/2006	EPA 415.3-1 (T) Ox	LCS - Recovery	9.97	0.5 mg/L as C	0	0	0.997	0.85	1.15
02/10/2006	EPA 415.3-1 (T) Ox	LCS - Recovery	9.89	0.5 mg/L as C	0	0	0.989	0.85	1.15
02/10/2006	EPA 415.3-1 (T) Ox	LCS - Recovery	9.97	0.5 mg/L as C	0	0	0.997	0.85	1.15

02/10/2006	EPA 415.3-1 (T) Ox	Check Standard	9.86	0.5 mg/L as C	0	0	0.986	0.85	1.15
02/10/2006	EPA 415.3-1 (T) Ox	Check Standard	9.86	0.5 mg/L as C	0	0	0.986	0.85	1.15
02/10/2006	EPA 415.3-1 (T) Ox	Method Blank	0	0.5 mg/L as C	0	0			
02/10/2006	EPA 415.3-1 (T) Ox	Method Blank	0	0.5 mg/L as C	0	0			
02/10/2006	EPA 415.3-1 (T) Ox	Method Blank	0	0.5 mg/L as C	0	0			
02/10/2006	EPA 415.3-1 (T) Ox	Method Blank	0	0.5 mg/L as C	0	0			
19/10/2006	EPA 415.3-1 (T) Ox	Method Blank	0	0.5 mg/L as C	0	0			
19/10/2006	EPA 415.3-1 (T) Ox	Check Standard	10.1	0.5 mg/L as C	0	0	1.01	0.85	1.15
19/10/2006	EPA 415.3-1 (T) Ox	Method Blank	0	0.5 mg/L as C	0	0			
19/10/2006	EPA 415.3-1 (T) Ox	Check Standard	10.1	0.5 mg/L as C	0	0	1.01	0.85	1.15
19/10/2006	EPA 415.3-1 (T) Ox	LCS - Recovery	9.61	0.5 mg/L as C	0	0	0.961	0.85	1.15
19/10/2006	EPA 415.3-1 (T) Ox	LCS - Recovery	9.64	0.5 mg/L as C	0	0	0.964	0.85	1.15
19/10/2006	EPA 415.3-1 (T) Ox	LCS - Recovery	9.61	0.5 mg/L as C	0	0	0.961	0.85	1.15
19/10/2006	EPA 415.3-1 (T) Ox	LCS - Recovery	9.64	0.5 mg/L as C	0	0	0.964	0.85	1.15
19/10/2006	EPA 415.3-1 (T) Ox	RPD - LCS Duplicate	0.3117	%	9.64	9.61		0	0.15
19/10/2006	EPA 415.3-1 (T) Ox	RPD - LCS Duplicate	0.3117	%	9.64	9.61		0	0.15
06/11/2006	EPA 415.3-1 (T) Ox	LCS - Recovery	9.98	0.5 mg/L as C	0	0	0.998	0.85	1.15
06/11/2006	EPA 415.3-1 (T) Ox	Check Standard	10.1	0.5 mg/L as C	0	0	1.01	0.85	1.15
06/11/2006	EPA 415.3-1 (T) Ox	RPD - Check Standard	5.769	%	10.1	10.7		0	0.15
06/11/2006	EPA 415.3-1 (T) Ox	RPD - LCS Duplicate	2.64	%	9.98	9.72		0	0.15
06/11/2006	EPA 415.3-1 (T) Ox	Method Blank	0	0.5 mg/L as C	0	0			
06/11/2006	EPA 415.3-1 (T) Ox	LCS - Recovery	9.72	0.5 mg/L as C	0	0	0.972	0.85	1.15
21/11/2006	EPA 415.3-1 (T) Ox	LCS - Recovery	10.1	0.5 mg/L as C	0	0	1.01	0.85	1.15
21/11/2006	EPA 415.3-1 (T) Ox	LCS - Recovery	10.2	0.5 mg/L as C	0	0	1.02	0.85	1.15
21/11/2006	EPA 415.3-1 (T) Ox	RPD - LCS Duplicate	0.9852	%	10.2	10.1		0	0.15
21/11/2006	EPA 415.3-1 (T) Ox	Check Standard	10.2	0.5 mg/L as C	0	0	1.02	0.85	1.15
21/11/2006	EPA 415.3-1 (T) Ox	Method Blank	0	0.5 mg/L as C	0	0			
21/11/2006	EPA 415.3-1 (T) Ox	RPD - Check Standard	2.899	%	10.2	10.5		0	0.15
04/12/2006	EPA 415.3-1 (T) Ox	RPD - Check Standard	1.573	%	9.46	9.61		0	0.15
04/12/2006	EPA 415.3-1 (T) Ox	RPD - LCS Duplicate	0.1026	%	9.75	9.74		0	0.15
04/12/2006	EPA 415.3-1 (T) Ox	Check Standard	9.46	0.5 mg/L as C	0	0	0.946	0.85	1.15
04/12/2006	EPA 415.3-1 (T) Ox	LCS - Recovery	9.74	0.5 mg/L as C	0	0	0.974	0.85	1.15
04/12/2006	EPA 415.3-1 (T) Ox	LCS - Recovery	9.75	0.5 mg/L as C	0	0	0.975	0.85	1.15
04/12/2006	EPA 415.3-1 (T) Ox	Method Blank	0	0.5 mg/L as C	0	0			
04/12/2006	EPA 415.3-1 (T) Ox	Check Standard	9.61	0.5 mg/L as C	0	0	0.961	0.85	1.15
04/12/2006	EPA 415.3-1 (T) Ox	Method Blank	0	0.5 mg/L as C	0	0			
21/12/2006	EPA 415.3-1 (T) Ox	LCS - Recovery	9.7	0.5 mg/L as C	0	0	0.97	0.85	1.15

21/12/2006	EPA 415.3-1 (T) Ox	RPD - LCS Duplicate	0.206	%	9.7	9.72	0	0.15	
21/12/2006	EPA 415.3-1 (T) Ox	LCS - Recovery	9.72	0.5 mg/L as C	0	0	0.972	0.85	1.15
21/12/2006	EPA 415.3-1 (T) Ox	Method Blank	0	0.5 mg/L as C	0	0			
02/01/2007	EPA 415.3-1 (T) Ox	Check Standard	10.4	0.5 mg/L as C	0	0	1.04	0.85	1.15
02/01/2007	EPA 415.3-1 (T) Ox	Method Blank	0	0.5 mg/L as C	0	0			
02/01/2007	EPA 415.3-1 (T) Ox	RPD - Check Standard	5.941	%	9.8	10.4	0	0.15	
02/01/2007	EPA 415.3-1 (T) Ox	RPD - LCS Duplicate	0.939	%	10.7	10.6	0	0.15	
02/01/2007	EPA 415.3-1 (T) Ox	Check Standard	9.8	0.5 mg/L as C	0	0	0.98	0.85	1.15
16/01/2007	EPA 415.3-1 (T) Ox	LCS - Recovery	9.85	0.5 mg/L as C	0	0	0.985	0.85	1.15
16/01/2007	EPA 415.3-1 (T) Ox	LCS - Recovery	9.77	0.5 mg/L as C	0	0	0.977	0.85	1.15
16/01/2007	EPA 415.3-1 (T) Ox	Check Standard	9.58	0.5 mg/L as C	0	0	0.958	0.85	1.15
16/01/2007	EPA 415.3-1 (T) Ox	RPD - Check Standard	1.964	%	9.77	9.58	0	0.15	
16/01/2007	EPA 415.3-1 (T) Ox	RPD - LCS Duplicate	0.8155	%	9.85	9.77	0	0.15	
16/01/2007	EPA 415.3-1 (T) Ox	Method Blank	0	0.5 mg/L as C	0	0			
16/01/2007	EPA 415.3-1 (T) Ox	Check Standard	9.77	0.5 mg/L as C	0	0	0.977	0.85	1.15
05/02/2007	EPA 415.3-1 (T) Ox	LCS - Recovery	9.6	0.5 mg/L as C	0	0	0.96	0.85	1.15
05/02/2007	EPA 415.3-1 (T) Ox	LCS - Recovery	9.7	0.5 mg/L as C	0	0	0.97	0.85	1.15
05/02/2007	EPA 415.3-1 (T) Ox	RPD - LCS Duplicate	1.036	%	9.7	9.6	0	0.15	
05/02/2007	EPA 415.3-1 (T) Ox	Check Standard	10.1	0.5 mg/L as C	0	0	1.01	0.85	1.15
05/02/2007	EPA 415.3-1 (T) Ox	Method Blank	0	0.5 mg/L as C	0	0			
05/02/2007	EPA 415.3-1 (T) Ox	Check Standard	9.8	0.5 mg/L as C	0	0	0.98	0.85	1.15
21/02/2007	EPA 415.3-1 (T) Ox	Check Standard	9.2	0.5 mg/L as C	0	0	0.92	0.85	1.15
21/02/2007	EPA 415.3-1 (T) Ox	Check Standard	9.2	0.5 mg/L as C	0	0	0.92	0.85	1.15
21/02/2007	EPA 415.3-1 (T) Ox	Check Standard	9.2	0.5 mg/L as C	0	0	0.92	0.85	1.15
21/02/2007	EPA 415.3-1 (T) Ox	Method Blank	0	0.5 mg/L as C	0	0			
21/02/2007	EPA 415.3-1 (T) Ox	LCS - Recovery	9.2	0.5 mg/L as C	0	0	0.92	0.85	1.15
21/02/2007	EPA 415.3-1 (T) Ox	LCS - Recovery	9.3	0.5 mg/L as C	0	0	0.93	0.85	1.15
21/02/2007	EPA 415.3-1 (T) Ox	LCS - Recovery	9.2	0.5 mg/L as C	0	0	0.92	0.85	1.15
21/02/2007	EPA 415.3-1 (T) Ox	LCS - Recovery	9.3	0.5 mg/L as C	0	0	0.93	0.85	1.15
21/02/2007	EPA 415.3-1 (T) Ox	RPD - LCS Duplicate	1.081	%	9.2	9.3	0	0.15	
21/02/2007	EPA 415.3-1 (T) Ox	Method Blank	0	0.5 mg/L as C	0	0			
21/02/2007	EPA 415.3-1 (T) Ox	Method Blank	0	0.5 mg/L as C	0	0			
21/02/2007	EPA 415.3-1 (T) Ox	LCS - Recovery	9.3	0.5 mg/L as C	0	0	0.93	0.85	1.15
21/02/2007	EPA 415.3-1 (T) Ox	RPD - LCS Duplicate	1.081	%	9.2	9.3	0	0.15	
21/02/2007	EPA 415.3-1 (T) Ox	RPD - LCS Duplicate	1.081	%	9.2	9.3	0	0.15	
21/02/2007	EPA 415.3-1 (T) Ox	LCS - Recovery	9.2	0.5 mg/L as C	0	0	0.92	0.85	1.15
06/03/2007	EPA 415.3-1 (T) Ox	LCS - Recovery	9.5	0.5 mg/L as C	0	0	0.95	0.85	1.15

06/03/2007	EPA 415.3-1 (T) Ox	Check Standard	9.9	0.5 mg/L as C	0	0	0.99	0.85	1.15
06/03/2007	EPA 415.3-1 (T) Ox	RPD - LCS Duplicate	2.128	%	9.3	9.5		0	0.15
06/03/2007	EPA 415.3-1 (T) Ox	LCS - Recovery	9.3	0.5 mg/L as C	0	0	0.93	0.85	1.15
06/03/2007	EPA 415.3-1 (T) Ox	Method Blank	0	0.5 mg/L as C	0	0			
21/03/2007	EPA 415.3-1 (T) Ox	Check Standard	9.5	0.5 mg/L as C	0	0	0.95	0.85	1.15
21/03/2007	EPA 415.3-1 (T) Ox	Method Blank	0	0.5 mg/L as C	0	0			
21/03/2007	EPA 415.3-1 (T) Ox	LCS - Recovery	9	0.5 mg/L as C	0	0	0.9	0.85	1.15
21/03/2007	EPA 415.3-1 (T) Ox	RPD - LCS Duplicate	3.209	%	9.5	9.2		0	0.15
21/03/2007	EPA 415.3-1 (T) Ox	LCS - Recovery	9.5	0.5 mg/L as C	0	0	0.95	0.85	1.15
21/03/2007	EPA 415.3-1 (T) Ox	Method Blank	0	0.5 mg/L as C	0	0			
21/03/2007	EPA 415.3-1 (T) Ox	LCS - Recovery	9.2	0.5 mg/L as C	0	0	0.92	0.85	1.15
21/03/2007	EPA 415.3-1 (T) Ox	RPD - LCS Duplicate	5.405	%	9.5	9		0	0.15
21/03/2007	EPA 415.3-1 (T) Ox	LCS - Recovery	9.5	0.5 mg/L as C	0	0	0.95	0.85	1.15
21/03/2007	EPA 415.3-1 (T) Ox	Check Standard	9.5	0.5 mg/L as C	0	0	0.95	0.85	1.15
03/04/2007	EPA 415.3-1 (T) Ox	Check Standard	8.8	0.5 mg/L as C	0	0	0.88	0.85	1.15
03/04/2007	EPA 415.3-1 (T) Ox	RPD - LCS Duplicate	1.07	%	9.4	9.3		0	0.15
03/04/2007	EPA 415.3-1 (T) Ox	LCS - Recovery	9.4	0.5 mg/L as C	0	0	0.94	0.85	1.15
03/04/2007	EPA 415.3-1 (T) Ox	Method Blank	0	0.5 mg/L as C	0	0			
03/04/2007	EPA 415.3-1 (T) Ox	LCS - Recovery	9.3	0.5 mg/L as C	0	0	0.93	0.85	1.15

Note

% is only for RPDs, Recovery, LowerConc and UpperConc. Recovery, Samp1_Conc, Samp2_Conc are mg/L