

1

Supporting Information

2 Occurrence of benzotriazole and 5-methylbenzotriazole in recycled 3 water, surface water and dishwashing detergents from Perth, 4 Western Australia

5
6 M.D. Alotaibi^{1,2,3}, B.M. Patterson^{*1,2}, A.J. McKinley¹, A.Y. Reeder⁴, A.J. Furness²

7 ¹School of Chemistry and Biochemistry, University of Western Australia, Crawley, Australia

8 ²CSIRO Land and Water Flagship, Private Bag 5, Wembley, 6913, WA, Australia

9 ³Department of Plant Protection, College of Food and Agricultural Sciences, King Saud University, Riyadh, Saudi Arabia

10 ⁴Centre for Microscopy, Characterization and Analysis, University of Western Australia, Crawley, Australia

11 *Corresponding author: Phone: +61-8-9333 6276; fax: +61-8-9333 6211; email: bradley.patterson@csiro.au

12

13

14

Table S1- Tabulates the averages of 11 replicates of BTri samples concentrations normalised to 100 ng L⁻¹.

Conc. ng/L	Avg	SD	T (95%)	Confidence Range	SD. Err	Confidence Level (95%)		
2	0.027	0.019	2.776	0.0163	0.0059	0.0274	±	0.0163
4	0.064	0.022	2.776	0.0181	0.0065	0.0641	±	0.0181
6	0.078	0.016	2.776	0.0134	0.0048	0.0781	±	0.0134
10	0.113	0.038	2.776	0.0315	0.0113	0.1133	±	0.0315
20	0.220	0.071	2.776	0.0595	0.0214	0.2199	±	0.0595
40	0.358	0.050	2.776	0.0415	0.0149	0.3576	±	0.0415
60	0.616	0.027	2.776	0.0305	0.0110	0.6159	±	0.0305
80	0.850	0.015	2.776	0.0167	0.0060	0.8504	±	0.0167
100	1.00	0	2.776	0	0	1	±	0

15 Where Avg= the average for 11 replicates data normalised to 100 ng/L, SD= standard deviation, T (95%)= statistical factor
16 that means T values at different confidence level summarise in a table which given T in confidence level 95%=
17 2.776, Confidence Range = (SD X T), St Err= Standard Error= (SD/ \sqrt{N}) and the Confidence level (95%)= (Avg+ Confidence
18 Range) ±(Avg- Confidence Range).

19

20

21

22

23

24

Table S2- Tabulates the averages of 11 replicates of 5-MeBT samples concentrations normalised to 100 ng L⁻¹.

Conc. ng/L	Avg	SD	T (95%)	Conf. Range	SD. Err	Conf. Level 95%		
2	0.025	0.0079	2.776	0.00662	0.00239	0.0250	±	0.0066
4	0.051	0.0074	2.776	0.00618	0.00223	0.0514	±	0.0062
6	0.094	0.0162	2.776	0.01359	0.00489	0.0941	±	0.0136
10	0.127	0.0212	2.776	0.01775	0.00639	0.1267	±	0.0178
20	0.229	0.0305	2.776	0.02550	0.00919	0.2286	±	0.0255
40	0.408	0.0261	2.776	0.02183	0.00786	0.4076	±	0.0218
60	0.602	0.0120	2.776	0.01359	0.00490	0.6015	±	0.0136
80	0.781	0.0254	2.776	0.02125	0.00766	0.7806	±	0.0213
100	1.00	0	2.776	0	0	1	±	0




Fig. S1. Calibration graphs for BTri and 5-MeBT normalised results with experimental errors¹.

REFERENCES

1. D. York, N. M. Evensen, M. L. Martinez and J. D. B. Delgado. Unified equations for the slope, intercept, and standard errors of the best straight line. *Am. J. Phys.* 2004, 72, 367-375.

